



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: pH

Supervisor Review By : JIGNESH

Run Number: LB98680

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.5	4.01	10/17/2018	17:10
2	CAL2	1	Water			22.7	7.00	10/17/2018	17:12
3	CAL3	1	Water			22.8	10.04	10/17/2018	17:14
4	ICV	1	Water			22.6	7.01	10/17/2018	17:16
5	CCV1	1	Water			22.5	2.00	10/17/2018	17:18
6	J5549-01	1	Solid	20.03	20	20.3	8.30	10/17/2018	17:30
7	J5549-01DUP	1	Solid	20.04	20	20.4	8.31	10/17/2018	17:32
8	J5549-02	1	Solid	20.02	20	20.4	8.41	10/17/2018	17:34
9	CCV2	1	Water			22.6	12.00	10/17/2018	17:38

WORKLIST(Hardcopy Internal Chain)

LB 98680

WorkList Name : pH101718-2

WorkList ID : 118137

Date : 10/17/2018 2:09:41 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/15/2018	Solid	J5549-01	pH	--Sel	USAR03	D42	CB-1018-S-TCLP-GRID-45	10/15/2018	9045D
10/15/2018	Solid	J5549-02	pH	--Sel	USAR03	D42	CB-1018-S-TCLP-GRID-50-2	10/15/2018	9045D

Date/Time 10/17/18 5:00
 Received by: SP
 Relinquished by: CP

Date/Time 10/17/18 5:40
 Received by: CP
 Relinquished by: SP