

Analytical Summary Report**Analysis Method:** 9045D**Analyst By :** Sweta**Parameter:** Corrosivity**Supervisor Review By :** apatel**Run Number:** LB98684**Slope :** 99.0**BalanceID:** WC SC-4**pH Meter ID :** WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.8	4.01	10/18/2018	10:40
2	CAL2	1	Water			22.6	7.00	10/18/2018	10:42
3	CAL3	1	Water			22.7	10.04	10/18/2018	10:44
4	ICV	1	Water			22.6	7.01	10/18/2018	10:46
5	CCV1	1	Water			22.5	2.00	10/18/2018	10:48
6	J5538-16	1	Solid	20.02	20	21.4	4.44	10/18/2018	10:50
7	J5538-16DUP	1	Solid	20.04	20	21.5	4.45	10/18/2018	10:52
8	J5538-17	1	Solid	20.02	20	21.6	9.03	10/18/2018	10:54
9	J5538-18	1	Solid	20.02	20	21.8	12.05	10/18/2018	10:56
10	J5538-19	1	Solid	20.03	20	21.0	2.23	10/18/2018	10:58
11	J5538-20	1	Solid	20.03	20	21.6	1.56	10/18/2018	11:00
12	J5538-21	1	Solid	20.02	20	21.5	1.60	10/18/2018	11:02
13	J5538-22	1	Solid	20.03	20	21.6	12.09	10/18/2018	11:04
14	J5538-23	1	Solid	20.02	20	21.8	12.30	10/18/2018	11:06
15	J5538-24	1	Solid	20.02	20	21.6	12.20	10/18/2018	11:08
16	CCV2	1	Water			22.5	12.00	10/18/2018	11:10

WORKLIST(Hardcopy Internal Chain)

LB 98684

WorkList Name : CORROSIONITY101818

WorkList ID : 118176

Date : 10/18/2018 10:04:29 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/12/2018	Solid	J5538-16	Corrosivity	Cool 4 deg C	ENVI77	NL_7	NL_7	10/12/2018	9045D
10/12/2018	Solid	J5538-17	Corrosivity	Cool 4 deg C	ENVI77	NL-7	NL-7	10/12/2018	9045D
10/12/2018	Solid	J5538-18	Corrosivity	Cool 4 deg C	ENVI77	BL01	BL01	10/12/2018	9045D
10/12/2018	Solid	J5538-19	Corrosivity	Cool 4 deg C	ENVI77	OXL01	OXL01	10/12/2018	9045D
10/12/2018	Solid	J5538-20	Corrosivity	Cool 4 deg C	ENVI77	AOXL	AOXL	10/12/2018	9045D
10/12/2018	Solid	J5538-21	Corrosivity	Cool 4 deg C	ENVI77	AL01	AL01	10/12/2018	9045D
10/12/2018	Solid	J5538-22	Corrosivity	Cool 4 deg C	ENVI77	CNL01	CNL01	10/12/2018	9045D
10/12/2018	Solid	J5538-23	Corrosivity	Cool 4 deg C	ENVI77	BS01	BS01	10/12/2018	9045D
10/12/2018	Solid	J5538-24	Corrosivity	Cool 4 deg C	ENVI77	CNS01	CNS01	10/12/2018	9045D

Date/Time 10/18/18 10:30
 Received by: SP
 Relinquished by: CP

Date/Time 10/18/18 11:30
 Received by: CP
 Relinquished by: SP