



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: pH

Supervisor Review By : JIGNESH

Run Number: LB98686

Slope : 99.0

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.8	4.01	10/18/2018	10:40
2	CAL2	1	Water			22.6	7.00	10/18/2018	10:42
3	CAL3	1	Water			22.7	10.04	10/18/2018	10:44
4	ICV	1	Water			22.6	7.01	10/18/2018	10:46
5	CCV1	1	Water			22.5	2.00	10/18/2018	10:48
6	J5538-16	1	Solid	20.02	20	21.4	4.44	10/18/2018	10:50
7	J5538-16DUP	1	Solid	20.04	20	21.5	4.45	10/18/2018	10:52
8	CCV2	1	Water			22.5	12.00	10/18/2018	11:10

WORKLIST(Hardcopy Internal Chain)

1398686

WorkList Name : pH101818

WorkList ID : 118178

Date : 10/18/2018 10:20:36 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/12/2018	Solid	J5538-16	pH	Cool 4 deg C	ENV177	D22	NL_7	10/12/2018	9045D

Date/Time 10/18/18 10:30
 Received by: SP
 Relinquished by: CP

Date/Time 10/18/18 11:30
 Received by: CP
 Relinquished by: SP