



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB98687

Slope : 97.5

pH Meter ID : WC PH MENTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.1	4.01	10/18/2018	13:35
2	CAL2	1	Water			20.1	7.01	10/18/2018	13:36
3	CAL3	1	Water			20.1	10.02	10/18/2018	13:37
4	ICV	1	Water			20.3	6.99	10/18/2018	13:40
5	CCV1	1	Water			20.2	2.01	10/18/2018	13:41
6	J5586-01	1	Water			16.5	6.86	10/18/2018	13:42
7	J5586-01DUP	1	Water			16.7	6.87	10/18/2018	13:43
8	CCV2	1	Water			20.2	12.01	10/18/2018	13:45

WorkList Name : TSS J5569

WorkList ID : 118154

Date : 10/18/2018 7:44:22 AM

WORKLIST(Hardcopy Internal Chain)

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/23/2018	Water	J5569-01 -H	TSS	Cool 4 deg C	CHAZ02	D31	RLF-GW-7S	10/16/2018	SM2540 D
10/23/2018	Water	J5569-02 -H	TSS	Cool 4 deg C	CHAZ02	D31	RLF-GW-9S	10/16/2018	SM2540 D
10/23/2018	Water	J5569-05 -H	TSS	Cool 4 deg C	CHAZ02	D31	RLF-GW-FD-01	10/16/2018	SM2540 D
10/23/2018	Water	J5569-06 -H	TSS	Cool 4 deg C	CHAZ02	D31	RLF-SW-1	10/16/2018	SM2540 D
10/23/2018	Water	J5569-07 -H	TSS	Cool 4 deg C	CHAZ02	D31	RLF-SW-3	10/16/2018	SM2540 D
10/23/2018	Water	J5569-08 -H	TSS	Cool 4 deg C	CHAZ02	D31	RLF-SW-4	10/16/2018	SM2540 D
10/23/2018	Water	J5569-09 -H	TSS	Cool 4 deg C	CHAZ02	D31	RLF-SW-5	10/16/2018	SM2540 D
10/23/2018	Water	J5569-12 -H	TSS	Cool 4 deg C	CHAZ02	D31	RLF-SW-8	10/16/2018	SM2540 D
10/23/2018	Water	J5569-13 -H	TSS	Cool 4 deg C	CHAZ02	D31	RLF-SW-FD-01	10/16/2018	SM2540 D
10/23/2018	Water	J5572-03 -F	TSS	Cool 4 deg C	SHAW22		OUTFALL001-181016	10/16/2018	SM2540 D

Date/Time 10-18-18 7:45 AM
 Received by: JA
 Relinquished by: JA

Date/Time 10-18-18 11:00 AM
 Received by: CS
 Relinquished by: JA