

245.1


LB98689

INSTRUMENT ID: CV2

Linear

 μ Abs.:

37096



A= 0.0000e+000

B= 2.7019e-004

C= 3.1856e-003

Rho= 0.9999821

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.010	0.010	24	0.000	24				
0.05	0.050	0.051	0.001	178	0.0 %	178				
0.2	0.200	0.182	-0.018	662	0.0 %	662				
2.5	2.500	2.528	0.028	9346	0.0 %	9346				
5.0	5.000	4.990	-0.010	18458	0.0 %	18458				
7.5	7.500	7.462	-0.038	27605	0.0 %	27605				
10.0	10.000	10.026	0.026	37096	0.0 %	37096				

LB98689 INSTRUMENT ID : CV1

Method: 245.1		Operator: Admin		Date of Analysis: 18 Oct 2018 14:45:12							
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type		
0.0 - 1	50	24	-	0.0000	245.1	PPB	18 Oct 2018 14:46:39	S	Std		
0.05 - 1	50.05	178	-	0.0500	245.1	PPB	18 Oct 2018 14:48:57	S	Std		
0.2 - 1	50.2	662	-	0.2000	245.1	PPB	18 Oct 2018 14:51:15	S	Std		
2.5 - 1	52.5	9346	-	2.5000	245.1	PPB	18 Oct 2018 14:53:33	S	Std		
5.0 - 1	55	18458	-	5.0000	245.1	PPB	18 Oct 2018 14:55:50	S	Std		
7.5 - 1	57.5	27605	-	7.5000	245.1	PPB	18 Oct 2018 14:58:06	S	Std		
10.0 - 1	510	37096	-	10.0000	245.1	PPB	18 Oct 2018 15:00:23	S	Std		
ICV29 - 1	ICV29	14719	3.9802	-245.1		PPB	18 Oct 2018 15:03:06	U	SMPL		
ICB29 - 1	ICB29	-71	-0.0160	-245.1		PPB	18 Oct 2018 15:05:21	U	SMPL		
CCV29 - 1	CCV29	18395	4.9734	-245.1		PPB	18 Oct 2018 15:07:37	U	SMPL		
CCB29 - 1	CCB29	-75	-0.0171	-245.1		PPB	18 Oct 2018 15:09:53	U	SMPL		
CRA - 1	CRA	175	0.0505	-245.1		PPB	18 Oct 2018 15:12:09	U	SMPL		
HighStd - 1	HighStd	37375	10.1017	-245.1		PPB	18 Oct 2018 15:14:25	U	SMPL		
ChkStd - 1	ChkStd	26003	7.0290	-245.1		PPB	18 Oct 2018 15:16:42	U	SMPL		
J5511-01 - 1	101118-SIDI-F-D-B	-208	-0.0530	-245.1		PPB	18 Oct 2018 15:18:59	U	SMPL		
J5511-02 - 1	101118-SIDI-F-D-M	-113	-0.0273	-245.1		PPB	18 Oct 2018 15:21:17	U	SMPL		
J5511-03 - 1	101118-SIDI-F-D-S	-90	-0.0211	-245.1		PPB	18 Oct 2018 15:23:34	U	SMPL		
J5511-04 - 1	101118-SIDI-F-U-B	-75	-0.0171	-245.1		PPB	18 Oct 2018 15:25:51	U	SMPL		
J5511-05 - 1	101118-SIDI-F-U-M	-80	-0.0184	-245.1		PPB	18 Oct 2018 15:28:09	U	SMPL		
J5511-06 - 1	101118-SIDI-F-U-S	-69	-0.0155	-245.1		PPB	18 Oct 2018 15:30:27	U	SMPL		
J5511-07 - 1	101118-TIPC-F-D-B	-80	-0.0184	-245.1		PPB	18 Oct 2018 15:32:44	U	SMPL		
CCV30 - 1	CCV30	17932	4.8483	-245.1		PPB	18 Oct 2018 15:34:59	U	SMPL		
CCB30 - 1	CCB30	-75	-0.0171	-245.1		PPB	18 Oct 2018 15:37:15	U	SMPL		
J5511-08 - 1	101118-TIPC-F-D-S	-83	-0.0192	-245.1		PPB	18 Oct 2018 15:39:31	U	SMPL		
J5511-09 - 1	101118-TIPC-F-U-B	-56	-0.0119	-245.1		PPB	18 Oct 2018 15:41:47	U	SMPL		
J5511-10 - 1	101118-TIPC-F-U-S	-85	-0.0198	-245.1		PPB	18 Oct 2018 15:44:03	U	SMPL		
J5541-01 - 1	101518-SIDI-E-D-B	-56	-0.0119	-245.1		PPB	18 Oct 2018 15:46:19	U	SMPL		
J5541-02 - 1	101518-SIDI-E-D-M	-56	-0.0119	-245.1		PPB	18 Oct 2018 15:48:36	U	SMPL		
J5541-03 - 1	101518-SIDI-E-D-S	-49	-0.0101	-245.1		PPB	18 Oct 2018 15:50:53	U	SMPL		
J5541-04 - 1	101518-SIDI-E-U-B	-64	-0.0141	-245.1		PPB	18 Oct 2018 15:53:10	U	SMPL		
J5541-05 - 1	101518-SIDI-E-U-M	-62	-0.0136	-245.1		PPB	18 Oct 2018 15:55:28	U	SMPL		
J5541-06 - 1	101518-SIDI-E-U-S	-60	-0.0130	-245.1		PPB	18 Oct 2018 15:57:46	U	SMPL		
J5541-07 - 1	101518-TIPC-F-D-B	-53	-0.0111	-245.1		PPB	18 Oct 2018 16:00:01	U	SMPL		
CCV31 - 1	CCV31	17828	4.8202	-245.1		PPB	18 Oct 2018 16:02:18	U	SMPL		
CCB31 - 1	CCB31	-70	-0.0157	-245.1		PPB	18 Oct 2018 16:04:34	U	SMPL		
J5541-08 - 1	101518-TIPC-F-D-S	-76	-0.0173	-245.1		PPB	18 Oct 2018 16:06:51	U	SMPL		
J5541-09 - 1	101518-TIPC-F-U-B	-63	-0.0138	-245.1		PPB	18 Oct 2018 16:09:07	U	SMPL		
J5541-10 - 1	101518-TIPC-F-U-S	-56	-0.0119	-245.1		PPB	18 Oct 2018 16:11:23	U	SMPL		
J5541-10DUP - 1	101518-TIPC-F-U-SDUP	-55	-0.0117	-245.1		PPB	18 Oct 2018 16:13:40	U	SMPL		
J5541-10MS - 1	101518-TIPC-F-U-SMS	13578	3.6719	-245.1		PPB	18 Oct 2018 16:15:57	U	SMPL		
J5541-10MSD - 1	101518-TIPC-F-U-SMSD	13422	3.6297	-245.1		PPB	18 Oct 2018 16:18:14	U	SMPL		
J5543-01 - 1	101618-TIPC-E-D-M	-89	-0.0209	-245.1		PPB	18 Oct 2018 16:20:31	U	SMPL		
J5543-02 - 1	101618-TIPC-E-U-M	-16	-0.0011	-245.1		PPB	18 Oct 2018 16:22:49	U	SMPL		
J5543-03 - 1	101618-SIDI-E-D-B	-56	-0.0119	-245.1		PPB	18 Oct 2018 16:25:06	U	SMPL		
J5543-04 - 1	101618-SIDI-E-D-M	-55	-0.0117	-245.1		PPB	18 Oct 2018 16:27:23	U	SMPL		
CCV32 - 1	CCV32	17614	4.7624	-245.1		PPB	18 Oct 2018 16:29:39	U	SMPL		
CCB32 - 1	CCB32	-73	-0.0165	-245.1		PPB	18 Oct 2018 16:31:56	U	SMPL		
J5543-05 - 1	101618-SIDI-E-D-S	-85	-0.0198	-245.1		PPB	18 Oct 2018 16:34:12	U	SMPL		
J5543-06 - 1	101618-SIDI-E-U-B	-56	-0.0119	-245.1		PPB	18 Oct 2018 16:36:29	U	SMPL		
J5543-07 - 1	101618-SIDI-E-U-M	-49	-0.0101	-245.1		PPB	18 Oct 2018 16:38:46	U	SMPL		
J5543-08 - 1	101618-SIDI-E-U-S	-51	-0.0106	-245.1		PPB	18 Oct 2018 16:41:02	U	SMPL		
J5543-09 - 1	101618-DBRI-E-D-B	-56	-0.0119	-245.1		PPB	18 Oct 2018 16:43:19	U	SMPL		
J5543-10 - 1	101618-DBRI-E-D-M	-62	-0.0136	-245.1		PPB	18 Oct 2018 16:45:35	U	SMPL		
J5543-11 - 1	101618-DBRI-E-D-S	-77	-0.0176	-245.1		PPB	18 Oct 2018 16:47:53	U	SMPL		
J5543-11DUP - 1	101618-DBRI-E-D-SDUP	-67	-0.0149	-245.1		PPB	18 Oct 2018 16:50:10	U	SMPL		
J5543-11MS - 1	101618-DBRI-E-D-SMS	13269	3.5884	-245.1		PPB	18 Oct 2018 16:52:28	U	SMPL		
J5543-11MSD - 1	101618-DBRI-E-D-SMSD	13555	3.6657	-245.1		PPB	18 Oct 2018 16:54:44	U	SMPL		
CCV33 - 1	CCV33	17650	4.7721	-245.1		PPB	18 Oct 2018 16:57:02	U	SMPL		
CCB33 - 1	CCB33	-90	-0.0211	-245.1		PPB	18 Oct 2018 16:59:18	U	SMPL		
J5541-10LX5 - 1		-51	-0.0106	-245.1		PPB	18 Oct 2018 17:01:35	U	SMPL		
J5541-10A - 1		14536	3.9307	-245.1		PPB	18 Oct 2018 17:03:52	U	SMPL		
J5543-11LX5 - 1		-86	-0.0201	-245.1		PPB	18 Oct 2018 17:06:08	U	SMPL		
J5543-11A - 1		14804	4.0031	-245.1		PPB	18 Oct 2018 17:08:25	U	SMPL		
PB113929BL - 1	PBW	-88	-0.0206	-245.1		PPB	18 Oct 2018 17:10:42	U	SMPL		
PB113929BS - 1	LCSW	14201	3.8402	-245.1		PPB	18 Oct 2018 17:12:59	U	SMPL		
PB113930BL - 1	PBW	-93	-0.0219	-245.1		PPB	18 Oct 2018 17:15:16	U	SMPL		
PB113930BS - 1	LCSW	13343	3.6084	-245.1		PPB	18 Oct 2018 17:17:34	U	SMPL		
CCV34 - 1	CCV34	17830	4.8207	-245.1		PPB	18 Oct 2018 17:19:52	U	SMPL		
CCB34 - 1	CCB34	-91	-0.0214	-245.1		PPB	18 Oct 2018 17:22:08	U	SMPL		