



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 10/19/2018

OVENTEMP IN Celsius(°C): 107
Time IN: 17:15
In Date: 10/18/2018
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M Oven-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:11
Out Date: 10/19/2018
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M Sc-1

QC:LB98703

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
J5195-03	JC-02-101818-A	1	1.13	9.57	8.6	88.5
J5195-05	JC-02-101818-B	2	1.18	9.7	8.75	88.8
J5574-01	FK-BPM-03-003-ABCD	3	1.14	9.81	8.6	86.0
J5574-02	FK-BPM-03-003-ABCD	4	1.18	9.96	8.9	87.9
J5585-01	SSSI-1017-S-DH-20	5	1.14	9.79	9.44	96.0
J5585-02	SSSI-1017-S-DH-20	6	1.13	9.57	9.19	95.5
J5585-03	SSSI-1017-S-DH-19	7	1.17	9.94	9.6	96.1
J5585-04	SSSI-1017-S-DH-19	8	1.16	9.98	9.55	95.1
J5585-05	SSSI-1017-S-DH-19A	9	1.13	9.72	9.17	93.6
J5585-06	SSSI-1017-S-DH-19A	10	1.11	9.93	9.5	95.1
J5590-01	LS-SOIL-1	11	1.14	9.71	9.02	91.9
J5590-02	LS-SOIL-1-A	12	1.14	9.71	9.02	91.9
J5590-03	LS-SOIL-1-B	13	1.14	9.71	9.02	91.9
J5590-04	LS-SOIL-1	14	1.14	9.71	9.02	91.9
J5590-05	LS-SOIL-2	15	1.12	9.79	8.51	85.2
J5590-06	LS-SOIL-2-A	16	1.12	9.79	8.51	85.2
J5590-07	LS-SOIL-2-B	17	1.12	9.79	8.51	85.2
J5590-08	LS-SOIL-2	18	1.12	9.79	8.51	85.2
J5592-03	M167-5A	19	1.14	9.86	8.21	81.1
J5592-06	M167-6A	20	1.1	9.82	8.1	80.3
J5592-09	M167-7A	21	1.15	9.95	8.39	82.3
J5592-11	M167-1A	22	1.17	9.97	8.22	80.1
J5592-14	M167-2A	23	1.11	9.62	7.27	72.4
J5592-17	M167-3A	24	1.16	9.89	8.38	82.7
J5592-20	M167-4A	25	1.19	9.94	8.42	82.6
J5592-22	M167-8A	26	1.16	9.89	8.33	82.1
J5592-25	M167-DUP1	27	1.1	9.91	7.53	73.0
J5596-01	A369177	28	1.00	2.00	2.00	100.0
J5596-02	B369318	29	1.00	2.00	2.00	100.0

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$