


LB 98704

7471B

INSTRUMENT ID: CV1

 μ Abs.:

53866



A= 0.0000e+000

B= 1.8564e-004

C= 1.4871e-002

Rho= 0.9999886

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.027	0.027	64	0.000	64				
0.2	0.200	0.193	-0.007	957	0.0 %	957				
2.5	2.500	2.475	-0.025	13254	0.0 %	13254				
5.0	5.000	5.006	0.006	26887	0.0 %	26887				
7.5	7.500	7.485	-0.015	40239	0.0 %	40239				
10.0	10.000	10.014	0.014	53866	0.0 %	53866				

LB98704 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 18 Oct 2018 17:39:24

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc/Method	Units	Date	Type	Type
0.0 - 1	50	64	-	0.00007471B	PPB	18 Oct 2018 17:40:54	S	Std
0.2 - 1	50.2	957	-	0.20007471B	PPB	18 Oct 2018 17:43:12	S	Std
2.5 - 1	52.5	13254	-	2.50007471B	PPB	18 Oct 2018 17:45:30	S	Std
5.0 - 1	55	26887	-	5.00007471B	PPB	18 Oct 2018 17:47:48	S	Std
7.5 - 1	57.5	40239	-	7.50007471B	PPB	18 Oct 2018 17:50:05	S	Std
10.0 - 1	510	53866	-	10.00007471B	PPB	18 Oct 2018 17:52:22	S	Std
ICV30 - 1	ICV30	21228	3.9556	-7471B	PPB	18 Oct 2018 17:55:07	U	SMPL
ICB30 - 1	ICB30	-11	0.0128	-7471B	PPB	18 Oct 2018 17:57:22	U	SMPL
CCV35 - 1	CCV35	26611	4.9549	-7471B	PPB	18 Oct 2018 17:59:38	U	SMPL
CCB35 - 1	CCB35	-18	0.0115	-7471B	PPB	18 Oct 2018 18:01:54	U	SMPL
CRA - 1	CRA	923	0.1862	-7471B	PPB	18 Oct 2018 18:04:10	U	SMPL
HighStd - 1	HighStd	53892	10.0193	-7471B	PPB	18 Oct 2018 18:06:26	U	SMPL
ChkStd - 1	ChkStd	37678	7.0094	-7471B	PPB	18 Oct 2018 18:08:41	U	SMPL
J5570-01 - 1	TS-06	4024	0.7619	-7471B	PPB	18 Oct 2018 18:10:59	U	SMPL
J5570-02 - 1	TS-07	3829	0.7257	-7471B	PPB	18 Oct 2018 18:13:17	U	SMPL
J5570-03 - 1	TS-08	4635	0.8753	-7471B	PPB	18 Oct 2018 18:15:35	U	SMPL
J5573-01 - 1	0689-KM-CA-COMP	94911	17.6340	-7471B	PPB	18 Oct 2018 18:17:52	U	SMPL
J5573-02 - 1	0692-KM-CA-AMD-COMP	97017	18.0250	-7471B	PPB	18 Oct 2018 18:20:10	U	SMPL
J5575-02 - 1	SSSI-1015-S-DH-28	-45	0.0065	-7471B	PPB	18 Oct 2018 18:22:31	U	SMPL
J5575-04 - 1	SSSI-1015-S-DH-27	331	0.0763	-7471B	PPB	18 Oct 2018 18:24:49	U	SMPL
CCV36 - 1	CCV36	26514	4.9369	-7471B	PPB	18 Oct 2018 18:27:05	U	SMPL
CCB36 - 1	CCB36	-47	0.0061	-7471B	PPB	18 Oct 2018 18:29:20	U	SMPL
J5575-06 - 1	SSSI-1015-S-DH-26	516	0.1107	-7471B	PPB	18 Oct 2018 18:31:37	U	SMPL
J5575-06DUP - 1	SSSI-1015-S-DH-26DUP	464	0.1010	-7471B	PPB	18 Oct 2018 18:33:53	U	SMPL
J5575-06MS - 1	SSSI-1015-S-DH-26MS	20307	3.7846	-7471B	PPB	18 Oct 2018 18:36:09	U	SMPL
J5575-06MSD - 1	SSSI-1015-S-DH-26MSD	19796	3.6898	-7471B	PPB	18 Oct 2018 18:38:26	U	SMPL
J5575-08 - 1	SSSI-1015-S-DH-25A-T	558	0.1185	-7471B	PPB	18 Oct 2018 18:40:42	U	SMPL
J5575-10 - 1	SSSI-1015-S-DH-25A-B	454	0.0992	-7471B	PPB	18 Oct 2018 18:42:59	U	SMPL
J5575-12 - 1	SSSI-1015-S-DH-25-T	882	0.1786	-7471B	PPB	18 Oct 2018 18:45:16	U	SMPL
J5575-14 - 1	SSSI-1015-S-DH-25-B	590	0.1244	-7471B	PPB	18 Oct 2018 18:47:33	U	SMPL
J5575-16 - 1	SSSI-1015-S-DH-24-T	581	0.1227	-7471B	PPB	18 Oct 2018 18:49:51	U	SMPL
J5575-18 - 1	SSSI-1015-S-DH-24-B	490	0.1058	-7471B	PPB	18 Oct 2018 18:52:07	U	SMPL
CCV37 - 1	CCV37	26602	4.9532	-7471B	PPB	18 Oct 2018 18:54:23	U	SMPL
CCB37 - 1	CCB37	-27	0.0099	-7471B	PPB	18 Oct 2018 18:56:39	U	SMPL
J5575-20 - 1	SSSI-1016-S-DH-23	801	0.1636	-7471B	PPB	18 Oct 2018 18:58:56	U	SMPL
J5577-02 - 1	SSSI-1016-S-DH-22	1301	0.2564	-7471B	PPB	18 Oct 2018 19:01:12	U	SMPL
J5577-04 - 1	SSSI-1016-S-DH-21	460	0.1003	-7471B	PPB	18 Oct 2018 19:03:29	U	SMPL
J5579-01 - 1	7-1-ROLLOFFS	7258	1.3622	-7471B	PPB	18 Oct 2018 19:05:45	U	SMPL
J5202-01 - 1	PL-03-101718-A	844	0.1716	-7471B	PPB	18 Oct 2018 19:08:02	U	SMPL
J5202-03 - 1	PL-03-101718-B	1017	0.2037	-7471B	PPB	18 Oct 2018 19:10:19	U	SMPL
J5202-05 - 1	PL-03-101718-C	1039	0.2077	-7471B	PPB	18 Oct 2018 19:12:36	U	SMPL
J5204-05 - 1	SU-03-101718-A	3867	0.7327	-7471B	PPB	18 Oct 2018 19:14:54	U	SMPL
J5204-07 - 1	SU-03-101718-B	3031	0.5775	-7471B	PPB	18 Oct 2018 19:17:12	U	SMPL
J5538-04 - 1	NL_7	342	0.0784	-7471B	PPB	18 Oct 2018 19:19:28	U	SMPL
CCV38 - 1	CCV38	26630	4.9584	-7471B	PPB	18 Oct 2018 19:21:44	U	SMPL
CCB38 - 1	CCB38	-20	0.0112	-7471B	PPB	18 Oct 2018 19:24:01	U	SMPL
J5538-05 - 1	NL-7	172	0.0468	-7471B	PPB	18 Oct 2018 19:26:17	U	SMPL
J5538-06 - 1	BL01	124	0.0379	-7471B	PPB	18 Oct 2018 19:28:34	U	SMPL
J5538-07 - 1	OXL01	1685	0.3277	-7471B	PPB	18 Oct 2018 19:30:51	U	SMPL
J5538-08 - 1	AOXL	4892	0.9230	-7471B	PPB	18 Oct 2018 19:33:07	U	SMPL
J5538-09 - 1	AL01	159	0.0444	-7471B	PPB	18 Oct 2018 19:35:24	U	SMPL
J5538-10 - 1	CNL01	308	0.0720	-7471B	PPB	18 Oct 2018 19:37:40	U	SMPL
J5565-01 - 1	RLF-SD-1	988	0.1983	-7471B	PPB	18 Oct 2018 19:39:58	U	SMPL
J5565-02 - 1	RLF-SD-3	1453	0.2846	-7471B	PPB	18 Oct 2018 19:42:15	U	SMPL
J5565-03 - 1	RLF-SD-4	2240	0.4307	-7471B	PPB	18 Oct 2018 19:44:33	U	SMPL
J5565-04 - 1	RLF-SD-5	2222	0.4274	-7471B	PPB	18 Oct 2018 19:46:50	U	SMPL
CCV39 - 1	CCV39	26997	5.0266	-7471B	PPB	18 Oct 2018 19:49:06	U	SMPL
CCB39 - 1	CCB39	3	0.0154	-7471B	PPB	18 Oct 2018 19:51:23	U	SMPL
J5565-04DUP - 1	RLF-SD-5DUP	1965	0.3797	-7471B	PPB	18 Oct 2018 19:53:39	U	SMPL
J5565-05 - 1	J5565-04MS	22235	4.1425	-7471B	PPB	18 Oct 2018 19:55:56	U	SMPL
J5565-06 - 1	J5565-04MSD	20515	3.8232	-7471B	PPB	18 Oct 2018 19:58:12	U	SMPL
J5565-07 - 1	RLF-SD-8	3644	0.6913	-7471B	PPB	18 Oct 2018 20:00:29	U	SMPL
J5565-08 - 1	RLF-SD-FD-01	1765	0.3425	-7471B	PPB	18 Oct 2018 20:02:46	U	SMPL
J5575-06LX5 - 1		113	0.0358	-7471B	PPB	18 Oct 2018 20:05:03	U	SMPL
J5575-06A - 1		20905	3.8956	-7471B	PPB	18 Oct 2018 20:07:21	U	SMPL
J5565-04LX5 - 1		405	0.0901	-7471B	PPB	18 Oct 2018 20:09:38	U	SMPL
J5565-04A - 1		22941	4.2736	-7471B	PPB	18 Oct 2018 20:11:56	U	SMPL
CCV40 - 1	CCV40	27190	5.0624	-7471B	PPB	18 Oct 2018 20:14:12	U	SMPL
CCB40 - 1	CCB40	-47	0.0061	-7471B	PPB	18 Oct 2018 20:16:30	U	SMPL
PB113943BL - 1	PBS	32	0.0208	-7471B	PPB	18 Oct 2018 20:18:46	U	SMPL
PB113943BS - 1	LCSS	20361	3.7947	-7471B	PPB	18 Oct 2018 20:21:03	U	SMPL
PB113944BL - 1	PBS	-16	0.0119	-7471B	PPB	18 Oct 2018 20:23:20	U	SMPL
PB113944BS - 1	LCSS	20312	3.7856	-7471B	PPB	18 Oct 2018 20:25:37	U	SMPL
CCV41 - 1	CCV41	26675	4.9668	-7471B	PPB	18 Oct 2018 20:32:28	U	SMPL
CCB41 - 1	CCB41	-25	0.0102	-7471B	PPB	18 Oct 2018 20:34:45	U	SMPL