

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: jignesh

Date: 10/19/2018

TEMP1 IN: 104 10/19/2018 07:00 TEMP OUT: 104 10/19/2018 08:05

TEMP2 IN: 104 10/19/2018 08:35 TEMP OUT: 104 10/19/2018 09:35

TEMP3 IN: 104 10/19/2018 10:50 TEMP OUT: 103 10/19/2018 13:35

TEMP4 IN: 104 10/19/2018 15:05 TEMP OUT: 104 10/19/2018 15:40

Run Number: LB98709

BalanceID: wc-sc-3

OvenID: wc-oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB98709BL	LB98709BL	2.3697	2.3697	100	2.3697	2.3697	2.3697	0.0000	0
2	LB98709BS	LB98709BS	2.1603	2.1603	100	2.2148	2.2148	2.2148	0.0545	545
3	J5593-01	101718-TIPC-E-D-M	1.4000	1.4000	400	1.4088	1.4088	1.4088	0.0088	22
4	J5593-02	101718-TIPC-E-U-M	1.3996	1.3996	400	1.4075	1.4075	1.4075	0.0079	19.8
5	J5593-03	101718-SIDI-E-D-B	1.3995	1.3995	400	1.4204	1.4204	1.4204	0.0209	52.2
6	J5593-04	101718-SIDI-E-D-M	1.4029	1.4029	400	1.4127	1.4127	1.4127	0.0098	24.5
7	J5593-05	101718-SIDI-E-D-S	1.3914	1.3914	400	1.3990	1.3990	1.3990	0.0076	19
8	J5593-06	101718-DBRI-E-D-B	1.4065	1.4065	400	1.4239	1.4239	1.4239	0.0174	43.5
9	J5593-07	101718-DBRI-E-D-M	1.4147	1.4147	400	1.4255	1.4255	1.4255	0.0108	27
10	J5593-08	101718-DBRI-E-D-S	1.3985	1.3985	400	1.4044	1.4044	1.4044	0.0059	14.8
11	J5593-09	101718-DBRI-E-U-B	1.3995	1.3995	400	1.4073	1.4073	1.4073	0.0078	19.5
12	J5593-10	101718-DBRI-E-U-M	1.3962	1.3962	200	1.3987	1.3987	1.3987	0.0025	12.5
13	J5593-10DUP	101718-DBRI-E-U-MDUP	1.4016	1.4016	200	1.4042	1.4042	1.4042	0.0026	13
14	J5593-11	101718-DBRI-E-U-S	1.4017	1.4017	400	1.4060	1.4060	1.4060	0.0043	10.8
15	J5593-12	101718-JETT-E-D-B	1.4053	1.4053	400	1.4127	1.4127	1.4127	0.0074	18.5

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: jignesh

Date: 10/19/2018

TEMP1 IN: 104 10/19/2018 07:00 TEMP OUT: 104 10/19/2018 08:05

TEMP2 IN: 104 10/19/2018 08:35 TEMP OUT: 104 10/19/2018 09:35

TEMP3 IN: 104 10/19/2018 10:50 TEMP OUT: 103 10/19/2018 13:35

TEMP4 IN: 104 10/19/2018 15:05 TEMP OUT: 104 10/19/2018 15:40

Run Number: LB98709

BalanceID: wc-sc-3

OvenID: wc-oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	J5593-13	101718-JETT-E-D-M	1.3972	1.3972	400	1.4023	1.4023	1.4023	0.0051	12.8
17	J5593-14	101718-JETT-E-D-S	1.4016	1.4016	400	1.4062	1.4062	1.4062	0.0046	11.5
18	J5593-15	101718-JETT-E-U-B	1.4056	1.4056	400	1.4152	1.4152	1.4152	0.0096	24
19	J5593-16	101718-JETT-E-U-M	1.3999	1.3999	400	1.4058	1.4058	1.4058	0.0059	14.8
20	J5593-17	101718-JETT-E-U-S	1.4011	1.4011	400	1.4057	1.4057	1.4057	0.0046	11.5

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

1308709

WorkList Name : TSS J5593

WorkList ID : 118220

Date : 10/19/2018 7:29:01 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/22/2018	Water	J5593-01 -C	TSS	Cool 4 deg C	HDR103	E11	101718-TIPC-E-D-M	10/17/2018	SM2540 D
10/22/2018	Water	J5593-02 -C	TSS	Cool 4 deg C	HDR103	E11	101718-TIPC-E-U-M	10/17/2018	SM2540 D
10/24/2018	Water	J5593-03 -C	TSS	Cool 4 deg C	HDR103	E11	101718-SIDI-E-D-B	10/17/2018	SM2540 D
10/24/2018	Water	J5593-04 -C	TSS	Cool 4 deg C	HDR103	E11	101718-SIDI-E-D-M	10/17/2018	SM2540 D
10/24/2018	Water	J5593-05 -C	TSS	Cool 4 deg C	HDR103	E11	101718-SIDI-E-D-S	10/17/2018	SM2540 D
10/24/2018	Water	J5593-06 -C	TSS	Cool 4 deg C	HDR103	E11	101718-DBRI-E-D-B	10/17/2018	SM2540 D
10/24/2018	Water	J5593-07 -C	TSS	Cool 4 deg C	HDR103	E11	101718-DBRI-E-D-M	10/17/2018	SM2540 D
10/24/2018	Water	J5593-08 -C	TSS	Cool 4 deg C	HDR103	E11	101718-DBRI-E-D-S	10/17/2018	SM2540 D
10/24/2018	Water	J5593-09 -C	TSS	Cool 4 deg C	HDR103	E11	101718-DBRI-E-U-B	10/17/2018	SM2540 D
10/24/2018	Water	J5593-10 -C	TSS	Cool 4 deg C	HDR103	E11	101718-DBRI-E-U-M	10/17/2018	SM2540 D
10/24/2018	Water	J5593-11 -C	TSS	Cool 4 deg C	HDR103	E11	101718-DBRI-E-U-S	10/17/2018	SM2540 D
10/24/2018	Water	J5593-12 -C	TSS	Cool 4 deg C	HDR103	E11	101718-JETT-E-D-B	10/17/2018	SM2540 D
10/24/2018	Water	J5593-13 -C	TSS	Cool 4 deg C	HDR103	E11	101718-JETT-E-D-M	10/17/2018	SM2540 D
10/24/2018	Water	J5593-14 -C	TSS	Cool 4 deg C	HDR103	E11	101718-JETT-E-D-S	10/17/2018	SM2540 D
10/24/2018	Water	J5593-15 -C	TSS	Cool 4 deg C	HDR103	E11	101718-JETT-E-U-B	10/17/2018	SM2540 D
10/24/2018	Water	J5593-16 -C	TSS	Cool 4 deg C	HDR103	E11	101718-JETT-E-U-M	10/17/2018	SM2540 D
10/24/2018	Water	J5593-17 -C	TSS	Cool 4 deg C	HDR103	E11	101718-JETT-E-U-S	10/17/2018	SM2540 D

Date/Time 10/19/18 7:35 AM
 Received by: JS
 Relinquished by: JS

Date/Time 10-14-18 12:100 PM
 Received by: JS
 Relinquished by: JS