



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: pH

Supervisor Review By : JIGNESH

Run Number: LB98714

Slope : 99.1

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.8	4.01	10/18/2018	17:15
2	CAL2	1	Water			22.6	7.00	10/18/2018	17:17
3	CAL3	1	Water			22.7	10.04	10/18/2018	17:19
4	ICV	1	Water			22.6	7.01	10/18/2018	17:21
5	CCV1	1	Water			22.5	2.00	10/18/2018	17:23
6	J5590-09	1	Water			15.1	9.42	10/18/2018	17:25
7	J5590-09DUP	1	Water			15.2	9.43	10/18/2018	17:27
8	CCV2	1	Water			22.6	12.00	10/18/2018	17:40

WORKLIST(Hardcopy Internal Chain)

LB 98714

WorkList Name : pH101818-2

WorkList ID : 118207

Date : 10/18/2018 3:58:54 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/18/2018	Water	J5590-09	pH	Cool 4 deg C	PSEG01	P63	LS-DRUM	10/18/2018	9040C

Date/Time 10/18/18 5:05
 Received by: SP
 Relinquished by: CP

Date/Time 10/18/18 5:45
 Received by: CP
 Relinquished by: SP