



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : JIGNESH

Run Number: LB98716

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.8	4.01	10/18/2018	17:15
2	CAL2	1	Water			22.6	7.00	10/18/2018	17:17
3	CAL3	1	Water			22.7	10.04	10/18/2018	17:19
4	ICV	1	Water			22.6	7.01	10/18/2018	17:21
5	CCV1	1	Water			22.5	2.00	10/18/2018	17:23
6	J5195-04	1	Solid	20.02	20	22.3	9.47	10/18/2018	17:30
7	J5195-04DUP	1	Solid	20.03	20	22.4	9.48	10/18/2018	17:32
8	J5195-06	1	Solid	20.02	20	22.2	9.37	10/18/2018	17:34
9	J5590-04	1	Solid	20.04	20	21.0	6.78	10/18/2018	17:36
10	J5590-08	1	Solid	20.03	20	21.9	5.89	10/18/2018	17:38
11	CCV2	1	Water			22.6	12.00	10/18/2018	17:40

WORKLIST(Hardcopy Internal Chain)

LB 98716

WorkList Name : CORROSIVITY101818-1 WorkList ID : 118206 Date : 10/18/2018 3:58:41 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/18/2018	Solid	J5195-04	Corrosivity	Cool 4 deg C	PSEG05		JC-02-101818-A	10/18/2018	9045D
10/18/2018	Solid	J5195-06	Corrosivity	Cool 4 deg C	PSEG05		JC-02-101818-B	10/18/2018	9045D
10/18/2018	Solid	J5590-04	Corrosivity	Cool 4 deg C	PSEG01	P63	LS-SOIL-1	10/18/2018	9045D
10/18/2018	Solid	J5590-08	Corrosivity	Cool 4 deg C	PSEG01	P63	LS-SOIL-2	10/18/2018	9045D

Date/Time 10/18/18 5:05
 Received by: SP
 Relinquished by: CP

Date/Time 10/18/18 5:05
 Received by: CP
 Relinquished by: SP