


ISM02.4

LB98717
INSTRUMENT ID: CV4

Linear

 μ Abs.:

40617



A= 0.0000e+000

B= 2.4615e-004

C= -1.6216e-002

Rho= 0.9999214

Accept=Accepted

Handwritten notes:
 slope
 $\gamma = 1.21 \times 10^{-4}$

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.039	-0.039	-91	0.000	-91					-
0.2	0.200	0.177	-0.023	785	0.0 %	785					-12
2.5	2.500	2.598	0.098	10622	0.0 %	10622					4
5.0	5.000	4.973	-0.027	20269	0.0 %	20269					-1
7.5	7.500	7.509	0.009	30571	0.0 %	30571					8
10.0	10.000	9.982	-0.018	40617	0.0 %	40617					

Handwritten signature:
 10/19/18

LB98717
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	-91	-		0 ISM02.4	PPB	10/19/2018 10:45	Std
	0.2 S0.2	785	-		0.2 ISM02.4	PPB	10/19/2018 10:47	Std
	2.5 S2.5	10622	-		2.5 ISM02.4	PPB	10/19/2018 10:49	Std
	5 S5	20269	-		5 ISM02.4	PPB	10/19/2018 10:52	Std
	7.5 S7.5	30571	-		7.5 ISM02.4	PPB	10/19/2018 10:54	Std
	10 S10.0	40617	-		10 ISM02.4	PPB	10/19/2018 11:00	Std
ICV	ICV	15903	3.8983 -		ISM02.4	PPB	10/19/2018 11:03	SMPL
ICB	ICB	-156	-0.0546 -		ISM02.4	PPB	10/19/2018 11:05	SMPL
CCV042	CCV042	20192	4.954 -		ISM02.4	PPB	10/19/2018 11:07	SMPL
CCB042	CCB042	-162	-0.0561 -		ISM02.4	PPB	10/19/2018 11:10	SMPL
PB113988BL	PBW0188	9	-0.014 -		ISM02.4	PPB	10/19/2018 11:12	SMPL
J5587-01	MC0AG2	-93	-0.0391 -		ISM02.4	PPB	10/19/2018 11:14	SMPL
J5587-02	MC0AG3	-89	-0.0381 -		ISM02.4	PPB	10/19/2018 11:16	SMPL
J5587-03	MC0AG4	-75	-0.0347 -		ISM02.4	PPB	10/19/2018 11:19	SMPL
J5587-04	MC0AG5	-94	-0.0394 -		ISM02.4	PPB	10/19/2018 11:21	SMPL
J5587-05	MC0AG6	-99	-0.0406 -		ISM02.4	PPB	10/19/2018 11:23	SMPL
J5587-06	MC0AG7	-85	-0.0371 -		ISM02.4	PPB	10/19/2018 11:25	SMPL
J5587-07	MC0AG7D	-89	-0.0381 -		ISM02.4	PPB	10/19/2018 11:28	SMPL
J5587-08	MC0AG7S	4150	1.0053 -		ISM02.4	PPB	10/19/2018 11:30	SMPL
CCV043	CCV043	20562	5.0451 -		ISM02.4	PPB	10/19/2018 11:32	SMPL
CCB043	CCB043	-149	-0.0529 -		ISM02.4	PPB	10/19/2018 11:35	SMPL