



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB98719

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.6	10/18/2018	11:20
2	J5538-16		1	8	>100.0	>212.0	>212.0	10/18/2018	11:50
3	J5538-16DUP		1	8	>100.0	>212.0	>212.0	10/18/2018	12:15
4	J5538-17		1	7	>100.0	>212.0	>212.0	10/18/2018	12:40
5	J5538-18		1	9	>100.0	>212.0	>212.0	10/18/2018	13:05
6	J5538-19		1	8	>100.0	>212.0	>212.0	10/18/2018	13:45
7	J5538-20		1	8	>100.0	>212.0	>212.0	10/18/2018	14:10
8	J5538-21		1	9	>100.0	>212.0	>212.0	10/18/2018	14:40
9	J5538-22		1	6	>100.0	>212.0	>212.0	10/18/2018	15:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 98719

WorkList Name : fp101818

WorkList ID : 118180

Date : 10/18/2018 10:36:20 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/19/2018	Solid	J5538-16 ✓	Flash Point	Cool 4 deg C	ENVI77	D22	NL_7	10/12/2018	1010A
10/19/2018	Solid	J5538-17 ✓	Flash Point	Cool 4 deg C	ENVI77	D22	NL-7	10/12/2018	1010A
10/19/2018	Solid	J5538-18 ✓	Flash Point	Cool 4 deg C	ENVI77	D22	BL01	10/12/2018	1010A
10/19/2018	Solid	J5538-19 ✓	Flash Point	Cool 4 deg C	ENVI77	D22	OXL01	10/12/2018	1010A
10/19/2018	Solid	J5538-20 ✓	Flash Point	Cool 4 deg C	ENVI77	D22	AOXL	10/12/2018	1010A
10/19/2018	Solid	J5538-21 ✓	Flash Point	Cool 4 deg C	ENVI77	D22	AL01	10/12/2018	1010A
10/19/2018	Solid	J5538-22 ✓	Flash Point	Cool 4 deg C	ENVI77	D22	CNL01	10/12/2018	1010A

Date/Time 10/18/18 10:45 AM

Received by: KS

Relinquished by: CP

Date/Time 10/18/18 15:15

Received by: CP

Relinquished by: KS