



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB98720

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.0	80.6	80.6	10/19/2018	09:10
2	J5538-01		1	9	>100.0	>212.0	>212.0	10/19/2018	09:35
3	J5538-01DUP		1	9	>100.0	>212.0	>212.0	10/19/2018	09:55
4	J5538-11		1	7	>100.0	>212.0	>212.0	10/19/2018	10:20
5	J5590-10		1	8	>100.0	>212.0	>212.0	10/19/2018	10:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 98720

WorkList Name : fp101818-1

WorkList ID : 118181

Date : 10/18/2018 10:41:45 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/19/2018	Water	J5538-01	Flash Point	Cool 4 deg C	ENVI77		TKD-01	10/12/2018	1010A
10/19/2018	Water	J5538-11	Flash Point	Cool 4 deg C	ENVI77		TKDSLUDGE	10/12/2018	1010A
10/23/2018	Water	J5590-10	Flash Point	Cool 4 deg C	PSEG01	P63	LS-DRUM	10/18/2018	1010A

Date/Time 10/19/18 8:30 AM
 Received by: KS
 Relinquished by: CP

Date/Time 10/19/18 10:55 AM
 Received by: CP
 Relinquished by: KS