

LB 98721


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

36668



A= 0.0000e+000

B= 2.6663e-004

C= 2.2651e-002

Rho= 0.9992199

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.002	0.002	-78	0.000	-78				
0.2	0.200	0.204	0.004	680	0.0 %	680				
2.5	2.500	2.345	-0.155	8709	0.0 %	8709				
5.0	5.000	5.144	0.144	19207	0.0 %	19207				
7.5	7.500	7.706	0.206	28816	0.0 %	28816				
10.0	10.000	9.800	-0.200	36668	0.0 %	36668				

LB98721 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Oct 2018 12:11:50

Sample ID	Extended ID	μ Abs	Conc	Std Conc/Method	Units	Date	Type	Type
0.0 - 1	50	-78	-	0.00007470A	PPB	19 Oct 2018 12:16:00	S	Std
0.2 - 1	50.2	680	-	0.20007470A	PPB	19 Oct 2018 12:18:09	S	Std
2.5 - 1	52.5	8709	-	2.50007470A	PPB	19 Oct 2018 12:20:18	S	Std
5.0 - 1	55	19207	-	5.00007470A	PPB	19 Oct 2018 12:22:26	S	Std
7.5 - 1	57.5	28816	-	7.50007470A	PPB	19 Oct 2018 12:24:34	S	Std
10.0 - 1	510	36668	-	10.00007470A	PPB	19 Oct 2018 12:29:11	S	Std
ICV31 - 1	ICV31	15023	4.0283	-7470A	PPB	19 Oct 2018 12:31:51	U	SMPL
ICB31 - 1	ICB31	-160	-0.0200	-7470A	PPB	19 Oct 2018 12:33:57	U	SMPL
CCV44 - 1	CCV44	18958	5.0775	-7470A	PPB	19 Oct 2018 12:36:03	U	SMPL
CCB44 - 1	CCB44	-175	-0.0240	-7470A	PPB	19 Oct 2018 12:38:10	U	SMPL
CRA - 1	CRA	615	0.1866	-7470A	PPB	19 Oct 2018 12:40:17	U	SMPL
HighStd - 1	HighStd	36086	9.6444	-7470A	PPB	19 Oct 2018 12:42:23	U	SMPL
ChkStd - 1	ChkStd	26606	7.1167	-7470A	PPB	19 Oct 2018 12:44:30	U	SMPL
J5200-10 - 1	OR-01-101618-A	-105	-0.0053	-7470A	PPB	19 Oct 2018 12:46:38	U	SMPL
J5200-10DUP - 1	OR-01-101618-ADUP	6	0.0243	-7470A	PPB	19 Oct 2018 12:48:46	U	SMPL
J5200-10MS - 1	OR-01-101618-AMS	12232	3.2841	-7470A	PPB	19 Oct 2018 12:50:54	U	SMPL
J5200-10MSD - 1	OR-01-101618-AMSD	12754	3.4233	-7470A	PPB	19 Oct 2018 12:53:02	U	SMPL
J5200-12 - 1	OR-01-101618-B	-33	0.0139	-7470A	PPB	19 Oct 2018 12:55:10	U	SMPL
J5202-02 - 1	PL-03-101718-A	0	0.0227	-7470A	PPB	19 Oct 2018 12:57:19	U	SMPL
J5202-04 - 1	PL-03-101718-B	43	0.0341	-7470A	PPB	19 Oct 2018 12:59:26	U	SMPL
CCV45 - 1	CCV45	18995	5.0874	-7470A	PPB	19 Oct 2018 13:01:33	U	SMPL
CCB45 - 1	CCB45	-146	-0.0163	-7470A	PPB	19 Oct 2018 13:03:39	U	SMPL
J5202-06 - 1	PL-03-101718-C	-6	0.0211	-7470A	PPB	19 Oct 2018 13:05:46	U	SMPL
J5204-06 - 1	SU-03-101718-A	-14	0.0189	-7470A	PPB	19 Oct 2018 13:07:53	U	SMPL
J5204-08 - 1	SU-03-101718-B	-5	0.0213	-7470A	PPB	19 Oct 2018 13:10:00	U	SMPL
J5536-04 - 1	351-SOIL	5	0.0240	-7470A	PPB	19 Oct 2018 13:12:07	U	SMPL
J5542-01 - 1	OILY-DEBRIS	2	0.0232	-7470A	PPB	19 Oct 2018 13:14:15	U	SMPL
J5542-02 - 1	DPC-81753-20	-3	0.0219	-7470A	PPB	19 Oct 2018 13:16:22	U	SMPL
J5542-03 - 1	DPC-81753-19	6	0.0243	-7470A	PPB	19 Oct 2018 13:18:31	U	SMPL
J5549-01 - 1	CB-1018-S-TCLP-GRID-45	28	0.0301	-7470A	PPB	19 Oct 2018 13:20:39	U	SMPL
J5549-02 - 1	CB-1018-S-TCLP-GRID-50-2	18	0.0275	-7470A	PPB	19 Oct 2018 13:22:48	U	SMPL
J5579-04 - 1	7-1-ROLLOFFS	-2	0.0221	-7470A	PPB	19 Oct 2018 13:24:55	U	SMPL
CCV46 - 1	CCV46	18730	5.0167	-7470A	PPB	19 Oct 2018 13:27:02	U	SMPL
CCB46 - 1	CCB46	-147	-0.0165	-7470A	PPB	19 Oct 2018 13:29:09	U	SMPL
J5573-03 - 1	0692-KM-CA-AMD-COMP	621	0.1882	-7470A	PPB	19 Oct 2018 13:31:16	U	SMPL
J5573-03DUP - 1	0692-KM-CA-AMD-COMPDUP	675	0.2026	-7470A	PPB	19 Oct 2018 13:33:23	U	SMPL
J5573-03MS - 1	0692-KM-CA-AMD-COMPMS	14028	3.7630	-7470A	PPB	19 Oct 2018 13:35:30	U	SMPL
J5573-03MSD - 1	0692-KM-CA-AMD-COMPMSD	14075	3.7755	-7470A	PPB	19 Oct 2018 13:37:38	U	SMPL
J5586-01 - 1	LCS-5-WEST-DRAIN-PIPE	-118	-0.0088	-7470A	PPB	19 Oct 2018 13:39:45	U	SMPL
J5586-01DUP - 1	LCS-5-WEST-DRAIN-PIPEDUP	-73	0.0032	-7470A	PPB	19 Oct 2018 13:41:53	U	SMPL
J5586-01MS - 1	LCS-5-WEST-DRAIN-PIPEMS	12856	3.4505	-7470A	PPB	19 Oct 2018 13:44:02	U	SMPL
J5586-01MSD - 1	LCS-5-WEST-DRAIN-PIPEMSD	13375	3.5889	-7470A	PPB	19 Oct 2018 13:46:10	U	SMPL
J5588-01 - 1	MH-6	-143	-0.0155	-7470A	PPB	19 Oct 2018 13:48:18	U	SMPL
J5590-09 - 1	LS-DRUM	-109	-0.0064	-7470A	PPB	19 Oct 2018 13:50:25	U	SMPL
CCV47 - 1	CCV47	18557	4.9706	-7470A	PPB	19 Oct 2018 13:52:33	U	SMPL
CCB47 - 1	CCB47	-172	-0.0232	-7470A	PPB	19 Oct 2018 13:54:40	U	SMPL
J5200-10LX5 - 1		-76	0.0024	-7470A	PPB	19 Oct 2018 13:56:47	U	SMPL
J5200-10A - 1		13644	3.6606	-7470A	PPB	19 Oct 2018 13:58:54	U	SMPL
J5573-03LX5 - 1		-17	0.0181	-7470A	PPB	19 Oct 2018 14:01:01	U	SMPL
J5573-03A - 1		15365	4.1195	-7470A	PPB	19 Oct 2018 14:03:09	U	SMPL
J5586-01LX5 - 1		-144	-0.0157	-7470A	PPB	19 Oct 2018 14:05:16	U	SMPL
J5586-01A - 1		13725	3.6822	-7470A	PPB	19 Oct 2018 14:07:24	U	SMPL
CCV48 - 1	CCV48	18968	5.0802	-7470A	PPB	19 Oct 2018 14:09:32	U	SMPL
CCB48 - 1	CCB48	-165	-0.0213	-7470A	PPB	19 Oct 2018 14:11:41	U	SMPL
PB113989BL - 1	PBW	1	0.0229	-7470A	PPB	19 Oct 2018 14:13:49	U	SMPL
PB113989BS - 1	LCSW	13878	3.7230	-7470A	PPB	19 Oct 2018 14:15:56	U	SMPL
PB113989TB - 1	PB113907TB	-42	0.0115	-7470A	PPB	19 Oct 2018 14:18:04	U	SMPL
PB113990BL - 1	PBW	18	0.0275	-7470A	PPB	19 Oct 2018 14:20:11	U	SMPL
PB113990BS - 1	LCSW	13156	3.5305	-7470A	PPB	19 Oct 2018 14:22:18	U	SMPL
PB113990TB - 1	PB113939TB	-151	-0.0176	-7470A	PPB	19 Oct 2018 14:24:25	U	SMPL
PB113991BL - 1	PBW	-1	0.0224	-7470A	PPB	19 Oct 2018 14:26:33	U	SMPL
PB113991BS - 1	LCSW	15117	4.0534	-7470A	PPB	19 Oct 2018 14:28:41	U	SMPL
CCV49 - 1	CCV49	18690	5.0060	-7470A	PPB	19 Oct 2018 14:30:49	U	SMPL
CCB49 - 1	CCB49	-167	-0.0219	-7470A	PPB	19 Oct 2018 14:32:57	U	SMPL