



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB98746

Analyst By : sweta

Supervisor Review By : JIGNESH

Slope : 99.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.8	4.01	10/19/2018	16:50
2	CAL2	1	Water			22.7	7.00	10/19/2018	16:52
3	CAL3	1	Water			22.8	10.04	10/19/2018	16:54
4	ICV	1	Water			22.6	7.01	10/19/2018	16:56
5	CCV1	1	Water			22.5	2.00	10/19/2018	16:58
6	J5604-06	1	Water			17.3	7.31	10/19/2018	17:00
7	J5604-06DUP	1	Water			17.4	7.32	10/19/2018	17:02
8	J5604-09	1	Water			17.8	7.04	10/19/2018	17:04
9	CCV2	1	Water			22.6	12.00	10/19/2018	17:20

WORKLIST(Hardcopy Internal Chain)

UB98746

WorkList Name : corrosivity101918-1

WorkList ID : 118275

Date : 10/19/2018 4:01:19 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/19/2018	Water	J5604-06	Corrosivity	Cool 4 deg C	PSEG01	E22	MS-VB-29736-BOX-2	10/19/2018	9040C
10/19/2018	Water	J5604-09	Corrosivity	Cool 4 deg C	PSEG01	E22	MS-VB-25868-BOX-1	10/19/2018	9040C

Date/Time 10/19/18 4:40
 Received by: SP
 Relinquished by: CP

Date/Time 10/19/18 5:30
 Received by: CP
 Relinquished by: SP