



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB98749

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP68919
Turbidity std, 40NTU	WP68913
turbidity std 5 NTU	WP68916
turbidity std 20 NTU	WP68914
Turbidity std 1 NTU	WP68917
turbidity std 0 NTU	WP68918
turbidity std 10 NTU	WP68915
Turbidity Std, 200NTU	WP68935
10 NTU Standard 500 ml	W2297

Intercept: 0.8226

Slope: 0.8929

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.280	10/20/2018	11:47
2	CAL2	1	1	0.992	10/20/2018	11:51
3	CAL3	5	1	4.860	10/20/2018	11:55
4	CAL4	10	1	9.680	10/20/2018	11:58
5	CAL5	20	1	18.790	10/20/2018	12:02
6	CAL6	40	1	38.600	10/20/2018	12:07
7	CAL7	200	1	179.000	10/19/2018	12:11



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB98749

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.350	9.550	10/20/2018	12:20
2	ICB		1	0.260	-0.630	10/20/2018	12:24
3	CCV1	10	1	9.580	9.808	10/20/2018	12:28
4	CCB1		1	0.000	-0.921	10/20/2018	12:34
5	LB98749BL		1	0.257	-0.633	10/20/2018	12:37
6	J5609-01		1	88.300	97.970	10/20/2018	12:41
7	J5609-01DUP		1	88.100	97.746	10/20/2018	12:44
8	CCV2	10	1	9.610	9.841	10/20/2018	12:49
9	CCB2		1	0.259	-0.631	10/20/2018	12:53

WORKLIST(Hardcopy Internal Chain)

CB 98449

WorkList Name : tur102018

WorkList ID : 118306

Date : 10/20/2018 10:03:51 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/21/2018	Water	J5609-01	Turbidity	Cool 4 deg C	LIRO01	E22	UDS-CH-4-01	10/19/2018	SM2130 B

Date/Time 10-20-18 10:45
 Received by: AP
 Relinquished by: AP

Date/Time 10-20-18 13:05
 Received by: AP
 Relinquished by: AP