


7470A

UB98753 INSTRUMENT ID: 101

 μ Abs.:

38540



Linear

A= 0.0000e+000

B= 2.6005e-004

C= 5.4866e-002

Rho= 0.9998788

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.045	0.045	-37	0.000	-37				
0.2	0.200	0.246	0.046	735	0.0 %	735				
2.5	2.500	2.440	-0.060	9170	0.0 %	9170				
5.0	5.000	4.936	-0.064	18770	0.0 %	18770				
7.5	7.500	7.456	-0.044	28461	0.0 %	28461				
10.0	10.000	10.077	0.077	38540	0.0 %	38540				

LB98753 INSTRUMENT ID:CV1

Method: 7470A

Operator: Admin

Date of Analysis: 22 Oct 2018 13:27:03

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	-37	-	0.0000	7470A	PPB	22 Oct 2018 13:27:10	S	Std
0.2 - 1	50.2	735	-	0.2000	7470A	PPB	22 Oct 2018 13:29:19	S	Std
2.5 - 1	52.5	9170	-	2.5000	7470A	PPB	22 Oct 2018 13:31:28	S	Std
5.0 - 1	55.0	18770	-	5.0000	7470A	PPB	22 Oct 2018 13:33:36	S	Std
7.5 - 1	57.5	28461	-	7.5000	7470A	PPB	22 Oct 2018 13:35:44	S	Std
10.0 - 1	59.0	38540	-	10.0000	7470A	PPB	22 Oct 2018 13:37:51	S	Std
ICV32 - 1	ICV32	14691	3.8752	-	7470A	PPB	22 Oct 2018 13:55:04	U	SMPL
ICB32 - 1	ICB32	-81	0.0338	-	7470A	PPB	22 Oct 2018 13:57:10	U	SMPL
CCV50 - 1	CCV50	18578	4.8861	-	7470A	PPB	22 Oct 2018 13:59:17	U	SMPL
CCB50 - 1	CCB50	-81	0.0338	-	7470A	PPB	22 Oct 2018 14:01:23	U	SMPL
CRA - 1	CRA	732	0.2452	-	7470A	PPB	22 Oct 2018 14:03:30	U	SMPL
HighStd - 1	HighStd	38732	10.1271	-	7470A	PPB	22 Oct 2018 14:05:36	U	SMPL
ChkStd - 1	ChkStd	27729	7.2658	-	7470A	PPB	22 Oct 2018 14:07:43	U	SMPL
J5538-03 - 1	WWT1-6	251	0.1201	-	7470A	PPB	22 Oct 2018 14:09:51	U	SMPL
J5538-03DUP - 1	WWT1-6DUP	71	0.0733	-	7470A	PPB	22 Oct 2018 14:11:59	U	SMPL
J5538-03MS - 1	WWT1-6MS	13701	3.6178	-	7470A	PPB	22 Oct 2018 14:14:06	U	SMPL
J5538-03MSD - 1	WWT1-6MSD	13033	3.4441	-	7470A	PPB	22 Oct 2018 14:16:14	U	SMPL
J5538-16 - 1	NL_7	-18	0.0502	-	7470A	PPB	22 Oct 2018 14:18:23	U	SMPL
J5538-17 - 1	NL-7	556	0.1995	-	7470A	PPB	22 Oct 2018 14:20:31	U	SMPL
J5538-18 - 1	BL01	82	0.0762	-	7470A	PPB	22 Oct 2018 14:22:39	U	SMPL
CCV51 - 1	CCV51	18387	4.8364	-	7470A	PPB	22 Oct 2018 14:24:45	U	SMPL
CCB51 - 1	CCB51	-55	0.0406	-	7470A	PPB	22 Oct 2018 14:26:52	U	SMPL
J5538-22 - 1	CNL01	352	0.1464	-	7470A	PPB	22 Oct 2018 14:28:59	U	SMPL
J5538-23 - 1	BS01	2	0.0554	-	7470A	PPB	22 Oct 2018 14:31:05	U	SMPL
J5538-24 - 1	CNS01	171	0.0993	-	7470A	PPB	22 Oct 2018 14:33:12	U	SMPL
J5538-03LX5 - 1		-57	0.0400	-	7470A	PPB	22 Oct 2018 14:35:19	U	SMPL
J5538-03A - 1		14578	3.8459	-	7470A	PPB	22 Oct 2018 14:37:26	U	SMPL
PB114029BL - 1	PBS0129	-2	0.0543	-	7470A	PPB	22 Oct 2018 14:39:34	U	SMPL
PB114029BS - 1	LCS0129	12247	3.2397	-	7470A	PPB	22 Oct 2018 14:41:42	U	SMPL
PB114029TB - 1	LEB0129	130	0.0887	-	7470A	PPB	22 Oct 2018 14:43:50	U	SMPL
PB114030BL - 1	PBS0130	73	0.0738	-	7470A	PPB	22 Oct 2018 14:45:59	U	SMPL
PB114030BS - 1	LCS0130	13440	3.5499	-	7470A	PPB	22 Oct 2018 14:48:05	U	SMPL
CCV52 - 1	CCV52	18903	4.9706	-	7470A	PPB	22 Oct 2018 14:50:12	U	SMPL
CCB52 - 1	CCB52	25	0.0614	-	7470A	PPB	22 Oct 2018 14:52:19	U	SMPL
PB114030TB - 1	LEB0130	163	0.0973	-	7470A	PPB	22 Oct 2018 14:54:26	U	SMPL
J5195-04 - 1	JC-02-101818-A	70	0.0731	-	7470A	PPB	22 Oct 2018 14:56:33	U	SMPL
J5195-04DUP - 1	JC-02-101818-ADUP	98	0.0804	-	7470A	PPB	22 Oct 2018 14:58:40	U	SMPL
J5195-04MS - 1	JC-02-101818-AMS	12400	3.2795	-	7470A	PPB	22 Oct 2018 15:00:47	U	SMPL
J5195-04MSD - 1	JC-02-101818-AMSD	13131	3.4696	-	7470A	PPB	22 Oct 2018 15:02:54	U	SMPL
J5195-06 - 1	JC-02-101818-B	74	0.0741	-	7470A	PPB	22 Oct 2018 15:05:02	U	SMPL
J5590-04 - 1	LS-SOIL-1	137	0.0905	-	7470A	PPB	22 Oct 2018 15:07:10	U	SMPL
J5590-08 - 1	LS-SOIL-2	145	0.0926	-	7470A	PPB	22 Oct 2018 15:09:18	U	SMPL
J5195-04LX5 - 1		95	0.0796	-	7470A	PPB	22 Oct 2018 15:11:27	U	SMPL
J5195-04A - 1		15123	3.9876	-	7470A	PPB	22 Oct 2018 15:13:34	U	SMPL
CCV53 - 1	CCV53	19154	5.0358	-	7470A	PPB	22 Oct 2018 15:15:41	U	SMPL
CCB53 - 1	CCB53	63	0.0712	-	7470A	PPB	22 Oct 2018 15:17:48	U	SMPL