



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : sweta

Parameter: pH

Supervisor Review By : JIGNESH

Run Number: LB98763

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.8	4.01	10/22/2018	12:35
2	CAL2	1	Water			22.7	7.00	10/22/2018	12:37
3	CAL3	1	Water			22.8	10.04	10/22/2018	12:39
4	ICV	1	Water			22.6	7.01	10/22/2018	12:41
5	CCV1	1	Water			22.8	2.00	10/22/2018	12:43
6	J5617-01	1	Water			15.3	6.54	10/22/2018	12:45
7	J5617-01DUP	1	Water			15.4	6.55	10/22/2018	12:47
8	CCV2	1	Water			22.7	12.00	10/22/2018	12:49

WORKLIST(Hardcopy Internal Chain)

1398763

WorkList Name : pH102218

WorkList ID : 118314

Date : 10/22/2018 12:23:15 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/20/2018	Water	J5617-01 A	pH	Cool 4 deg C	WARR01	E61	PANTRY-COOLER	10/20/2018	SM 4500-H B

Date/Time 10/22/18 12:25
 Received by: SP
 Relinquished by: SP

Date/Time 10/22/18 12:55
 Received by: SP
 Relinquished by: SP