


7470A

 μ Abs.:

33943



Linear

A= 0.0000e+000

B= 2.9561e-004

C= -3.3369e-002

Rho= 0.9999922

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.010	-0.010	80	0.000	80				
0.2	0.200	0.205	0.005	808	0.0 %	808				
2.5	2.500	2.493	-0.007	8546	0.0 %	8546				
5.0	5.000	5.028	0.028	17123	0.0 %	17123				
7.5	7.500	7.483	-0.017	25426	0.0 %	25426				
10.0	10.000	10.000	0.000	33943	0.0 %	33943				

LB98765 INSTRUMENT ID:CV1

Method: 7470A

Operator: Admin

Date of Analysis: 23 Oct 2018 12:55:06

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	80	-	0.0000	7470A	PPB	23 Oct 2018 13:12:29	S	Std
0.2 - 1	50.2	808	-	0.2000	7470A	PPB	23 Oct 2018 13:14:38	S	Std
2.5 - 1	52.5	8546	-	2.5000	7470A	PPB	23 Oct 2018 13:16:47	S	Std
5.0 - 1	55.0	17123	-	5.0000	7470A	PPB	23 Oct 2018 13:18:55	S	Std
7.5 - 1	57.5	25426	-	7.5000	7470A	PPB	23 Oct 2018 13:21:03	S	Std
10.0 - 1	59.0	33943	-	10.0000	7470A	PPB	23 Oct 2018 13:23:11	S	Std
ICV33 - 1	ICV33	14510	4.2559	-	7470A	PPB	23 Oct 2018 13:32:00	U	SMPL
ICB33 - 1	ICB33	39	-0.0218	-	7470A	PPB	23 Oct 2018 13:34:06	U	SMPL
CCV54 - 1	CCV54	18492	5.4330	-	7470A	PPB	23 Oct 2018 13:36:13	U	SMPL
CCB54 - 1	CCB54	35	-0.0230	-	7470A	PPB	23 Oct 2018 13:38:19	U	SMPL
CRA - 1	CRA	842	0.2155	-	7470A	PPB	23 Oct 2018 13:40:28	U	SMPL
HighStd - 1	HighStd	39120	11.5307	-	7470A	PPB	23 Oct 2018 13:42:35	U	SMPL
ChkStd - 1	ChkStd	25829	7.6018	-	7470A	PPB	23 Oct 2018 13:52:18	U	SMPL
J5590-10 - 1	LS-DRUM	317	0.0603	-	7470A	PPB	23 Oct 2018 13:54:25	U	SMPL
J5590-10DUP - 1	LS-DRUMDUP	383	0.0798	-	7470A	PPB	23 Oct 2018 13:56:32	U	SMPL
J5590-10MS - 1	LS-DRUMMS	13967	4.0954	-	7470A	PPB	23 Oct 2018 13:58:40	U	SMPL
J5590-10MSD - 1	LS-DRUMMSD	14093	4.1326	-	7470A	PPB	23 Oct 2018 14:00:48	U	SMPL
J5604-06 - 1	MS-VB-29736-BOX-2	165	0.0154	-	7470A	PPB	23 Oct 2018 14:02:56	U	SMPL
J5604-09 - 1	MS-VB-25868-BOX-1	231	0.0349	-	7470A	PPB	23 Oct 2018 14:05:04	U	SMPL
J5197-04 - 1	HD-02-101918	174	0.0181	-	7470A	PPB	23 Oct 2018 14:07:11	U	SMPL
CCV55 - 1	CCV55	18689	5.4912	-	7470A	PPB	23 Oct 2018 14:09:18	U	SMPL
CCB55 - 1	CCB55	-81	-0.0573	-	7470A	PPB	23 Oct 2018 14:11:24	U	SMPL
J5197-04DUP - 1	HD-02-101918DUP	361	0.0733	-	7470A	PPB	23 Oct 2018 14:13:31	U	SMPL
J5197-04MS - 1	HD-02-101918MS	14481	4.2473	-	7470A	PPB	23 Oct 2018 14:15:38	U	SMPL
J5197-04MSD - 1	HD-02-101918MSD	13882	4.0702	-	7470A	PPB	23 Oct 2018 14:17:45	U	SMPL
J5604-02 - 1	MS-VB-29736-BOX-2	586	0.1399	-	7470A	PPB	23 Oct 2018 14:19:52	U	SMPL
J5604-04 - 1	MS-VB-25868-BOX-1	148	0.0104	-	7470A	PPB	23 Oct 2018 14:21:59	U	SMPL
J5605-06 - 1	TP-A	194	0.0240	-	7470A	PPB	23 Oct 2018 14:24:07	U	SMPL
J5605-12 - 1	TP-B	148	0.0104	-	7470A	PPB	23 Oct 2018 14:26:15	U	SMPL
J5606-03 - 1	SOIL-PILE	248	0.0399	-	7470A	PPB	23 Oct 2018 14:28:23	U	SMPL
PB114074BL - 1	PB114074BL	91	-0.0065	-	7470A	PPB	23 Oct 2018 14:30:32	U	SMPL
PB114074BS - 1	PB114074BS	12453	3.6478	-	7470A	PPB	23 Oct 2018 14:32:39	U	SMPL
CCV56 - 1	CCV56	18569	5.4557	-	7470A	PPB	23 Oct 2018 14:34:45	U	SMPL
CCB56 - 1	CCB56	24	-0.0263	-	7470A	PPB	23 Oct 2018 14:36:52	U	SMPL
PB114074TB - 1	PB114074TB	139	0.0077	-	7470A	PPB	23 Oct 2018 14:38:59	U	SMPL
PB114073BL - 1	PB114073BL	86	-0.0079	-	7470A	PPB	23 Oct 2018 14:41:06	U	SMPL
PB114073BS - 1	PB114073BS	12215	3.5775	-	7470A	PPB	23 Oct 2018 14:43:13	U	SMPL
PB114073TB - 1	PB114073TB	129	0.0048	-	7470A	PPB	23 Oct 2018 14:45:20	U	SMPL
J5590-10LX5 - 1		129	0.0048	-	7470A	PPB	23 Oct 2018 14:47:27	U	SMPL
J5590-10A - 1		14982	4.3954	-	7470A	PPB	23 Oct 2018 14:49:35	U	SMPL
J5197-04LX5 - 1		72	-0.0121	-	7470A	PPB	23 Oct 2018 14:51:42	U	SMPL
J5197-04A - 1		15498	4.5479	-	7470A	PPB	23 Oct 2018 14:53:51	U	SMPL
CCV57 - 1	CCV57	18230	5.3555	-	7470A	PPB	23 Oct 2018 15:14:21	U	SMPL
CCB57 - 1	CCB57	10	-0.0304	-	7470A	PPB	23 Oct 2018 15:16:28	U	SMPL