



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: jignesh

Date: 10/23/2018

TEMP1 IN: 103 10/23/2018 07:15 TEMP OUT: 103 10/23/2018 08:15

TEMP2 IN: 103 10/23/2018 08:35 TEMP OUT: 103 10/23/2018 09:35

TEMP3 IN: 104 10/23/2018 10:45 TEMP OUT: 104 10/23/2018 12:15

TEMP4 IN: 104 10/23/2018 12:35 TEMP OUT: 104 10/23/2018 14:05

Run Number: LB98771

BalanceID: wc-sc-3

OvenID: wc-oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB98771BL	LB98771BL	2.3698	2.3698	100	2.3698	2.3698	2.3698	0.0000	0
2	LB98771BS	LB98771BS	1.4025	1.4025	100	1.4573	1.4573	1.4573	0.0548	548
3	J5602-01	RAV-01D	1.3986	1.3986	800	1.4331	1.4331	1.4331	0.0345	43.1
4	J5602-02	RAV-01G	1.3956	1.3956	800	1.3957	1.3957	1.3957	0.0001	0.1
5	J5602-04	RAV-RAINEY-AFTER	1.3932	1.3932	400	1.3947	1.3947	1.3947	0.0015	3.8
6	J5602-04DUP	RAV-RAINEY-AFTERDUP	1.3879	1.3879	400	1.3892	1.3892	1.3892	0.0013	3.3
7	J5623-01	101918-SIDI-F-D-B	1.3949	1.3949	400	1.4127	1.4127	1.4127	0.0178	44.5
8	J5623-02	101918-SIDI-F-D-M	1.4019	1.4019	400	1.4111	1.4111	1.4111	0.0092	23
9	J5623-03	101918-SIDI-F-D-S	1.3924	1.3924	400	1.3972	1.3972	1.3972	0.0048	12
10	J5623-04	101918-SIDI-F-U-B	1.4006	1.4006	400	1.4132	1.4132	1.4132	0.0126	31.5
11	J5623-05	101918-SIDI-F-U-M	1.4017	1.4017	400	1.4085	1.4085	1.4085	0.0068	17
12	J5623-06	101918-SIDI-F-U-S	1.3998	1.3998	400	1.4041	1.4041	1.4041	0.0043	10.8
13	J5623-07	101918-TIPC-E-D-M	1.4026	1.4026	400	1.4107	1.4107	1.4107	0.0081	20.2
14	J5623-08	101918-TIPC-E-U-M	1.4043	1.4043	400	1.4116	1.4116	1.4116	0.0073	18.3
15	J5623-09	101918-DBRI-E-D-B	1.3991	1.3991	400	1.4126	1.4126	1.4126	0.0135	33.8

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: jignesh

Date: 10/23/2018

TEMP1 IN: 103 10/23/2018 07:15 TEMP OUT: 103 10/23/2018 08:15

TEMP2 IN: 103 10/23/2018 08:35 TEMP OUT: 103 10/23/2018 09:35

TEMP3 IN: 104 10/23/2018 10:45 TEMP OUT: 104 10/23/2018 12:15

TEMP4 IN: 104 10/23/2018 12:35 TEMP OUT: 104 10/23/2018 14:05

Run Number: LB98771

BalanceID: wc-sc-3

OvenID: wc-oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	J5623-10	101918-DBRI-E-D-M	1.3959	1.3959	400	1.4016	1.4016	1.4016	0.0057	14.3
17	J5623-11	101918-DBRI-E-D-S	1.3928	1.3928	400	1.3982	1.3982	1.3982	0.0054	13.5
18	J5623-12	101918-DBRI-E-U-B	1.4041	1.4041	400	1.4158	1.4158	1.4158	0.0117	29.3
19	J5623-13	101918-DBRI-E-U-M	1.3987	1.3987	400	1.4046	1.4046	1.4046	0.0059	14.8
20	J5623-14	101918-DBRI-E-U-S	1.4014	1.4014	400	1.4063	1.4063	1.4063	0.0049	12.3

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS J5623

WorkList ID : 118331

Date : 10/23/2018 7:10:15 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/25/2018	Water	J5602-01- 1	TSS	Cool 4 deg C	FOST03	E21	RAV-01D	10/18/2018	SM2540 D
10/25/2018	Water	J5602-02- 1	TSS	Cool 4 deg C	FOST03	E21	RAV-01G	10/18/2018	SM2540 D
10/25/2018	Water	J5602-04- 1	TSS	Cool 4 deg C	FOST03	E21	RAV-RAINEY-AFTER	10/18/2018	SM2540 D
10/26/2018	Water	J5623-01- 1	TSS	Cool 4 deg C	HDR103	E32	101918-SIDI-F-D-B	10/19/2018	SM2540 D
10/26/2018	Water	J5623-02- 1	TSS	Cool 4 deg C	HDR103	E32	101918-SIDI-F-D-M	10/19/2018	SM2540 D
10/26/2018	Water	J5623-03- 1	TSS	Cool 4 deg C	HDR103	E32	101918-SIDI-F-D-S	10/19/2018	SM2540 D
10/26/2018	Water	J5623-04- 1	TSS	Cool 4 deg C	HDR103	E32	101918-SIDI-F-U-B	10/19/2018	SM2540 D
10/26/2018	Water	J5623-05- 1	TSS	Cool 4 deg C	HDR103	E32	101918-SIDI-F-U-M	10/19/2018	SM2540 D
10/26/2018	Water	J5623-06- 1	TSS	Cool 4 deg C	HDR103	E32	101918-SIDI-F-U-S	10/19/2018	SM2540 D
10/24/2018	Water	J5623-07- 1	TSS	Cool 4 deg C	HDR103	E32	101918-TIPC-E-D-M	10/19/2018	SM2540 D
10/24/2018	Water	J5623-08- 1	TSS	Cool 4 deg C	HDR103	E32	101918-TIPC-E-U-M	10/19/2018	SM2540 D
10/26/2018	Water	J5623-09- 1	TSS	Cool 4 deg C	HDR103	E32	101918-DBRI-E-D-B	10/19/2018	SM2540 D
10/26/2018	Water	J5623-10- 1	TSS	Cool 4 deg C	HDR103	E32	101918-DBRI-E-D-M	10/19/2018	SM2540 D
10/26/2018	Water	J5623-11- 1	TSS	Cool 4 deg C	HDR103	E32	101918-DBRI-E-D-S	10/19/2018	SM2540 D
10/26/2018	Water	J5623-12- 1	TSS	Cool 4 deg C	HDR103	E32	101918-DBRI-E-U-B	10/19/2018	SM2540 D
10/26/2018	Water	J5623-13- 1	TSS	Cool 4 deg C	HDR103	E32	101918-DBRI-E-U-M	10/19/2018	SM2540 D
10/26/2018	Water	J5623-14- 1	TSS	Cool 4 deg C	HDR103	E32	101918-DBRI-E-U-S	10/19/2018	SM2540 D

Date/Time 10-23-18 7:15 AM
 Received by: JA
 Relinquished by: (signature)

Date/Time 10-23-18 10:40 AM
 Received by: (signature)
 Relinquished by: JA