



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: pH

Supervisor Review By : JIGNESH

Run Number: LB98772

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.9	4.01	10/22/2018	16:50
2	CAL2	1	Water			22.8	7.00	10/22/2018	16:52
3	CAL3	1	Water			22.8	10.04	10/22/2018	16:54
4	ICV	1	Water			22.7	7.01	10/22/2018	16:56
5	CCV1	1	Water			22.6	2.00	10/22/2018	16:58
6	J5621-05	1	Solid	20.02	20	21.4	6.32	10/22/2018	17:00
7	J5621-05DUP	1	Solid	20.04	20	21.5	6.33	10/22/2018	17:02
8	J5621-06	1	Solid	20.02	20	21.3	5.96	10/22/2018	17:04
9	J5621-07	1	Solid	20.03	20	21.1	5.82	10/22/2018	17:06
10	J5624-05	1	Solid	20.03	20	21.2	9.96	10/22/2018	17:08
11	J5624-06	1	Solid	20.04	20	20.5	8.72	10/22/2018	17:10
12	J5624-07	1	Solid	20.02	20	20.7	10.37	10/22/2018	17:12
13	J5624-08	1	Solid	20.04	20	21.2	9.73	10/22/2018	17:14
14	J5624-09	1	Solid	20.02	20	20.8	9.24	10/22/2018	17:16
15	J5626-01	1	Solid	20.02	20	21.3	9.08	10/22/2018	17:18
16	CCV2	1	Water			22.7	12.00	10/22/2018	17:20

WORKLIST(Hardcopy Internal Chain)

LB 98772

WorkList Name : pH102218-1

WorkList ID : 118318

Date : 10/22/2018 3:00:04 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/19/2018	Solid	J5621-05	pH	Cool 4 deg C	BEMS01	E52	SA-SB-01A	10/19/2018	9045D
10/19/2018	Solid	J5621-06	pH	Cool 4 deg C	BEMS01	E52	SA-SB-01B	10/19/2018	9045D
10/19/2018	Solid	J5621-07	pH	Cool 4 deg C	BEMS01	E52	SA-SB-01C	10/19/2018	9045D
10/19/2018	Solid	J5624-05	pH	Cool 4 deg C	BEMS01	E52	PA-SB-03A	10/19/2018	9045D
10/19/2018	Solid	J5624-06	pH	Cool 4 deg C	BEMS01	E52	PA-SB-03B	10/19/2018	9045D
10/19/2018	Solid	J5624-07	pH	Cool 4 deg C	BEMS01	E52	PA-SB-02A	10/19/2018	9045D
10/19/2018	Solid	J5624-08	pH	Cool 4 deg C	BEMS01	E52	PA-SB-02B	10/19/2018	9045D
10/19/2018	Solid	J5624-09	pH	Cool 4 deg C	BEMS01	E52	PA-SB-01A	10/19/2018	9045D
10/19/2018	Solid	J5626-01	pH	Cool 4 deg C	BEMS01	E52	PA-SB-01B	10/19/2018	9045D

Date/Time 10/22/18 4:35

Received by: SP

Relinquished by: CP

Date/Time 10/22/18 5:45

Received by: CP

Relinquished by: JSP