



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : sweta

Parameter: pH

Supervisor Review By : JIGNESH

Run Number: LB98782

Slope : 99.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			23.4	4.01	10/23/2018	11:22
2	CAL2	1	Water			23.2	7.00	10/23/2018	11:24
3	CAL3	1	Water			22.9	10.04	10/23/2018	11:26
4	ICV	1	Water			23.2	7.00	10/23/2018	11:28
5	CCV1	1	Water			23.2	2.00	10/23/2018	11:30
6	J5634-01	1	Water			20.8	6.84	10/23/2018	11:32
7	J5634-01DUP	1	Water			20.9	6.85	10/23/2018	11:34
8	CCV2	1	Water			23.0	12.00	10/23/2018	11:36

WORKLIST(Hardcopy Internal Chain)

LB 98782

WorkList Name : pH102318

WorkList ID : 118360

Date : 10/23/2018 11:09:00 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/19/2018	Water	J5634-01	A pH	Cool 4 deg C	STER14	E31	JS-SYSTEM	10/19/2018	SM 4500-H B

Date/Time 10/23/18 11:15
 Received by: SP
 Relinquished by: SP

Date/Time 10/23/18 11:40
 Received by: CP
 Relinquished by: SP