



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB98789

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP68999
Turbidity std, 40NTU	WP68993
turbidity std 5 NTU	WP68996
turbidity std 20 NTU	WP68994
Turbidity std 1 NTU	WP68997
turbidity std 0 NTU	WP68998
turbidity std 10 NTU	WP68995

Intercept: 0.048

Slope: 0.9614

Regression: 0.9997

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.280	10/22/2018	08:30
2	CAL2	1	1	0.991	10/22/2018	08:33
3	CAL3	5	1	4.860	10/22/2018	08:36
4	CAL4	10	1	9.820	10/22/2018	08:40
5	CAL5	20	1	18.600	10/22/2018	08:43
6	CAL6	40	1	38.800	10/22/2018	08:46



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB98789

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.320	9.644	10/22/2018	08:49
2	ICB		1	0.260	0.221	10/22/2018	08:53
3	CCV1	10	1	9.460	9.790	10/22/2018	08:57
4	CCB1		1	0.265	0.226	10/22/2018	09:00
5	LB98789BL		1	0.270	0.231	10/22/2018	09:03
6	J5617-01		1	0.522	0.493	10/22/2018	09:06
7	J5617-01DUP		1	0.520	0.491	10/22/2018	09:09
8	CCV2	10	1	9.520	9.852	10/22/2018	09:13
9	CCB2		1	0.278	0.239	10/22/2018	09:16

LD 98789

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TUR102218

WorkList ID : 118372

Date : 10/22/2018 8:06:27 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/22/2018	Water	J5617-01	Turbidity	Cool 4 deg C	WARR01	E61	PANTRY-COOLER	10/20/2018	SM2130 B

Date/Time 10-22-18 8:20
Received by: AC
Relinquished by: CA

Date/Time 10-22-18 10:00
Received by: AC
Relinquished by: AC