



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB98802

Analyst By : sweta

Supervisor Review By : JIGNESH

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			23.4	4.01	10/23/2018	17:40
2	CAL2	1	Water			23.2	7.00	10/23/2018	17:42
3	CAL3	1	Water			22.9	10.04	10/23/2018	17:44
4	ICV	1	Water			23.2	7.01	10/23/2018	17:46
5	CCV1	1	Water			23.2	2.00	10/23/2018	17:48
6	J5639-14	1	Water			15.2	7.48	10/23/2018	17:50
7	J5639-14DUP	1	Water			15.3	7.49	10/23/2018	17:52
8	CCV2	1	Water			22.9	12.00	10/23/2018	18:14

WORKLIST(Hardcopy Internal Chain)

LB 98802

WorkList Name : corrosivity102318-1

WorkList ID : 118376

Date : 10/23/2018 5:00:08 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/23/2018	Water	J5639-14	Corrosivity	Cool 4 deg C	PSEG01	P23	BS-DRUM	10/23/2018	9040C

Date/Time 10/23/18 5:30
 Received by: SP
 Relinquished by: CP

Date/Time 10/23/18 6:35
 Received by: CP
 Relinquished by: SP