

7470A


LB 98806

INSTRUMENT ID: CVI

Linear

 μ Abs.:

37187



A= 0.0000e+000

B= 2.6717e-004

C= -1.8882e-002

Rho= 0.9998489

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.009	0.009	106	0.000	106				
0.2	0.200	0.190	-0.010	780	0.0 %	780				
2.5	2.500	2.472	-0.028	9323	0.0 %	9323				
5.0	5.000	4.985	-0.015	18728	0.0 %	18728				
7.5	7.500	7.628	0.128	28623	0.0 %	28623				
10.0	10.000	9.916	-0.084	37187	0.0 %	37187				

LB98806 INSTRUMENT ID:CV1

Method: 7470A

Operator: Admin

Date of Analysis: 24 Oct 2018 12:41:02

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD-0	106	-	0.00007470A	PPB		24 Oct 2018 12:43:12	S	Std
0.2 - 1	SD-2	780	-	0.20007470A	PPB		24 Oct 2018 12:45:21	S	Std
2.5 - 1	SD-5	9323	-	2.50007470A	PPB		24 Oct 2018 12:47:30	S	Std
5.0 - 1	SD-10	18728	-	5.00007470A	PPB		24 Oct 2018 12:49:38	S	Std
7.5 - 1	SD-15	28623	-	7.50007470A	PPB		24 Oct 2018 12:51:46	S	Std
10.0 - 1	SD-20	37187	-	10.00007470A	PPB		24 Oct 2018 12:53:56	S	Std
ICV36 - 1	ICV36	15069	4.0071	-7470A	PPB		24 Oct 2018 12:58:34	U	SMPL
ICB36 - 1	ICB36	78	0.0020	-7470A	PPB		24 Oct 2018 13:00:40	U	SMPL
CCV64 - 1	CCV64	18846	5.0162	-7470A	PPB		24 Oct 2018 13:02:46	U	SMPL
CCB64 - 1	CCB64	84	0.0036	-7470A	PPB		24 Oct 2018 13:04:53	U	SMPL
CRA - 1	CRA	793	0.1930	-7470A	PPB		24 Oct 2018 13:06:59	U	SMPL
HighStd - 1	HighStd	37313	9.9499	-7470A	PPB		24 Oct 2018 13:09:06	U	SMPL
ChkStd - 1	ChkStd	28724	7.6552	-7470A	PPB		24 Oct 2018 13:19:21	U	SMPL
PB114106BL - 1	PB114106BL	100	0.0078	-7470A	PPB		24 Oct 2018 13:21:28	U	SMPL
PB114106BS - 1	PB114106BS	13786	3.6643	-7470A	PPB		24 Oct 2018 13:23:36	U	SMPL
J5604-05 - 1	MS-VB-29736-BOX-2	7683	2.0338	-7470A	PPB		24 Oct 2018 13:25:44	U	SMPL
J5604-08 - 1	MS-VB-25868-BOX-1	20542	5.4693	-7470A	PPB		24 Oct 2018 13:27:52	U	SMPL
J5621-01 - 1	SA-SW-01	-66	-0.0365	-7470A	PPB		24 Oct 2018 13:30:00	U	SMPL
J5621-02 - 1	SA-SW-02	7	-0.0170	-7470A	PPB		24 Oct 2018 13:32:11	U	SMPL
J5621-08 - 1	SA-TWP-01	12881	3.4225	-7470A	PPB		24 Oct 2018 13:34:18	U	SMPL
CCV65 - 1	CCV65	18607	4.9523	-7470A	PPB		24 Oct 2018 13:36:24	U	SMPL
CCB65 - 1	CCB65	71	0.0001	-7470A	PPB		24 Oct 2018 13:38:33	U	SMPL
J5621-09 - 1	PA-SW-01	126	0.0148	-7470A	PPB		24 Oct 2018 13:40:40	U	SMPL
J5621-09DUP - 1	PA-SW-01	143	0.0193	-7470A	PPB		24 Oct 2018 13:42:46	U	SMPL
J5621-10 - 1	J5621-09MS	15590	4.1463	-7470A	PPB		24 Oct 2018 13:44:53	U	SMPL
J5621-11 - 1	J5621-09MSD	15399	4.0952	-7470A	PPB		24 Oct 2018 13:47:00	U	SMPL
J5624-10 - 1	PA-SW-02	104	0.0089	-7470A	PPB		24 Oct 2018 13:49:08	U	SMPL
J5626-04 - 1	PA-TWP-01	12153	3.2280	-7470A	PPB		24 Oct 2018 13:51:16	U	SMPL
J5626-06 - 1	EB-2018.10.19-SED	-122	-0.0515	-7470A	PPB		24 Oct 2018 13:53:24	U	SMPL
J5626-08 - 1	EB-2018.10.19-SW	-2	-0.0194	-7470A	PPB		24 Oct 2018 13:55:32	U	SMPL
J5627-01 - 1	BP-LF1-SPW-1T-10-18	96	0.0068	-7470A	PPB		24 Oct 2018 13:57:41	U	SMPL
J5627-03 - 1	BP-LF1-SPW-2T-10-18	104	0.0089	-7470A	PPB		24 Oct 2018 13:59:48	U	SMPL
CCV66 - 1	CCV66	18661	4.9667	-7470A	PPB		24 Oct 2018 14:01:54	U	SMPL
CCB66 - 1	CCB66	65	-0.0015	-7470A	PPB		24 Oct 2018 14:04:01	U	SMPL
J5621-09LX5 - 1		85	0.0038	-7470A	PPB		24 Oct 2018 14:06:08	U	SMPL
J5621-09A - 1		18055	4.8048	-7470A	PPB		24 Oct 2018 14:08:15	U	SMPL
PB114107BL - 1	PB114107BL	96	0.0068	-7470A	PPB		24 Oct 2018 14:10:22	U	SMPL
PB114107BS - 1	PB114107BS	14431	3.8366	-7470A	PPB		24 Oct 2018 14:12:29	U	SMPL
PB114107TB - 1	PB114107TB	121	0.0134	-7470A	PPB		24 Oct 2018 14:14:37	U	SMPL
J5619-01 - 1	NB-OILY-PADS	208	0.0367	-7470A	PPB		24 Oct 2018 14:16:45	U	SMPL
J5620-01 - 1	CG-OILY-WIPES	24	-0.0125	-7470A	PPB		24 Oct 2018 14:18:53	U	SMPL
J5622-02 - 1	KG-VB-16195	88	0.0046	-7470A	PPB		24 Oct 2018 14:21:01	U	SMPL
J5622-02DUP - 1	KG-VB-16195	107	0.0097	-7470A	PPB		24 Oct 2018 14:23:10	U	SMPL
J5622-02MS - 1	KG-VB-16195	15710	4.1783	-7470A	PPB		24 Oct 2018 14:25:17	U	SMPL
CCV67 - 1	CCV67	18401	4.8973	-7470A	PPB		24 Oct 2018 14:27:24	U	SMPL
CCB67 - 1	CCB67	65	-0.0015	-7470A	PPB		24 Oct 2018 14:29:31	U	SMPL
J5622-02MSD - 1	KG-VB-16195	14430	3.8363	-7470A	PPB		24 Oct 2018 14:31:38	U	SMPL
J5622-04 - 1	KG-V23608	191	0.0321	-7470A	PPB		24 Oct 2018 14:33:45	U	SMPL
J5622-02LX5 - 1	KG-VB-16195	55	-0.0042	-7470A	PPB		24 Oct 2018 14:35:53	U	SMPL
J5622-02A - 1	KG-VB-16195	14995	3.9873	-7470A	PPB		24 Oct 2018 14:38:00	U	SMPL
CCV68 - 1	CCV68	18472	4.9162	-7470A	PPB		24 Oct 2018 14:40:07	U	SMPL
CCB68 - 1	CCB68	-114	-0.0493	-7470A	PPB		24 Oct 2018 14:42:15	U	SMPL