



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: pH

Supervisor Review By : JIGNESH

Run Number: LB98814

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.6	4.01	10/24/2018	14:30
2	CAL2	1	Water			21.8	7.00	10/24/2018	14:32
3	CAL3	1	Water			21.7	10.04	10/24/2018	14:34
4	ICV	1	Water			21.7	7.01	10/24/2018	14:36
5	CCV1	1	Water			21.6	2.00	10/24/2018	14:38
6	J5649-01	1	Solid	20.02	20	20.6	6.93	10/24/2018	14:40
7	J5649-01DUP	1	Solid	20.03	20	20.7	6.94	10/24/2018	14:42
8	J5649-02	1	Solid	20.02	20	20.7	5.54	10/24/2018	14:44
9	J5649-03	1	Solid	20.04	20	20.7	4.87	10/24/2018	14:46
10	J5649-04	1	Solid	20.02	20	20.6	5.11	10/24/2018	14:48
11	J5649-05	1	Solid	20.03	20	20.7	5.00	10/24/2018	14:50
12	J5649-06	1	Solid	20.03	20	20.3	5.56	10/24/2018	14:52
13	J5649-09	1	Solid	20.04	20	20.3	5.05	10/24/2018	14:54
14	J5649-10	1	Solid	20.02	20	20.2	4.96	10/24/2018	14:56
15	J5649-11	1	Solid	20.03	20	20.5	5.06	10/24/2018	14:58
16	CCV2	1	Water			21.8	12.00	10/24/2018	15:00
17	J5649-12	1	Solid	20.02	20	20.4	8.48	10/24/2018	15:02
18	J5649-13	1	Solid	20.04	20	20.5	7.85	10/24/2018	15:04
19	CCV3	1	Water			21.7	2.00	10/24/2018	15:06

WORKLIST(Hardcopy Internal Chain)

VB 98814

WorkList Name : pH102418

WorkList ID : 118414

Date : 10/24/2018 2:04:23 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/23/2018	Solid	J5649-01	F pH	Cool 4 deg C	BEMS01	E62	DL-01A	10/23/2018	9045D
10/23/2018	Solid	J5649-02	F pH	Cool 4 deg C	BEMS01	E62	DL-01B	10/23/2018	9045D
10/23/2018	Solid	J5649-03	F pH	Cool 4 deg C	BEMS01	E62	DL-02A	10/23/2018	9045D
10/23/2018	Solid	J5649-04	F pH	Cool 4 deg C	BEMS01	E62	DL-02B	10/23/2018	9045D
10/23/2018	Solid	J5649-05	F pH	Cool 4 deg C	BEMS01	E62	DL-02C	10/23/2018	9045D
10/23/2018	Solid	J5649-06	F pH	Cool 4 deg C	BEMS01	E62	DL-03A	10/23/2018	9045D
10/23/2018	Solid	J5649-09	F pH	Cool 4 deg C	BEMS01	E62	DL-03B	10/23/2018	9045D
10/23/2018	Solid	J5649-10	F pH	Cool 4 deg C	BEMS01	E62	DL-04A	10/23/2018	9045D
10/23/2018	Solid	J5649-11	F pH	Cool 4 deg C	BEMS01	E62	DL-04B	10/23/2018	9045D
10/23/2018	Solid	J5649-12	F pH	Cool 4 deg C	BEMS01	E62	DL-05A	10/23/2018	9045D
10/23/2018	Solid	J5649-13	F pH	Cool 4 deg C	BEMS01	E62	DL-05B	10/23/2018	9045D

Date/Time 10/24/18 2:10

Received by: SP

Relinquished by: CP

Date/Time 10/24/18 3:15

Received by: CP

Relinquished by: SP