



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: pH

Supervisor Review By : JIGNESH

Run Number: LB98827

Slope : 98.7

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.6	4.01	10/24/2018	14:30
2	CAL2	1	Water			21.8	7.00	10/24/2018	14:32
3	CAL3	1	Water			21.7	10.04	10/24/2018	14:34
4	ICV	1	Water			21.7	7.01	10/24/2018	14:36
5	CCV1	1	Water			21.6	2.00	10/24/2018	16:14
6	J5656-01	1	Water			12.8	7.12	10/24/2018	16:15
7	J5656-01DUP	1	Water			12.9	7.13	10/24/2018	16:17
8	CCV2	1	Water			21.6	12.00	10/24/2018	16:21

WORKLIST(Hardcopy Internal Chain)

UB 98827

WorkList Name : pH102418-1

WorkList ID : 118423

Date : 10/24/2018 2:58:23 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/24/2018	Water	J5656-01	pH	Cool 4 deg C	PSEG01	P63	WS-DRUM	10/24/2018	9040C

Date/Time 10/24/18 4:12
 Received by: SP
 Relinquished by: CP

Date/Time 10/24/18 4:40
 Received by: CP
 Relinquished by: SP