



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : JIGNESH

Run Number: LB98829

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.6	4.01	10/24/2018	14:30
2	CAL2	1	Water			21.8	7.00	10/24/2018	14:32
3	CAL3	1	Water			21.7	10.04	10/24/2018	14:34
4	ICV	1	Water			21.7	7.01	10/24/2018	14:36
5	CCV1	1	Water			21.8	2.00	10/24/2018	16:24
6	J5651-03	1	Solid	20.03	20	20.5	8.27	10/24/2018	16:26
7	J5651-03DUP	1	Solid	20.04	20	20.6	8.28	10/24/2018	16:28
8	J5652-04	1	Solid	20.02	20	19.7	10.89	10/24/2018	16:30
9	J5661-02	1	Solid	20.02	20	19.4	7.40	10/24/2018	16:32
10	CCV2	1	Water			21.7	12.00	10/24/2018	16:34

WORKLIST(Hardcopy Internal Chain)

96529

WorkList Name : corrosivity102418

WorkList ID : 118420

Date : 10/24/2018 2:58:02 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/24/2018	Solid	J5651-03	Corrosivity	Cool 4 deg C	PSEG01	P62	RBR-200030	10/24/2018	9045D
10/24/2018	Solid	J5652-04	Corrosivity	Cool 4 deg C	PSEG01	P63	VB16206-VB16207-V236-VB1	10/24/2018	9045D
10/24/2018	Solid	J5661-02	Corrosivity	Cool 4 deg C	PSEG01	P62	VB44949	10/24/2018	9045D

Date/Time 10/24/18 4:50
 Received by: SP
 Relinquished by: SP

10/24/18
 Date/Time 10/24/18 4:40
 Received by: CP
 Relinquished by: SP