

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB98830

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP65054
calibration std. phosphate 1 ppm	WP69033
calibration std. phosphate 0.5 ppm	WP69032
calibration std. phosphate 0.3 ppm	WP69031
calibration std. 0 ppm	WP69028
phosphate ICV std.	WP69037
phosphate CCV std.	WP69036
calibration std. phosphate 0.1 ppm	WP69030
calibration std. phosphate 0.05 ppm	WP69029
11 N sulfuric acid	WP66905
Phenolphthalein indicator	WP66907
Sodium hydroxide, 1N	WP66924

Intercept: -0.0009Slope: 0.6504Regression: 0.999829

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	10/24/2018	11:42
2	CAL2	0.05	1	50	50	0.033	10/24/2018	11:42
3	CAL3	0.10	1	50	50	0.065	10/24/2018	11:43
4	CAL4	0.30	1	50	50	0.186	10/24/2018	11:43
5	CAL5	0.50	1	50	50	0.330	10/24/2018	11:43
6	CAL6	1.00	1	50	50	0.649	10/24/2018	11:44



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB98830

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.324	0.500	10/24/2018	11:45
2	ICB		1	50	50	0.000	0.001	10/24/2018	11:45
3	CCV1	0.50	1	50	50	0.326	0.503	10/24/2018	11:46
4	CCB1		1	50	50	0.000	0.001	10/24/2018	11:46
5	PB114153BL		1	50	50	0.000	0.001	10/24/2018	11:47
6	PB114153BS	0.50	1	50	50	0.342	0.527	10/24/2018	11:48
7	J5569-01		1	50	50	0.063	0.098	10/24/2018	11:48
8	J5569-02		1	50	50	0.047	0.074	10/24/2018	11:49
9	J5569-05		1	50	50	0.047	0.074	10/24/2018	11:49
10	J5569-06		1	50	50	0.080	0.124	10/24/2018	11:50
11	J5569-06DUP		1	50	50	0.081	0.126	10/24/2018	11:51
12	J5569-06MS	0.50	1	50	50	0.410	0.632	10/24/2018	11:51
13	J5569-06MSD	0.50	1	50	50	0.411	0.633	10/24/2018	11:52
14	J5569-07		1	50	50	0.082	0.127	10/24/2018	11:52
15	CCV2	0.50	1	50	50	0.327	0.504	10/24/2018	11:53
16	CCB2		1	50	50	0.000	0.001	10/24/2018	11:53
17	J5569-08		1	50	50	0.073	0.114	10/24/2018	11:54
18	J5569-09		1	50	50	0.107	0.166	10/24/2018	11:54
19	J5569-12		1	50	50	0.062	0.097	10/24/2018	11:55
20	J5569-13		1	50	50	0.096	0.149	10/24/2018	11:55
21	J5586-01		1	50	50	0.032	0.051	10/24/2018	11:56
22	J5591-01		1	50	50	1.789	2.752	10/24/2018	11:57
23	J5591-01		10	50	50	0.584	0.899	10/24/2018	12:00
24	CCV3	0.50	1	50	50	0.328	0.506	10/24/2018	12:01
25	CCB3		1	50	50	0.000	0.001	10/24/2018	12:01