


CB 98834 INSTRUMENT ID: CVI

7471B

μ Abs.:

52996



Linear

A= 0.0000e+000

B= 1.8873e-004

C= -2.9933e-002

Rho= 0.9998886

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.002	0.002	168	0.000	168				
0.2	0.200	0.187	-0.013	1148	0.0 %	1148				
2.5	2.500	2.447	-0.053	13127	0.0 %	13127				
5.0	5.000	5.118	0.118	27275	0.0 %	27275				
7.5	7.500	7.475	-0.025	39765	0.0 %	39765				
10.0	10.000	9.972	-0.028	52996	0.0 %	52996				

LB98834 INSTRUMENT ID :CV1

Method: 7471B

Operator: Admin

Date of Analysis: 25 Oct 2018 12:41:05

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD - 0	168	-	0.0000	7471B	PPB	25 Oct 2018 12:50:57	S	Std
0.2 - 1	SD - 2	1148	-	0.2000	7471B	PPB	25 Oct 2018 12:53:16	S	Std
2.5 - 1	SD - 5	13127	-	2.5000	7471B	PPB	25 Oct 2018 12:55:33	S	Std
5.0 - 1	SD - 0	27275	-	5.0000	7471B	PPB	25 Oct 2018 12:57:50	S	Std
7.5 - 1	SD - 5	39765	-	7.5000	7471B	PPB	25 Oct 2018 13:00:07	S	Std
10.0 - 1	SD - 0	52996	-	10.0000	7471B	PPB	25 Oct 2018 13:02:24	S	Std
ICV38 - 1	ICV38	21574	4.0416	-	7471B	PPB	25 Oct 2018 13:06:35	U	SMPL
ICB38 - 1	ICB38	87	-0.0135	-	7471B	PPB	25 Oct 2018 13:08:50	U	SMPL
CCV73 - 1	CCV73	27361	5.1338	-	7471B	PPB	25 Oct 2018 13:11:06	U	SMPL
CCB73 - 1	CCB73	80	-0.0148	-	7471B	PPB	25 Oct 2018 13:13:22	U	SMPL
CRA - 1	CRA	1131	0.1835	-	7471B	PPB	25 Oct 2018 13:15:38	U	SMPL
HighStd - 1	HighStd	52696	9.9151	-	7471B	PPB	25 Oct 2018 13:17:54	U	SMPL
ChkStd - 1	ChkStd	39372	7.4006	-	7471B	PPB	25 Oct 2018 13:23:23	U	SMPL
PB114105BL - 1	PBS	51	-0.0203	-	7471B	PPB	25 Oct 2018 13:25:39	U	SMPL
PB114105BS - 1	LCSS	18841	3.5258	-	7471B	PPB	25 Oct 2018 13:27:57	U	SMPL
J5610-01 - 1	SB-1(0-2)	7288	1.3455	-	7471B	PPB	25 Oct 2018 13:30:14	U	SMPL
J5610-01DUP - 1	SB-1(0-2)DUP	7365	1.3600	-	7471B	PPB	25 Oct 2018 13:32:31	U	SMPL
J5610-01MS - 1	SB-1(0-2)MS	25956	4.8686	-	7471B	PPB	25 Oct 2018 13:34:48	U	SMPL
J5610-01MSD - 1	SB-1(0-2)MSD	30921	5.8056	-	7471B	PPB	25 Oct 2018 13:37:06	U	SMPL
J5610-02 - 1	SB-1(7-9)	9413	1.7465	-	7471B	PPB	25 Oct 2018 13:39:23	U	SMPL
CCV74 - 1	CCV74	27111	5.0866	-	7471B	PPB	25 Oct 2018 13:41:40	U	SMPL
CCB74 - 1	CCB74	73	-0.0162	-	7471B	PPB	25 Oct 2018 13:43:56	U	SMPL
J5610-03 - 1	SB-2(0-2)	5806	1.0658	-	7471B	PPB	25 Oct 2018 13:46:11	U	SMPL
J5610-04 - 1	SB-2(6-8)	11696	2.1774	-	7471B	PPB	25 Oct 2018 13:48:27	U	SMPL
J5610-05 - 1	SB-3(0-2)	137412	25.9032	-	7471B	PPB	25 Oct 2018 13:50:43	U	SMPL
J5610-06 - 1	SB-3(6-8)	9603	1.7824	-	7471B	PPB	25 Oct 2018 13:53:00	U	SMPL
J5610-07 - 1	SB-4(0-2)	13438	2.5062	-	7471B	PPB	25 Oct 2018 13:55:20	U	SMPL
J5610-08 - 1	SB-4(8-10)	10893	2.0259	-	7471B	PPB	25 Oct 2018 13:57:37	U	SMPL
J5610-09 - 1	SB-5(0-2)	2362	0.4158	-	7471B	PPB	25 Oct 2018 13:59:54	U	SMPL
J5610-10 - 1	SB-6(0-2)	8775	1.6261	-	7471B	PPB	25 Oct 2018 14:02:11	U	SMPL
J5610-11 - 1	SB-7(0-2)	26384	4.9494	-	7471B	PPB	25 Oct 2018 14:04:29	U	SMPL
J5618-02 - 1	SSSI-1018-S-DH-13	2484	0.4389	-	7471B	PPB	25 Oct 2018 14:06:45	U	SMPL
CCV75 - 1	CCV75	27698	5.1974	-	7471B	PPB	25 Oct 2018 14:09:03	U	SMPL
CCB75 - 1	CCB75	53	-0.0199	-	7471B	PPB	25 Oct 2018 14:11:19	U	SMPL
J5618-04 - 1	SSSI-1018-S-DH-14	3244	0.5823	-	7471B	PPB	25 Oct 2018 14:13:35	U	SMPL
J5618-06 - 1	SSSI-1018-S-DH-15	1904	0.3294	-	7471B	PPB	25 Oct 2018 14:15:51	U	SMPL
J5618-08 - 1	SSSI-1019-S-DH-16	13164	2.4544	-	7471B	PPB	25 Oct 2018 14:18:08	U	SMPL
J5622-01 - 1	KG-VB-16195	53161	10.0029	-	7471B	PPB	25 Oct 2018 14:20:24	U	SMPL
J5622-03 - 1	KG-V23608	34831	6.5436	-	7471B	PPB	25 Oct 2018 14:22:41	U	SMPL
J5610-01LX5 - 1		1503	0.2537	-	7471B	PPB	25 Oct 2018 14:25:00	U	SMPL
J5610-01A - 1		27681	5.1942	-	7471B	PPB	25 Oct 2018 14:27:19	U	SMPL
J5610-05DLX10 - 1		14764	2.7564	-	7471B	PPB	25 Oct 2018 14:40:13	U	SMPL
CCV76 - 1	CCV76	27552	5.1698	-	7471B	PPB	25 Oct 2018 14:42:30	U	SMPL
CCB76 - 1	CCB76	94	-0.0122	-	7471B	PPB	25 Oct 2018 14:44:47	U	SMPL