

LB98845

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

10/25/2018 15:06

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.021	0.0	0.113	
ICB1	0.000	0.0	0.004	
CCV1	1.011	0.0	0.112	
CCB1	0.004	0.0	0.004	
PB113984BL	0.004	0.0	0.004	
PB113984BS	0.970	0.0	0.108	
J5569-01	0.024	0.0	0.006	
J5569-01DUP	0.020	0.0	0.006	
J5569-02	0.004	0.0	0.004	
J5569-05	0.009	0.0	0.005	
J5569-06	0.033	0.0	0.007	
J5569-07	0.003	0.0	0.004	
CCV2	1.014	0.0	0.112	
CCB2	0.003	0.0	0.004	
J5569-08	0.032	0.0	0.007	
J5569-09	0.007	0.0	0.004	
J5569-12	0.006	0.0	0.004	
J5569-13	0.006	0.0	0.004	
J5586-01	0.008	0.0	0.005	
J5588-01	0.002	0.0	0.004	
J5569-01MS	1.008	0.0	0.112	
J5569-01MSD	1.014	0.0	0.112	
CCV3	1.011	0.0	0.112	
CCB3	0.003	0.0	0.004	

N	24
Mean	0.301
SD	0.4631
CV%	153.94

Aquakem v. 7.2AQ1

Results from time period:

Thu Oct 25 13:53:22 2018

Thu Oct 25 15:01:55 2018

AO

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0572	mg/l	10/25/2018 13:53:22	
0.1PPM	A	Phenolics-	P	0.1014	mg/l	10/25/2018 13:53:23	
0.25PPM	A	Phenolics-	P	0.2487	mg/l	10/25/2018 13:53:24	
0.50PPM	A	Phenolics-	P	0.4948	mg/l	10/25/2018 13:53:25	
1.0PPM	A	Phenolics-	P	0.9935	mg/l	10/25/2018 13:53:26	
2.0PPM	A	Phenolics-	P	2.0045	mg/l	10/25/2018 13:53:27	
ICV1	S	Phenolics-	P	1.0214	mg/l	10/25/2018 14:28:02	
ICB1	S	Phenolics-	P	0.0004	mg/l	10/25/2018 14:28:03	
CCV1	S	Phenolics-	P	1.0106	mg/l	10/25/2018 14:28:04	
CCB1	S	Phenolics-	P	0.0045	mg/l	10/25/2018 14:28:05	
PB113984BL	S	Phenolics-	P	0.0042	mg/l	10/25/2018 14:28:06	
PB113984BS	S	Phenolics-	P	0.9702	mg/l	10/25/2018 14:28:07	
J5569-01	S	Phenolics-	P	0.0245	mg/l	10/25/2018 14:28:08	
J5569-01DUP	S	Phenolics-	P	0.0198	mg/l	10/25/2018 14:28:09	
J5569-02	S	Phenolics-	P	0.004	mg/l	10/25/2018 14:28:12	
J5569-05	S	Phenolics-	P	0.0095	mg/l	10/25/2018 14:28:13	
J5569-06	S	Phenolics-	P	0.0326	mg/l	10/25/2018 14:38:04	
J5569-07	S	Phenolics-	P	0.0034	mg/l	10/25/2018 14:38:05	
CCV2	S	Phenolics-	P	1.0136	mg/l	10/25/2018 14:38:06	
CCB2	S	Phenolics-	P	0.0028	mg/l	10/25/2018 14:38:07	
J5569-08	S	Phenolics-	P	0.0325	mg/l	10/25/2018 14:38:08	
J5569-09	S	Phenolics-	P	0.0072	mg/l	10/25/2018 14:38:09	
J5569-12	S	Phenolics-	P	0.0063	mg/l	10/25/2018 14:38:10	
J5569-13	S	Phenolics-	P	0.006	mg/l	10/25/2018 14:38:11	
J5586-01	S	Phenolics-	P	0.008	mg/l	10/25/2018 14:38:12	
J5588-01	S	Phenolics-	P	0.0023	mg/l	10/25/2018 14:38:13	
J5569-01MS	S	Phenolics-	P	1.0084	mg/l	10/25/2018 15:01:52	
J5569-01MSD	S	Phenolics-	P	1.0143	mg/l	10/25/2018 15:01:53	
CCV3	S	Phenolics-	P	1.0106	mg/l	10/25/2018 15:01:54	
CCB3	S	Phenolics-	P	0.0027	mg/l	10/25/2018 15:01:55	

6398845
CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

10/25/2018 13:55


Test Phenolics-

Accepted 10/25/2018 13:55

Factor 9.336
Bias 0.004

Coeff. of det. 0.999948

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.010	0.0572	0.0500	
2	Phenol-2	0.015	0.1014	0.1000	
3	Phenol-2	0.030	0.2487	0.2500	
4	Phenol-2	0.057	0.4948	0.5000	
5	Phenol-2	0.110	0.9935	1.0000	
6	Phenol-2	0.218	2.0045	2.0000	