



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB98847

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.9	10/25/2018	09:50
2	J5656-02		1	8	>100.0	>212.0	>212.0	10/25/2018	10:20
3	J5656-02DUP		1	8	>100.0	>212.0	>212.0	10/25/2018	10:45
4	J5659-01		1	7	75.0	167.0	167.3	10/25/2018	11:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB98847

WorkList Name : fp102518

WorkList ID : 118452

Date : 10/25/2018 9:14:26 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/29/2018	Water	J5656-02	Flash Point	Cool 4 deg C	PSEG01	P63	WS-DRUM	10/24/2018	1010A
10/31/2018	Water	J5659-01	Flash Point	Cool 4 deg C	ENV177	F12	OIL01	10/23/2018	1010A

Date/Time 10/25/18 9:15AM

Received by: KS

Relinquished by: CP

Date/Time 10/25/18 11:30AM

Received by: CP

Relinquished by: KS