


LB 98849 INSTRUMENT ID: CVI

7471B

Linear

 μ Abs.:

53098



A= 0.0000e+000

B= 1.8797e-004

C= -5.2951e-002

Rho= 0.9998949

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.027	-0.027	139	0.000	139				
0.2	0.200	0.160	-0.040	1133	0.0 %	1133				
2.5	2.500	2.520	0.020	13686	0.0 %	13686				
5.0	5.000	5.093	0.093	27375	0.0 %	27375				
7.5	7.500	7.526	0.026	40320	0.0 %	40320				
10.0	10.000	9.928	-0.072	53098	0.0 %	53098				

LB98849 INSTRUMENT ID:CV1

Method: 7471B

Operator: Admin

Date of Analysis: 25 Oct 2018 15:18:00

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	139	-	0.0000	7471B	PPB	25 Oct 2018 15:19:13	S	Std
0.2 - 1	50.2	1133	-	0.2000	7471B	PPB	25 Oct 2018 15:21:31	S	Std
2.5 - 1	52.5	13686	-	2.5000	7471B	PPB	25 Oct 2018 15:23:48	S	Std
5.0 - 1	55.0	27375	-	5.0000	7471B	PPB	25 Oct 2018 15:26:06	S	Std
7.5 - 1	57.5	40320	-	7.5000	7471B	PPB	25 Oct 2018 15:28:23	S	Std
10.0 - 1	59.0	53098	-	10.0000	7471B	PPB	25 Oct 2018 15:48:37	S	Std
ICV39 - 1	ICV39	20582	3.8159	-7471B	PPB	25 Oct 2018 15:50:52	U	SMPL	
ICB39 - 1	ICB39	59	-0.0419	-7471B	PPB	25 Oct 2018 15:53:10	U	SMPL	
CCV77 - 1	CCV77	27666	5.1475	-7471B	PPB	25 Oct 2018 15:55:26	U	SMPL	
CCB77 - 1	CCB77	42	-0.0451	-7471B	PPB	25 Oct 2018 15:57:42	U	SMPL	
CRA - 1	CRA	1121	0.1578	-7471B	PPB	25 Oct 2018 15:59:59	U	SMPL	
HighStd - 1	HighStd	52979	9.9057	-7471B	PPB	25 Oct 2018 16:02:14	U	SMPL	
ChkStd - 1	ChkStd	40046	7.4747	-7471B	PPB	25 Oct 2018 16:04:30	U	SMPL	
PB114121BL - 1	PB114121BL	47	-0.0441	-7471B	PPB	25 Oct 2018 16:06:49	U	SMPL	
PB114121BS - 1	PB114121BS	18525	3.4293	-7471B	PPB	25 Oct 2018 16:09:07	U	SMPL	
J5624-01 - 1	PA-SED-01	3490	0.6031	-7471B	PPB	25 Oct 2018 16:11:23	U	SMPL	
J5624-01DUP - 1	PA-SED-01DUP	3249	0.5578	-7471B	PPB	25 Oct 2018 16:13:41	U	SMPL	
J5624-02 - 1	J5624-01MS	23491	4.3628	-7471B	PPB	25 Oct 2018 16:15:58	U	SMPL	
J5624-03 - 1	J5624-01MSD	21174	3.9272	-7471B	PPB	25 Oct 2018 16:18:16	U	SMPL	
J5624-04 - 1	PA-SED-02	4958	0.8790	-7471B	PPB	25 Oct 2018 16:20:32	U	SMPL	
CCV78 - 1	CCV78	26868	4.9975	-7471B	PPB	25 Oct 2018 16:22:48	U	SMPL	
CCB78 - 1	CCB78	80	-0.0379	-7471B	PPB	25 Oct 2018 16:25:03	U	SMPL	
J5624-05 - 1	PA-SB-03A	5604	1.0005	-7471B	PPB	25 Oct 2018 16:27:19	U	SMPL	
J5624-06 - 1	PA-SB-03B	232031	43.5629	-7471B	PPB	25 Oct 2018 16:29:35	U	SMPL	
J5624-07 - 1	PA-SB-02A	21320	3.9547	-7471B	PPB	25 Oct 2018 16:31:51	U	SMPL	
J5624-08 - 1	PA-SB-02B	4516	0.7959	-7471B	PPB	25 Oct 2018 16:34:14	U	SMPL	
J5624-09 - 1	PA-SB-01A	34241	6.3835	-7471B	PPB	25 Oct 2018 16:36:32	U	SMPL	
J5625-01 - 1	DW-01	10917	1.9992	-7471B	PPB	25 Oct 2018 16:38:49	U	SMPL	
J5625-03 - 1	DW-02	10235	1.8710	-7471B	PPB	25 Oct 2018 16:41:08	U	SMPL	
J5625-05 - 1	DW-03	10998	2.0144	-7471B	PPB	25 Oct 2018 16:43:25	U	SMPL	
J5625-07 - 1	DW-04	9200	1.6764	-7471B	PPB	25 Oct 2018 16:45:43	U	SMPL	
J5625-09 - 1	DW-05	9823	1.7935	-7471B	PPB	25 Oct 2018 16:47:59	U	SMPL	
CCV79 - 1	CCV79	27156	5.0517	-7471B	PPB	25 Oct 2018 16:50:16	U	SMPL	
CCB79 - 1	CCB79	40	-0.0454	-7471B	PPB	25 Oct 2018 16:52:32	U	SMPL	
J5625-11 - 1	DW-06	10319	1.8868	-7471B	PPB	25 Oct 2018 16:54:49	U	SMPL	
J5625-13 - 1	DW-07	8574	1.5587	-7471B	PPB	25 Oct 2018 16:57:05	U	SMPL	
J5625-15 - 1	DW-08	11680	2.1426	-7471B	PPB	25 Oct 2018 16:59:22	U	SMPL	
J5626-01 - 1	PA-SB-01B	2280	0.3756	-7471B	PPB	25 Oct 2018 17:01:38	U	SMPL	
J5626-02 - 1	J5626-01MS	20195	3.7432	-7471B	PPB	25 Oct 2018 17:03:55	U	SMPL	
J5626-03 - 1	J5626-01MSD	21449	3.9789	-7471B	PPB	25 Oct 2018 17:06:12	U	SMPL	
J5624-01LX5 - 1		644	0.0681	-7471B	PPB	25 Oct 2018 17:08:32	U	SMPL	
J5624-01A - 1		21525	3.9932	-7471B	PPB	25 Oct 2018 17:10:49	U	SMPL	
J5626-01A - 1		20841	3.8646	-7471B	PPB	25 Oct 2018 17:13:07	U	SMPL	
J5622-01DLX5 - 1		11083	2.0304	-7471B	PPB	25 Oct 2018 17:15:24	U	SMPL	
CCV80 - 1	CCV80	27266	5.0724	-7471B	PPB	25 Oct 2018 17:17:43	U	SMPL	
CCB80 - 1	CCB80	79	-0.0381	-7471B	PPB	25 Oct 2018 17:19:59	U	SMPL	
J5624-06DLX10 - 1		25898	4.8152	-7471B	PPB	25 Oct 2018 17:22:17	U	SMPL	
CCV081 - 1	CCV081	26971	5.0169	-7471B	PPB	25 Oct 2018 17:24:33	U	SMPL	
CCB081 - 1	CCB081	46	-0.0443	-7471B	PPB	25 Oct 2018 17:26:50	U	SMPL	