



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: pH

Supervisor Review By : JIGNESH

Run Number: LB98851

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.9	4.01	10/25/2018	14:00
2	CAL2	1	Water			21.8	7.00	10/25/2018	14:02
3	CAL3	1	Water			21.8	10.04	10/25/2018	14:04
4	ICV	1	Water			21.7	7.01	10/25/2018	14:06
5	CCV1	1	Water			21.9	2.00	10/25/2018	14:08
6	J5665-01	1	Solid	20.02	20	19.6	4.68	10/25/2018	14:10
7	J5665-02	1	Solid	20.03	20	19.7	7.40	10/25/2018	14:12
8	J5665-02DUP	1	Solid	20.04	20	19.8	7.41	10/25/2018	14:14
9	J5665-03	1	Solid	20.03	20	19.0	4.44	10/25/2018	14:16
10	J5665-04	1	Solid	20.02	20	19.5	7.29	10/25/2018	14:18
11	J5665-05	1	Solid	20.03	20	19.1	5.82	10/25/2018	14:20
12	J5665-06	1	Solid	20.02	20	19.0	5.77	10/25/2018	14:22
13	CCV2	1	Water			21.8	12.00	10/25/2018	14:24

WORKLIST(Hardcopy Internal Chain)

10/25/18

WorkList Name : pH102518

WorkList ID : 118476

Date : 10/25/2018 1:44:02 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/24/2018	Solid	J5665-01	pH	Cool 4 deg C	BEMS01	F21	DL-06A	10/24/2018	9045D
10/24/2018	Solid	J5665-02	pH	Cool 4 deg C	BEMS01	F21	DL-06B	10/24/2018	9045D
10/24/2018	Solid	J5665-03	pH	Cool 4 deg C	BEMS01	F21	DL-07A	10/24/2018	9045D
10/24/2018	Solid	J5665-04	pH	Cool 4 deg C	BEMS01	F21	DL-07B	10/24/2018	9045D
10/24/2018	Solid	J5665-05	pH	Cool 4 deg C	BEMS01	F21	DL-08A	10/24/2018	9045D
10/24/2018	Solid	J5665-06	pH	Cool 4 deg C	BEMS01	F21	DL-08B	10/24/2018	9045D

Date/Time 10/25/18 1:50

Received by: SP

Relinquished by: CP

Date/Time 10/25/18 2:30

Received by: CP

Relinquished by: JSP