



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : JIGNESH

Run Number: LB98852

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.9	4.01	10/25/2018	14:00
2	CAL2	1	Water			21.8	7.00	10/25/2018	14:02
3	CAL3	1	Water			21.8	10.04	10/25/2018	14:04
4	ICV	1	Water			21.7	7.01	10/25/2018	14:06
5	CCV1	1	Water			21.9	2.00	10/25/2018	14:08
6	J5665-11	1	Solid	20.03	20	19.7	7.40	10/25/2018	14:12
7	J5665-11DUP	1	Solid	20.04	20	19.8	7.41	10/25/2018	14:14
8	J5665-12	1	Solid	20.03	20	19.1	5.82	10/25/2018	14:20
9	CCV2	1	Water			21.8	12.00	10/25/2018	14:24

WORKLIST(Hardcopy Internal Chain)

LB 98852

WorkList Name : corrosivity102518

WorkList ID : 118477

Date : 10/25/2018 1:44:10 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/24/2018	Solid	J5665-11	Corrosivity	Cool 4 deg C	BEMS01	F21	DL-06B	10/24/2018	9045D
10/24/2018	Solid	J5665-12	Corrosivity	Cool 4 deg C	BEMS01	F21	DL-08A	10/24/2018	9045D

Date/Time 10/25/18 7:50
 Received by: SP
 Relinquished by: SP

Date/Time 10/25/18 2:30
 Received by: SP
 Relinquished by: SP