

CB 98859

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : Ab

10/25/2018 16:17

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.031	0.0	0.114	
ICB1	0.004	0.0	0.004	
CCV1	1.012	0.0	0.112	
CCB1	0.002	0.0	0.004	
J5565-03	0.000	0.0	0.004	
J5565-04	0.002	0.0	0.004	
J5565-04DUP	0.002	0.0	0.004	
J5565-05MS	0.940	0.0	0.104	
J5565-06MSD	0.954	0.0	0.106	
J5565-07	0.001	0.0	0.004	
J5565-08	-0.001	0.0	0.004	
CCV2	1.020	0.0	0.113	
CCB2	0.003	0.0	0.004	
PB114018BL	0.001	0.0	0.004	
PB114018BS	1.048	0.0	0.116	
J5565-01	0.000	0.0	0.004	
J5565-02	-0.002	0.0	0.003	
CCV3	1.028	0.0	0.114	
CCB3	0.003	0.0	0.004	

N	19
Mean	0.371
SD	0.4980
CV%	134.25

Aquakem v. 7.2AQ1

Results from time period:

Thu Oct 25 13:53:22 2018

Thu Oct 25 16:10:40 2018



Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0572	mg/l	10/25/2018 13:53:22	
0.1PPM	A	Phenolics-	P	0.1014	mg/l	10/25/2018 13:53:23	
0.25PPM	A	Phenolics-	P	0.2487	mg/l	10/25/2018 13:53:24	
0.50PPM	A	Phenolics-	P	0.4948	mg/l	10/25/2018 13:53:25	
1.0PPM	A	Phenolics-	P	0.9935	mg/l	10/25/2018 13:53:26	
2.0PPM	A	Phenolics-	P	2.0045	mg/l	10/25/2018 13:53:27	
ICV1	S	Phenolics-	P	1.0312	mg/l	10/25/2018 15:28:00	
ICB1	S	Phenolics-	P	0.0038	mg/l	10/25/2018 15:28:01	
CCV1	S	Phenolics-	P	1.0123	mg/l	10/25/2018 15:28:02	
CCB1	S	Phenolics-	P	0.0016	mg/l	10/25/2018 15:28:03	
J5565-03	S	Phenolics-	P	0.0001	mg/l	10/25/2018 15:28:04	
J5565-04	S	Phenolics-	P	0.002	mg/l	10/25/2018 15:28:05	
J5565-04DUP	S	Phenolics-	P	0.0015	mg/l	10/25/2018 15:28:06	
J5565-05MS	S	Phenolics-	P	0.94	mg/l	10/25/2018 15:28:07	
J5565-06MSD	S	Phenolics-	P	0.9536	mg/l	10/25/2018 15:28:08	
J5565-07	S	Phenolics-	P	0.001	mg/l	10/25/2018 15:28:09	
J5565-08	S	Phenolics-	P	-0.0012	mg/l	10/25/2018 15:28:10	
CCV2	S	Phenolics-	P	1.0204	mg/l	10/25/2018 15:35:18	
CCB2	S	Phenolics-	P	0.003	mg/l	10/25/2018 15:35:19	
PB114018BL	S	Phenolics-	P	0.0011	mg/l	10/25/2018 16:10:35	
PB114018BS	S	Phenolics-	P	1.0482	mg/l	10/25/2018 16:10:36	
J5565-01	S	Phenolics-	P	0.0005	mg/l	10/25/2018 16:10:37	
J5565-02	S	Phenolics-	P	-0.0018	mg/l	10/25/2018 16:10:38	
CCV3	S	Phenolics-	P	1.0284	mg/l	10/25/2018 16:10:39	
CCB3	S	Phenolics-	P	0.0026	mg/l	10/25/2018 16:10:40	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

10/25/2018 13:55


Test Phenolics-

Accepted 10/25/2018 13:55

Factor 9.336
Bias 0.004

Coeff. of det. 0.999948

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.010	0.0572	0.0500	
2	Phenol-2	0.015	0.1014	0.1000	
3	Phenol-2	0.030	0.2487	0.2500	
4	Phenol-2	0.057	0.4948	0.5000	
5	Phenol-2	0.110	0.9935	1.0000	
6	Phenol-2	0.218	2.0045	2.0000	