



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB98860

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP69099
Turbidity std, 40NTU	WP69094
turbidity std 5 NTU	WP69097
turbidity std 20 NTU	WP69095
Turbidity std 1 NTU	WP69098
turbidity std 0 NTU	WP69111
turbidity std 10 NTU	WP69096
10 NTU Standard 500 ml	W2297

Intercept: 0.1775

Slope: 0.9391

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.303	10/26/2018	11:25
2	CAL2	1	1	0.991	10/26/2018	11:17
3	CAL3	5	1	4.860	10/26/2018	11:20
4	CAL4	10	1	9.870	10/26/2018	11:25
5	CAL5	20	1	18.600	10/26/2018	11:30
6	CAL6	40	1	37.800	10/26/2018	11:34
7	CAL7	200	1	188.000	10/26/2018	11:36



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB98860

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.390	9.810	10/26/2018	11:38
2	ICB		1	0.297	0.127	10/26/2018	11:41
3	CCV1	10	1	9.520	9.948	10/26/2018	11:45
4	CCB1		1	0.289	0.119	10/26/2018	11:48
5	LB98860BL		1	0.296	0.126	10/26/2018	11:51
6	J5694-01		1	105.000	111.620	10/26/2018	11:54
7	J5694-01DUP		1	104.000	110.555	10/26/2018	11:59
8	CCV2	10	1	9.560	9.991	10/26/2018	12:03
9	CCB2		1	0.298	0.128	10/26/2018	12:06