

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB98864

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP65054
calibration std. phosphate 1 ppm	WP69087
calibration std. phosphate 0.5 ppm	WP69086
calibration std. phosphate 0.3 ppm	WP69085
calibration std. 0 ppm	WP69082
phosphate ICV std.	WP69089
phosphate CCV std.	WP69088
calibration std. phosphate 0.1 ppm	WP69084
calibration std. phosphate 0.05 ppm	WP69083
Combined reagent	WP69091
11 N sulfuric acid	WP66905
Phenolphthalein indicator	WP66907
Sodium hydroxide, 1N	WP66924

Intercept: -0.0011Slope: 0.6502Regression: 0.999819

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	10/26/2018	11:50
2	CAL2	0.05	1	50	50	0.033	10/26/2018	11:50
3	CAL3	0.10	1	50	50	0.065	10/26/2018	11:51
4	CAL4	0.30	1	50	50	0.185	10/26/2018	11:51
5	CAL5	0.50	1	50	50	0.329	10/26/2018	11:52
6	CAL6	1.00	1	50	50	0.649	10/26/2018	11:53



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB98864

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.329	0.508	10/26/2018	11:53
2	ICB		1	50	50	0.000	0.002	10/26/2018	11:53
3	CCV1	0.50	1	50	50	0.328	0.506	10/26/2018	12:31
4	CCB1		1	50	50	0.000	0.002	10/26/2018	12:31
5	PB114231BL		1	50	50	0.000	0.002	10/26/2018	12:32
6	PB114231BS	0.50	1	50	50	0.322	0.497	10/26/2018	12:33
7	J5671-01		1	50	50	0.444	0.685	10/26/2018	12:32
8	J5671-01DUP		1	50	50	0.505	0.778	10/26/2018	12:33
9	J5671-01MS	0.50	1	50	50	0.773	1.191	10/26/2018	12:34
10	J5671-01MSD	0.50	1	50	50	0.770	1.186	10/26/2018	12:34
11	J5671-01MS	0.50	2	50	50	0.382	0.589	10/26/2018	12:40
12	J5671-01MSD	0.50	2	50	50	0.370	0.571	10/26/2018	12:41
13	CCV2	0.50	1	50	50	0.326	0.503	10/26/2018	12:42
14	CCB2		1	50	50	0.000	0.002	10/26/2018	12:42