

LB 98881


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

38729



A= 0.0000e+000

B= 2.5930e-004

C= 2.9904e-003

Rho= 0.9999428

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.016	0.016	50	0.000	50				
0.2	0.200	0.209	0.009	796	0.0 %	796				
2.5	2.500	2.525	0.025	9728	0.0 %	9728				
5.0	5.000	4.923	-0.077	18974	0.0 %	18974				
7.5	7.500	7.481	-0.019	28839	0.0 %	28839				
10.0	10.000	10.045	0.045	38729	0.0 %	38729				

LB98881 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 26 Oct 2018 17:16:03

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	50	-	0.0000	7470A	PPB	26 Oct 2018 17:17:00	S	Std
0.2 - 1	50.2	796	-	0.2000	7470A	PPB	26 Oct 2018 17:19:08	S	Std
2.5 - 1	52.5	9728	-	2.5000	7470A	PPB	26 Oct 2018 17:21:17	S	Std
5.0 - 1	55	18974	-	5.0000	7470A	PPB	26 Oct 2018 17:23:25	S	Std
7.5 - 1	57.5	28839	-	7.5000	7470A	PPB	26 Oct 2018 17:25:33	S	Std
10.0 - 1	510	38729	-	10.0000	7470A	PPB	26 Oct 2018 17:27:40	S	Std
ICV42 - 1	ICV42	16321	4.2350	-	7470A	PPB	26 Oct 2018 17:43:33	U	SMPL
ICB42 - 1	ICB42	83	0.0245	-	7470A	PPB	26 Oct 2018 17:45:39	U	SMPL
CCV92 - 1	CCV92	19150	4.9686	-	7470A	PPB	26 Oct 2018 17:47:45	U	SMPL
CCB92 - 1	CCB92	76	0.0227	-	7470A	PPB	26 Oct 2018 17:49:52	U	SMPL
CRA - 1	CRA	809	0.2128	-	7470A	PPB	26 Oct 2018 17:51:59	U	SMPL
HighStd - 1	HighStd	37690	9.7760	-	7470A	PPB	26 Oct 2018 17:54:05	U	SMPL
ChkStd - 1	ChkStd	28374	7.3603	-	7470A	PPB	26 Oct 2018 17:56:12	U	SMPL
PB114183BL - 1	PBW	114	0.0326	-	7470A	PPB	26 Oct 2018 17:58:19	U	SMPL
PB114183TB - 1	PB114201TB	191	0.0525	-	7470A	PPB	26 Oct 2018 18:02:36	U	SMPL
J5651-03 - 1	BS-DRUM	122	0.0346	-	7470A	PPB	26 Oct 2018 18:04:43	U	SMPL
J5651-03DUP - 1	BS-DRUMDUP	163	0.0453	-	7470A	PPB	26 Oct 2018 18:06:52	U	SMPL
J5651-03MS - 1	BS-DRUMMS	15419	4.0011	-	7470A	PPB	26 Oct 2018 18:09:00	U	SMPL
J5651-03MSD - 1	BS-DRUMMSD	15429	4.0037	-	7470A	PPB	26 Oct 2018 18:11:07	U	SMPL
PB114183BS - 1	LCSW	17011	4.4139	-	7470A	PPB	26 Oct 2018 18:13:13	U	SMPL
CCV93 - 1	CCV93	19113	4.9590	-	7470A	PPB	26 Oct 2018 18:15:20	U	SMPL
CCB93 - 1	CCB93	69	0.0209	-	7470A	PPB	26 Oct 2018 18:17:27	U	SMPL
J5652-04 - 1	CCB93	165	0.0458	-	7470A	PPB	26 Oct 2018 18:19:33	U	SMPL
J5661-02 - 1	TWP-01	175	0.0484	-	7470A	PPB	26 Oct 2018 18:21:40	U	SMPL
J5651-03LX5 - 1	OIL01	136	0.0383	-	7470A	PPB	26 Oct 2018 18:23:47	U	SMPL
J5651-03A - 1		15298	3.9697	-	7470A	PPB	26 Oct 2018 18:25:55	U	SMPL
PB114182BL - 1	PB114182BL	138	0.0388	-	7470A	PPB	26 Oct 2018 18:28:02	U	SMPL
PB114182BS - 1	PB114182BS	15333	3.9788	-	7470A	PPB	26 Oct 2018 18:30:10	U	SMPL
PB114182TB - 1	PB114182TB	212	0.0580	-	7470A	PPB	26 Oct 2018 18:32:19	U	SMPL
J5639-14 - 1		92	0.0268	-	7470A	PPB	26 Oct 2018 18:34:27	U	SMPL
J5639-14DUP - 1		129	0.0364	-	7470A	PPB	26 Oct 2018 18:36:34	U	SMPL
J5639-14MS - 1		15273	3.9633	-	7470A	PPB	26 Oct 2018 18:38:42	U	SMPL
CCV94 - 1	CCV94	19136	4.9649	-	7470A	PPB	26 Oct 2018 18:40:48	U	SMPL
CCB94 - 1	CCB94	30	0.0108	-	7470A	PPB	26 Oct 2018 18:42:56	U	SMPL
J5639-14MSD - 1		14934	3.8754	-	7470A	PPB	26 Oct 2018 18:45:04	U	SMPL
J5656-02 - 1		96	0.0279	-	7470A	PPB	26 Oct 2018 18:47:11	U	SMPL
J5639-14LX5 - 1		33	0.0115	-	7470A	PPB	26 Oct 2018 18:49:18	U	SMPL
J5639-14A - 1		14818	3.8453	-	7470A	PPB	26 Oct 2018 18:51:25	U	SMPL
CCV95 - 1	CCV95	18903	4.9045	-	7470A	PPB	26 Oct 2018 18:53:33	U	SMPL
CCB95 - 1	CCB95	26	0.0097	-	7470A	PPB	26 Oct 2018 18:55:41	U	SMPL