



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : JIGNESH

Run Number: LB98894

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.6	4.01	10/26/2018	16:20
2	CAL2	1	Water			21.4	7.00	10/26/2018	16:22
3	CAL3	1	Water			21.6	10.04	10/26/2018	16:24
4	ICV	1	Water			21.5	7.01	10/26/2018	16:26
5	CCV1	1	Water			21.3	2.00	10/26/2018	16:28
6	J5194-08	1	Solid	20.02	20	19.8	9.08	10/26/2018	16:34
7	J5194-08DUP	1	Solid	20.03	20	19.9	9.09	10/26/2018	16:36
8	J5194-10	1	Solid	20.02	20	19.9	8.73	10/26/2018	16:38
9	J5194-12	1	Solid	20.04	20	20.0	9.12	10/26/2018	16:40
10	J5697-02	1	Solid	20.03	20	19.8	7.18	10/26/2018	16:42
11	CCV2	1	Water			21.4	12.00	10/26/2018	16:44

WORKLIST(Hardcopy Internal Chain)

LB 98894

WorkList Name : corrosivity102618

WorkList ID : 118540

Date : 10/26/2018 4:14:45 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/26/2018	Solid	J5697-02	Corrosivity	Cool 4 deg C	PSEG01	P63	CHRT-58-66-98-92	10/26/2018	9045D
10/26/2018	Solid	J5194-08	Corrosivity	Cool 4 deg C	PSEG05		HR-05-102618-A	10/26/2018	9045D
10/26/2018	Solid	J5194-10	Corrosivity	Cool 4 deg C	PSEG05		HR-05-102618-B	10/26/2018	9045D
10/26/2018	Solid	J5194-12	Corrosivity	Cool 4 deg C	PSEG05		HR-05-102618-C	10/26/2018	9045D

Date/Time 10/26/18 4:16
 Received by: SP
 Relinquished by: CP

Date/Time 10/26/18 4:50
 Received by: CP
 Relinquished by: SP