

LB98896


7470A

INSTRUMENT ID: CVL

Linear

μ Abs.:

38131



A= 0.0000e+000

B= 2.6432e-004

C= -6.8318e-002

Rho= 0.9997862

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.037	-0.037	119	0.000	119				
0.2	0.200	0.179	-0.021	934	0.0 %	934				
2.5	2.500	2.532	0.032	9837	0.0 %	9837				
5.0	5.000	5.134	0.134	19682	0.0 %	19682				
7.5	7.500	7.382	-0.118	28185	0.0 %	28185				
10.0	10.000	10.011	0.011	38131	0.0 %	38131				

LB98896 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 29 Oct 2018 12:42:12

Sample ID	Extended ID	µ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	119	-	0.00007470A	PPB	29 Oct 2018 12:46:18	S	Std
0.2 - 1	50.2	934	-	0.20007470A	PPB	29 Oct 2018 12:55:46	S	Std
2.5 - 1	52.5	9837	-	2.50007470A	PPB	29 Oct 2018 12:57:54	S	Std
5.0 - 1	55	19682	-	5.00007470A	PPB	29 Oct 2018 13:00:02	S	Std
7.5 - 1	57.5	28185	-	7.50007470A	PPB	29 Oct 2018 13:02:10	S	Std
10.0 - 1	510	38131	-	10.00007470A	PPB	29 Oct 2018 13:04:18	S	Std
ICV43 - 1	ICV43	15477	4.0226	-7470A	PPB	29 Oct 2018 13:06:51	U	SMPL
ICB43 - 1	ICB43	401	0.0377	-7470A	PPB	29 Oct 2018 13:08:58	U	SMPL
CCV01 - 1	CCV01	18701	4.8748	-7470A	PPB	29 Oct 2018 13:11:04	U	SMPL
CCB01 - 1	CCB01	402	0.0379	-7470A	PPB	29 Oct 2018 13:13:11	U	SMPL
CRA - 1	CRA	960	0.1854	-7470A	PPB	29 Oct 2018 13:15:17	U	SMPL
HighStd - 1	HighStd	38166	10.0199	-7470A	PPB	29 Oct 2018 13:17:24	U	SMPL
ChkStd - 1	ChkStd	26889	7.0391	-7470A	PPB	29 Oct 2018 13:19:31	U	SMPL
PB114201BL - 1	PBW	171	-0.0231	-7470A	PPB	29 Oct 2018 13:24:14	U	SMPL
PB114201BS - 1	LCSW	14231	3.6933	-7470A	PPB	29 Oct 2018 13:26:21	U	SMPL
J5639-13 - 1	BS-DRUM	-137	-0.1045	-7470A	PPB	29 Oct 2018 13:28:29	U	SMPL
J5639-13DUP - 1	BS-DRUMDUP	-93	-0.0929	-7470A	PPB	29 Oct 2018 13:30:37	U	SMPL
J5639-13MS - 1	BS-DRUMMS	13997	3.6314	-7470A	PPB	29 Oct 2018 13:32:46	U	SMPL
J5639-13MSD - 1	BS-DRUMMSD	13923	3.6119	-7470A	PPB	29 Oct 2018 13:34:55	U	SMPL
J5649-14 - 1	TWP-01	1015	0.2000	-7470A	PPB	29 Oct 2018 13:37:01	U	SMPL
CCV02 - 1	CCV02	18577	4.8420	-7470A	PPB	29 Oct 2018 13:39:08	U	SMPL
CCB02 - 1	CCB02	394	0.0358	-7470A	PPB	29 Oct 2018 13:41:15	U	SMPL
J5656-01 - 1	WS-DRUM	621	0.0958	-7470A	PPB	29 Oct 2018 13:43:21	U	SMPL
J5659-01 - 1	OIL01	62	-0.0519	-7470A	PPB	29 Oct 2018 13:45:28	U	SMPL
J5204-12 - 1	SU-02-102518-B	151	-0.0284	-7470A	PPB	29 Oct 2018 13:47:35	U	SMPL
J5204-14 - 1	SU-01-102518-A	267	0.0023	-7470A	PPB	29 Oct 2018 13:49:42	U	SMPL
J5204-16 - 1	SU-01-102518-B	216	-0.0112	-7470A	PPB	29 Oct 2018 13:51:49	U	SMPL
J5205-06 - 1	TR-05-102518	188	-0.0186	-7470A	PPB	29 Oct 2018 13:53:57	U	SMPL
J5665-11 - 1	DL-06B	302	0.0115	-7470A	PPB	29 Oct 2018 13:56:05	U	SMPL
J5665-12 - 1	DL-08A	163	-0.0252	-7470A	PPB	29 Oct 2018 13:58:13	U	SMPL
J5666-01 - 1	BS-OILY-DEBRI	172	-0.0229	-7470A	PPB	29 Oct 2018 14:00:21	U	SMPL
CCV03 - 1	CCV03	18635	4.8574	-7470A	PPB	29 Oct 2018 14:02:28	U	SMPL
CCB03 - 1	CCB03	407	0.0393	-7470A	PPB	29 Oct 2018 14:04:35	U	SMPL
J5666-02 - 1	BS-OIL	394	0.0358	-7470A	PPB	29 Oct 2018 14:06:42	U	SMPL
J5675-04 - 1	TP-5	117	-0.0374	-7470A	PPB	29 Oct 2018 14:08:49	U	SMPL
J5675-08 - 1	TP-13	147	-0.0295	-7470A	PPB	29 Oct 2018 14:10:57	U	SMPL
J5676-01 - 1	CB-1018-S-TCLP-GRID40	163	-0.0252	-7470A	PPB	29 Oct 2018 14:13:04	U	SMPL
J5683-03 - 1	SP-SAMPLE	1	-0.0681	-7470A	PPB	29 Oct 2018 14:15:11	U	SMPL
J5688-03 - 1	DPC-91567-OD-01	-10	-0.0710	-7470A	PPB	29 Oct 2018 14:17:18	U	SMPL
PB114264BL - 1	PBW	192	-0.0176	-7470A	PPB	29 Oct 2018 14:19:26	U	SMPL
PB114264BS - 1	LCSW	15391	3.9999	-7470A	PPB	29 Oct 2018 14:21:34	U	SMPL
PB114264TB - 1	PB114194TB	259	0.0001	-7470A	PPB	29 Oct 2018 14:23:43	U	SMPL
CCV04 - 1	CCV04	19865	5.1825	-7470A	PPB	29 Oct 2018 14:25:51	U	SMPL
CCB04 - 1	CCB04	48	-0.0556	-7470A	PPB	29 Oct 2018 14:27:58	U	SMPL
J5639-13LX5 - 1		13	-0.0649	-7470A	PPB	29 Oct 2018 14:30:06	U	SMPL
J5639-13A - 1		14569	3.7826	-7470A	PPB	29 Oct 2018 14:32:13	U	SMPL
J5204-10 - 1	SU-02-102518-A	144	-0.0303	-7470A	PPB	29 Oct 2018 14:34:20	U	SMPL
J5204-10DUP - 1	SU-02-102518-ADUP	211	-0.0125	-7470A	PPB	29 Oct 2018 14:36:27	U	SMPL
J5204-10MS - 1	SU-02-102518-AMS	16782	4.3676	-7470A	PPB	29 Oct 2018 14:38:35	U	SMPL
J5204-10MSD - 1	SU-02-102518-AMSD	15852	4.1218	-7470A	PPB	29 Oct 2018 14:40:42	U	SMPL
J5204-10LX5 - 1		32	-0.0599	-7470A	PPB	29 Oct 2018 14:42:49	U	SMPL
J5204-10A - 1		15073	3.9159	-7470A	PPB	29 Oct 2018 14:44:57	U	SMPL
CCV05 - 1	CCV05	19609	5.1148	-7470A	PPB	29 Oct 2018 14:47:05	U	SMPL
CCB05 - 1	CCB05	32	-0.0599	-7470A	PPB	29 Oct 2018 14:49:14	U	SMPL