



## Analytical Summary Report

Analysis Method: 180.1

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB98900

| Reagent/Standard          | Lot/Loq # |
|---------------------------|-----------|
| TURBIDITY CCV 10 NTU      | WP69169   |
| Turbidity std, 40NTU      | WP69161   |
| turbidity std 5 NTU       | WP69164   |
| turbidity std 20 NTU      | WP69162   |
| Turbidity std 1 NTU       | WP69165   |
| turbidity std 0 NTU       | WP69166   |
| turbidity std 10 NTU      | WP69163   |
| TURBIDITY LOD STD, 0.5NTU | WP69167   |
| TURBIDITY LOQ STD, 2.0NTU | WP69168   |

Intercept: 0.0865

Slope: 0.9567

Regression: 0.9997

| Seq | Lab ID | True Value (NTU) | Dilution | Reading | Anal Date  | Anal Time |
|-----|--------|------------------|----------|---------|------------|-----------|
| 1   | CAL1   | 0                | 1        | 0.236   | 10/29/2018 | 11:15     |
| 2   | CAL2   | 1                | 1        | 0.995   | 10/29/2018 | 11:19     |
| 3   | CAL3   | 5                | 1        | 4.880   | 10/29/2018 | 11:24     |
| 4   | CAL4   | 10               | 1        | 9.920   | 10/29/2018 | 11:29     |
| 5   | CAL5   | 20               | 1        | 18.600  | 10/29/2018 | 11:34     |
| 6   | CAL6   | 40               | 1        | 38.600  | 10/29/2018 | 11:39     |



## Analytical Summary Report

Analysis Method: 180.1

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB98900

| Seq | Lab ID    | True Value<br>(NTU) | Dilution | Reading | Result<br>(NTU) | AnalDate   | Anal Time |
|-----|-----------|---------------------|----------|---------|-----------------|------------|-----------|
| 1   | ICV       | 10                  | 1        | 9.320   | 9.651           | 10/29/2018 | 11:45     |
| 2   | ICB       |                     | 1        | 0.238   | 0.158           | 10/29/2018 | 11:49     |
| 3   | CCV1      | 10                  | 1        | 9.580   | 9.923           | 10/29/2018 | 11:53     |
| 4   | CCB1      |                     | 1        | 0.250   | 0.171           | 10/29/2018 | 11:58     |
| 5   | LB98900BL |                     | 1        | 0.262   | 0.183           | 10/29/2018 | 12:05     |
| 6   | J5164-07  |                     | 1        | 0.575   | 0.511           | 10/29/2018 | 12:09     |
| 7   | J5164-08  |                     | 1        | 2.315   | 2.329           | 10/29/2018 | 12:16     |
| 8   | J5164-09  |                     | 1        | 0.568   | 0.503           | 10/29/2018 | 12:22     |
| 9   | CCV2      | 10                  | 1        | 9.500   | 9.840           | 10/29/2018 | 12:29     |
| 10  | CCB2      |                     | 1        | 0.245   | 0.166           | 10/29/2018 | 12:35     |

# WORKLIST(Hardcopy Internal Chain)

6398900

WorkList Name : LODLOQTUR

WorkList ID : 118039

Date : 10/15/2018 1:54:38 PM

| Due Date   | Matrix | Sample   | Test      | Preservative | Customer | Storage Location | Customer Sample         | Collect Date | Method |
|------------|--------|----------|-----------|--------------|----------|------------------|-------------------------|--------------|--------|
| 09/30/2018 | Water  | J5164-07 | Turbidity | Cool 4 deg C | CHEM02   | Wet Chem L       | LOD-MDL-WATER-01-QT4-20 | 09/28/2018   | 180.1  |
| 09/30/2018 | Water  | J5164-08 | Turbidity | Cool 4 deg C | CHEM02   | Wet Chem L       | LOQ-WATER-02-QT4-2018   | 09/28/2018   | 180.1  |
| 09/30/2018 | Water  | J5164-09 | Turbidity | Cool 4 deg C | CHEM02   | Wet Chem L       | MDL-WATER-03-QT4-2018   | 09/28/2018   | 180.1  |

Date/Time \_\_\_\_\_  
 Received by: \_\_\_\_\_  
 Relinquished by: \_\_\_\_\_

Date/Time \_\_\_\_\_  
 Received by: \_\_\_\_\_  
 Relinquished by: \_\_\_\_\_