

LB 98913


245.1

INSTRUMENT ID: CVJ

Linear

 μ Abs.:

36539



A= 0.0000e+000

B= 2.7275e-004

C= -3.7423e-002

Rho= 0.9998670

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.000	0.000	139	0.000	139				
0.05	0.050	0.056	0.006	341	0.0 %	341				
0.2	0.200	0.193	-0.007	845	0.0 %	845				
2.5	2.500	2.417	-0.083	9000	0.0 %	9000				
5.0	5.000	5.106	0.106	18859	0.0 %	18859				
7.5	7.500	7.548	0.048	27812	0.0 %	27812				
10.0	10.000	9.929	-0.071	36539	0.0 %	36539				

LB98913 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 29 Oct 2018 17:10:06

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	58	139	-	0.0000	245.1	PPB	29 Oct 2018 17:11:45	S	Std
0.05 - 1	50.5	341	-	0.0500	245.1	PPB	29 Oct 2018 17:14:03	S	Std
0.2 - 1	50.2	845	-	0.2000	245.1	PPB	29 Oct 2018 17:16:21	S	Std
2.5 - 1	52.5	9000	-	2.5000	245.1	PPB	29 Oct 2018 17:18:39	S	Std
5.0 - 1	55	18859	-	5.0000	245.1	PPB	29 Oct 2018 17:20:56	S	Std
7.5 - 1	57.5	27812	-	7.5000	245.1	PPB	29 Oct 2018 17:23:13	S	Std
10.0 - 1	510	36539	-	10.0000	245.1	PPB	29 Oct 2018 17:25:29	S	Std
ICV45 - 1	ICV45	14941	4.0378	-245.1		PPB	29 Oct 2018 17:30:45	U	SMPL
ICB45 - 1	ICB45	17	-0.0328	-245.1		PPB	29 Oct 2018 17:33:00	U	SMPL
CCV10 - 1	CCV10	18585	5.0317	-245.1		PPB	29 Oct 2018 17:35:16	U	SMPL
CCB10 - 1	CCB10	26	-0.0303	-245.1		PPB	29 Oct 2018 17:37:31	U	SMPL
CRA - 1	CRA	329	0.0523	-245.1		PPB	29 Oct 2018 17:39:47	U	SMPL
HighStd - 1	HighStd	36754	9.9873	-245.1		PPB	29 Oct 2018 17:42:03	U	SMPL
ChkStd - 1	ChkStd	25419	6.8957	-245.1		PPB	29 Oct 2018 17:44:19	U	SMPL
PB114175BL - 1	PBW	-118	-0.0696	-245.1		PPB	29 Oct 2018 17:46:37	U	SMPL
PB114175BS - 1	LCSW	14313	3.8665	-245.1		PPB	29 Oct 2018 17:48:55	U	SMPL
J5623-01 - 1	101918-SIDI-F-D-B	39	-0.0268	-245.1		PPB	29 Oct 2018 17:51:12	U	SMPL
J5623-01DUP - 1	101918-SIDI-F-D-BDUP	77	-0.0164	-245.1		PPB	29 Oct 2018 17:53:29	U	SMPL
J5623-01MS - 1	101918-SIDI-F-D-BMS	16636	4.5001	-245.1		PPB	29 Oct 2018 17:55:47	U	SMPL
J5623-01MSD - 1	101918-SIDI-F-D-BMSD	16473	4.4556	-245.1		PPB	29 Oct 2018 17:58:05	U	SMPL
J5623-02 - 1	101918-SIDI-F-D-M	-58	-0.0532	-245.1		PPB	29 Oct 2018 18:00:21	U	SMPL
CCV11 - 1	CCV11	18470	5.0003	-245.1		PPB	29 Oct 2018 18:02:37	U	SMPL
CCB11 - 1	CCB11	8	-0.0352	-245.1		PPB	29 Oct 2018 18:04:53	U	SMPL
J5623-03 - 1	101918-SIDI-F-D-S	-49	-0.0508	-245.1		PPB	29 Oct 2018 18:07:09	U	SMPL
J5623-04 - 1	101918-SIDI-F-U-B	45	-0.0251	-245.1		PPB	29 Oct 2018 18:09:25	U	SMPL
J5623-05 - 1	101918-SIDI-F-U-M	3	-0.0366	-245.1		PPB	29 Oct 2018 18:11:41	U	SMPL
J5623-06 - 1	101918-SIDI-F-U-S	-25	-0.0442	-245.1		PPB	29 Oct 2018 18:13:57	U	SMPL
J5623-07 - 1	101918-TIPC-E-D-M	6	-0.0358	-245.1		PPB	29 Oct 2018 18:16:14	U	SMPL
J5623-08 - 1	101918-TIPC-E-U-M	-28	-0.0451	-245.1		PPB	29 Oct 2018 18:18:31	U	SMPL
J5623-09 - 1	101918-DBRI-E-D-B	27	-0.0301	-245.1		PPB	29 Oct 2018 18:20:49	U	SMPL
J5623-10 - 1	101918-DBRI-E-D-M	-7	-0.0393	-245.1		PPB	29 Oct 2018 18:23:06	U	SMPL
J5623-11 - 1	101918-DBRI-E-D-S	6	-0.0358	-245.1		PPB	29 Oct 2018 18:25:24	U	SMPL
J5623-12 - 1	101918-DBRI-E-U-B	-7	-0.0393	-245.1		PPB	29 Oct 2018 18:27:40	U	SMPL
CCV12 - 1	CCV12	18435	4.9908	-245.1		PPB	29 Oct 2018 18:29:56	U	SMPL
CCB12 - 1	CCB12	-29	-0.0453	-245.1		PPB	29 Oct 2018 18:32:13	U	SMPL
J5623-13 - 1	101918-DBRI-E-U-M	-51	-0.0513	-245.1		PPB	29 Oct 2018 18:34:31	U	SMPL
J5623-14 - 1	101918-DBRI-E-U-S	37	-0.0273	-245.1		PPB	29 Oct 2018 18:36:47	U	SMPL
J5650-01 - 1	102218-SIDI-E-U-M	64	-0.0200	-245.1		PPB	29 Oct 2018 18:39:03	U	SMPL
J5650-02 - 1	102218-SIDI-E-U-S	32	-0.0287	-245.1		PPB	29 Oct 2018 18:41:19	U	SMPL
J5623-01LX5 - 1		74	-0.0172	-245.1		PPB	29 Oct 2018 18:43:36	U	SMPL
J5623-01A - 1		15299	4.1354	-245.1		PPB	29 Oct 2018 18:45:53	U	SMPL
CCV13 - 1	CCV13	18257	4.9422	-245.1		PPB	29 Oct 2018 18:48:11	U	SMPL
CCB13 - 1	CCB13	-12	-0.0407	-245.1		PPB	29 Oct 2018 18:50:28	U	SMPL