

LB98921


245.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

34572



A= 0.0000e+000

B= 2.8695e-004

C= -5.2461e-002

Rho= 0.9997563

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.003	-0.003	172	0.000	172				
0.05	0.050	0.041	-0.009	324	0.0 %	324				
0.2	0.200	0.187	-0.013	834	0.0 %	834				
2.5	2.500	2.473	-0.027	8801	0.0 %	8801				
5.0	5.000	5.017	0.017	17666	0.0 %	17666				
7.5	7.500	7.668	0.168	26906	0.0 %	26906				
10.0	10.000	9.868	-0.132	34572	0.0 %	34572				

LB98921 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 30 Oct 2018 11:58:34

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	172	-	0.0000	245.1	PPB	30 Oct 2018 11:59:38	S	Std
0.05 - 1	50.05	324	-	0.0500	245.1	PPB	30 Oct 2018 12:01:56	S	Std
0.2 - 1	50.2	834	-	0.2000	245.1	PPB	30 Oct 2018 12:04:14	S	Std
2.5 - 1	52.5	8801	-	2.5000	245.1	PPB	30 Oct 2018 12:06:31	S	Std
5.0 - 1	55	17666	-	5.0000	245.1	PPB	30 Oct 2018 12:08:49	S	Std
7.5 - 1	57.5	26906	-	7.5000	245.1	PPB	30 Oct 2018 12:11:05	S	Std
10.0 - 1	510	34572	-	10.0000	245.1	PPB	30 Oct 2018 12:18:46	S	Std
ICV47 - 1	ICV47	14688	4.1622	-245.1		PPB	30 Oct 2018 12:21:35	U	SMPL
ICB47 - 1	ICB47	5	-0.0510	-245.1		PPB	30 Oct 2018 12:23:50	U	SMPL
CCV21 - 1	CCV21	18205	5.1714	-245.1		PPB	30 Oct 2018 12:26:06	U	SMPL
CCB21 - 1	CCB21	-2	-0.0530	-245.1		PPB	30 Oct 2018 12:28:22	U	SMPL
CRA - 1	CRA	380	0.0566	-245.1		PPB	30 Oct 2018 12:33:15	U	SMPL
HighStd - 1	HighStd	34056	9.7198	-245.1		PPB	30 Oct 2018 12:35:30	U	SMPL
ChkStd - 1	ChkStd	23885	6.8013	-245.1		PPB	30 Oct 2018 12:37:47	U	SMPL
PB114281BL - 1	PBW	205	0.0064	-245.1		PPB	30 Oct 2018 12:40:05	U	SMPL
PB114281BS - 1	LCSW	13865	3.9261	-245.1		PPB	30 Oct 2018 12:42:22	U	SMPL
J5640-01 - 1	102218-JETT-F-D-B	291	0.0310	-245.1		PPB	30 Oct 2018 12:44:39	U	SMPL
J5640-01DUP - 1	102218-JETT-F-D-BDUP	244	0.0176	-245.1		PPB	30 Oct 2018 12:49:35	U	SMPL
J5640-01MS - 1	102218-JETT-F-D-BMS	16037	4.5493	-245.1		PPB	30 Oct 2018 12:51:53	U	SMPL
J5640-01MSD - 1	102218-JETT-F-D-BMSD	15608	4.4262	-245.1		PPB	30 Oct 2018 12:54:10	U	SMPL
J5640-02 - 1	102218-JETT-F-D-M	7	-0.0505	-245.1		PPB	30 Oct 2018 12:56:26	U	SMPL
CCV22 - 1	CCV22	18285	5.1944	-245.1		PPB	30 Oct 2018 12:58:42	U	SMPL
CCB22 - 1	CCB22	39	-0.0413	-245.1		PPB	30 Oct 2018 13:00:58	U	SMPL
J5640-03 - 1	102218-JETT-F-D-S	167	-0.0045	-245.1		PPB	30 Oct 2018 13:03:13	U	SMPL
J5640-04 - 1	102218-JETT-F-U-B	209	0.0075	-245.1		PPB	30 Oct 2018 13:08:28	U	SMPL
J5640-05 - 1	102218-JETT-F-U-M	160	-0.0065	-245.1		PPB	30 Oct 2018 13:13:19	U	SMPL
J5640-06 - 1	102218-JETT-F-U-S	146	-0.0106	-245.1		PPB	30 Oct 2018 13:15:35	U	SMPL
J5640-07 - 1	102218-TIPC-E-D-B	75	-0.0309	-245.1		PPB	30 Oct 2018 13:17:51	U	SMPL
J5640-08 - 1	102218-TIPC-E-D-S	155	-0.0080	-245.1		PPB	30 Oct 2018 13:20:08	U	SMPL
J5640-09 - 1	102218-TIPC-E-U-B	74	-0.0312	-245.1		PPB	30 Oct 2018 13:22:26	U	SMPL
J5640-10 - 1	102218-TIPC-E-U-S	137	-0.0131	-245.1		PPB	30 Oct 2018 13:24:43	U	SMPL
J5640-11 - 1	102218-DBRI-F-D-B	152	-0.0088	-245.1		PPB	30 Oct 2018 13:29:42	U	SMPL
J5640-12 - 1	102218-DBRI-F-D-M	241	0.0167	-245.1		PPB	30 Oct 2018 13:31:58	U	SMPL
CCV23 - 1	CCV23	18484	5.2515	-245.1		PPB	30 Oct 2018 13:34:14	U	SMPL
CCB23 - 1	CCB23	46	-0.0393	-245.1		PPB	30 Oct 2018 13:36:31	U	SMPL
J5640-13 - 1	102218-DBRI-F-D-S	115	-0.0195	-245.1		PPB	30 Oct 2018 13:38:47	U	SMPL
J5640-14 - 1	102218-DBRI-F-U-B	88	-0.0272	-245.1		PPB	30 Oct 2018 13:41:04	U	SMPL
J5640-15 - 1	102218-DBRI-F-U-M	83	-0.0286	-245.1		PPB	30 Oct 2018 13:43:20	U	SMPL
J5640-16 - 1	102218-DBRI-F-U-S	73	-0.0315	-245.1		PPB	30 Oct 2018 13:45:36	U	SMPL
J5640-17 - 1	102218-SIDI-E-D-B	66	-0.0335	-245.1		PPB	30 Oct 2018 13:47:53	U	SMPL
J5640-18 - 1	102218-SIDI-E-D-M	53	-0.0373	-245.1		PPB	30 Oct 2018 13:50:10	U	SMPL
J5640-19 - 1	102218-SIDI-E-D-S	62	-0.0347	-245.1		PPB	30 Oct 2018 13:52:28	U	SMPL
J5640-20 - 1	102218-SIDI-E-U-B	51	-0.0378	-245.1		PPB	30 Oct 2018 13:54:45	U	SMPL
PB114282BL - 1	PBW	283	0.0287	-245.1		PPB	30 Oct 2018 13:57:03	U	SMPL
PB114282BS - 1	LCSW	14134	4.0033	-245.1		PPB	30 Oct 2018 13:59:19	U	SMPL
CCV24 - 1	CCV24	17910	5.0868	-245.1		PPB	30 Oct 2018 14:01:35	U	SMPL
CCB24 - 1	CCB24	58	-0.0358	-245.1		PPB	30 Oct 2018 14:03:52	U	SMPL
J5617-01 - 1	PANTRY-COOLER	108	-0.0215	-245.1		PPB	30 Oct 2018 14:06:08	U	SMPL
J5617-01DUP - 1	PANTRY-COOLERDUP	125	-0.0166	-245.1		PPB	30 Oct 2018 14:08:25	U	SMPL
J5617-01MS - 1	PANTRY-COOLERMS	14086	3.9895	-245.1		PPB	30 Oct 2018 14:10:41	U	SMPL
J5617-01MSD - 1	PANTRY-COOLERMSD	14516	4.1129	-245.1		PPB	30 Oct 2018 14:12:57	U	SMPL
J5662-01 - 1	102318-SIDI-F-D-B	161	-0.0063	-245.1		PPB	30 Oct 2018 14:15:14	U	SMPL
J5662-02 - 1	102318-SIDI-F-D-M	238	0.0158	-245.1		PPB	30 Oct 2018 14:17:31	U	SMPL
J5662-03 - 1	102318-SIDI-F-D-S	121	-0.0177	-245.1		PPB	30 Oct 2018 14:19:48	U	SMPL
J5662-04 - 1	102318-SIDI-F-U-B	129	-0.0154	-245.1		PPB	30 Oct 2018 14:22:05	U	SMPL
J5662-05 - 1	102318-SIDI-F-U-M	120	-0.0180	-245.1		PPB	30 Oct 2018 14:24:24	U	SMPL
J5662-06 - 1	102318-SIDI-F-U-S	70	-0.0324	-245.1		PPB	30 Oct 2018 14:26:41	U	SMPL
CCV25 - 1	CCV25	17928	5.0919	-245.1		PPB	30 Oct 2018 14:28:57	U	SMPL
CCB25 - 1	CCB25	53	-0.0373	-245.1		PPB	30 Oct 2018 14:31:14	U	SMPL
J5662-07 - 1	102318-DBRI-F-D-B	122	-0.0175	-245.1		PPB	30 Oct 2018 14:33:30	U	SMPL
J5662-08 - 1	102318-DBRI-F-D-M	128	-0.0157	-245.1		PPB	30 Oct 2018 14:35:46	U	SMPL
J5662-09 - 1	102318-DBRI-F-D-S	123	-0.0172	-245.1		PPB	30 Oct 2018 14:38:03	U	SMPL
J5662-10 - 1	102318-DBRI-F-U-B	169	-0.0040	-245.1		PPB	30 Oct 2018 14:40:19	U	SMPL
J5662-11 - 1	102318-DBRI-F-U-M	154	-0.0083	-245.1		PPB	30 Oct 2018 14:42:36	U	SMPL
J5662-12 - 1	102318-DBRI-F-U-S	144	-0.0111	-245.1		PPB	30 Oct 2018 14:44:53	U	SMPL
J5662-13 - 1	102318-TIPC-F-D-B	141	-0.0120	-245.1		PPB	30 Oct 2018 14:47:10	U	SMPL
J5662-14 - 1	102318-TIPC-F-D-S	160	-0.0065	-245.1		PPB	30 Oct 2018 14:49:28	U	SMPL
J5662-15 - 1	102318-TIPC-F-U-B	154	-0.0083	-245.1		PPB	30 Oct 2018 14:51:45	U	SMPL
J5662-16 - 1	102318-TIPC-F-U-S	129	-0.0154	-245.1		PPB	30 Oct 2018 14:54:02	U	SMPL
CCV26 - 1	CCV26	17733	5.0360	-245.1		PPB	30 Oct 2018 14:56:18	U	SMPL
CCB26 - 1	CCB26	63	-0.0344	-245.1		PPB	30 Oct 2018 14:58:35	U	SMPL
J5665-07 - 1	TWP-02	335	0.0437	-245.1		PPB	30 Oct 2018 15:00:52	U	SMPL
J5665-09 - 1	TWP-04	240	0.0164	-245.1		PPB	30 Oct 2018 15:03:08	U	SMPL
PB114283BL - 1	PBW	218	0.0101	-245.1		PPB	30 Oct 2018 15:05:25	U	SMPL
PB114283BS - 1	LCSW	14134	4.0033	-245.1		PPB	30 Oct 2018 15:07:43	U	SMPL
J5686-01 - 1	102418-SIDI-F-D-B	246	0.0181	-245.1		PPB	30 Oct 2018 15:10:00	U	SMPL
J5686-01DUP - 1	102418-SIDI-F-D-BDUP	262	0.0227	-245.1		PPB	30 Oct 2018 15:12:17	U	SMPL
J5686-01MS - 1	102418-SIDI-F-D-BMS	16788	4.7648	-245.1		PPB	30 Oct 2018 15:14:35	U	SMPL
J5686-01MSD - 1	102418-SIDI-F-D-BMSD	16647	4.7244	-245.1		PPB	30 Oct 2018 15:16:52	U	SMPL
J5686-02 - 1	102418-SIDI-F-D-M	50	-0.0381	-245.1		PPB	30 Oct 2018 15:19:10	U	SMPL
J5686-03 - 1	102418-SIDI-F-D-S	63	-0.0344	-245.1		PPB	30 Oct 2018 15:21:27	U	SMPL

LB98921 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 30 Oct 2018 11:58:34

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
CCV27 - 1	CCV27	17582	4.9926	-245.1	PPB	30 Oct 2018 15:23:44	U	SMPL
CCB27 - 1	CCB27	2	-0.0519	-245.1	PPB	30 Oct 2018 15:26:01	U	SMPL
J5686-04 - 1	102418-SIDI-F-U-B	320	0.0394	-245.1	PPB	30 Oct 2018 15:28:18	U	SMPL
J5686-05 - 1	102418-SIDI-F-U-M	78	-0.0301	-245.1	PPB	30 Oct 2018 15:30:35	U	SMPL
J5686-06 - 1	102418-SIDI-F-U-S	91	-0.0263	-245.1	PPB	30 Oct 2018 15:32:52	U	SMPL
J5686-07 - 1	102418-DBRI-F-D-B	224	0.0118	-245.1	PPB	30 Oct 2018 15:35:10	U	SMPL
J5686-08 - 1	102418-DBRI-F-D-M	262	0.0227	-245.1	PPB	30 Oct 2018 15:37:27	U	SMPL
J5686-09 - 1	102418-DBRI-F-D-S	119	-0.0183	-245.1	PPB	30 Oct 2018 15:39:44	U	SMPL
J5686-10 - 1	102418-DBRI-F-U-B	167	-0.0045	-245.1	PPB	30 Oct 2018 15:42:01	U	SMPL
J5686-11 - 1	102418-DBRI-F-U-M	121	-0.0177	-245.1	PPB	30 Oct 2018 15:44:18	U	SMPL
J5686-12 - 1	102418-DBRI-F-U-S	71	-0.0321	-245.1	PPB	30 Oct 2018 15:46:36	U	SMPL
J5686-13 - 1	102418-TIPC-F-D-B	100	-0.0238	-245.1	PPB	30 Oct 2018 15:48:53	U	SMPL
CCV28 - 1	CCV28	17496	4.9680	-245.1	PPB	30 Oct 2018 15:51:10	U	SMPL
CCB28 - 1	CCB28	17	-0.0476	-245.1	PPB	30 Oct 2018 15:53:27	U	SMPL
J5686-14 - 1	102418-TIPC-F-D-S	209	0.0075	-245.1	PPB	30 Oct 2018 15:55:44	U	SMPL
J5686-15 - 1	102418-TIPC-F-U-B	151	-0.0091	-245.1	PPB	30 Oct 2018 15:58:01	U	SMPL
J5686-16 - 1	102418-TIPC-F-U-S	100	-0.0238	-245.1	PPB	30 Oct 2018 16:00:18	U	SMPL
PB114284BL - 1	PBW	126	-0.0163	-245.1	PPB	30 Oct 2018 16:02:35	U	SMPL
PB114284BS - 1	LCSW	13848	3.9212	-245.1	PPB	30 Oct 2018 16:04:52	U	SMPL
J5706-01 - 1	102518-SIDI-F-D-B	1001	0.2348	-245.1	PPB	30 Oct 2018 16:07:10	U	SMPL
J5706-01DUP - 1	102518-SIDI-F-D-BDUP	1066	0.2534	-245.1	PPB	30 Oct 2018 16:09:27	U	SMPL
J5706-01MS - 1	102518-SIDI-F-D-BMS	16632	4.7200	-245.1	PPB	30 Oct 2018 16:14:07	U	SMPL
J5706-01MSD - 1	102518-SIDI-F-D-BMSD	14435	4.0896	-245.1	PPB	30 Oct 2018 16:19:01	U	SMPL
J5706-02 - 1	102518-SIDI-F-D-M	961	0.2233	-245.1	PPB	30 Oct 2018 16:21:18	U	SMPL
CCV29 - 1	CCV29	17585	4.9935	-245.1	PPB	30 Oct 2018 16:23:35	U	SMPL
CCB29 - 1	CCB29	-23	-0.0591	-245.1	PPB	30 Oct 2018 16:25:52	U	SMPL
J5706-03 - 1	102518-SIDI-F-D-S	899	0.2055	-245.1	PPB	30 Oct 2018 16:28:09	U	SMPL
J5706-04 - 1	102518-SIDI-F-U-B	1077	0.2566	-245.1	PPB	30 Oct 2018 16:30:26	U	SMPL
J5706-05 - 1	102518-SIDI-F-U-M	971	0.2262	-245.1	PPB	30 Oct 2018 16:32:43	U	SMPL
J5706-06 - 1	102518-SIDI-F-U-S	925	0.2130	-245.1	PPB	30 Oct 2018 16:35:00	U	SMPL
J5706-07 - 1	102518-TIPC-F-D-B	1014	0.2385	-245.1	PPB	30 Oct 2018 16:37:17	U	SMPL
J5706-08 - 1	102518-TIPC-F-D-S	1004	0.2356	-245.1	PPB	30 Oct 2018 16:39:34	U	SMPL
J5706-09 - 1	102518-TIPC-F-U-B	1067	0.2537	-245.1	PPB	30 Oct 2018 16:41:52	U	SMPL
J5706-10 - 1	102518-TIPC-F-U-S	971	0.2262	-245.1	PPB	30 Oct 2018 16:44:09	U	SMPL
CCV30 - 1	CCV30	17594	4.9961	-245.1	PPB	30 Oct 2018 17:13:38	U	SMPL
CCB30 - 1	CCB30	437	0.0729	-245.1	PPB	30 Oct 2018 17:18:33	U	SMPL
J5640-01A - 1		16211	4.5992	-245.1	PPB	30 Oct 2018 17:31:14	U	SMPL
J5686-01LX5 - 1		147	-0.0103	-245.1	PPB	30 Oct 2018 18:15:36	U	SMPL
J5686-01A - 1		15322	4.3441	-245.1	PPB	30 Oct 2018 18:17:53	U	SMPL
CCV31 - 1	CCV31	15976	4.5318	-245.1	PPB	30 Oct 2018 18:20:10	U	SMPL
CCB31 - 1	CCB31	24	-0.0456	-245.1	PPB	30 Oct 2018 18:22:27	U	SMPL
J5706-01LX5 - 1		-59	-0.0694	-245.1	PPB	30 Oct 2018 18:24:45	U	SMPL
J5706-01A - 1		14153	4.0087	-245.1	PPB	30 Oct 2018 18:27:03	U	SMPL
J5640-01LX5 - 1		-93	-0.0791	-245.1	PPB	30 Oct 2018 18:29:21	U	SMPL
J5617-01LX5 - 1		-42	-0.0645	-245.1	PPB	30 Oct 2018 18:34:50	U	SMPL
J5617-01A - 1		13562	3.8391	-245.1	PPB	30 Oct 2018 18:37:07	U	SMPL
CCV32 - 1	CCV32	17580	4.9921	-245.1	PPB	30 Oct 2018 18:39:25	U	SMPL
CCB32 - 1	CCB32	-54	-0.0680	-245.1	PPB	30 Oct 2018 18:41:43	U	SMPL