



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB98922

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.0	80.6	80.6	10/30/2018	12:20
2	J5718-02		1	9	>100.0	>212.0	>212.0	10/30/2018	12:45
3	J5718-02DUP		1	9	>100.0	>212.0	>212.0	10/30/2018	13:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 98922

WorkList Name : fp103018

WorkList ID : 118632

Date : 10/30/2018 11:42:01 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/01/2018	Water	J5718-02	C Flash Point	Cool 4 deg C	PSEG01	P63	WS-CFA-111C	10/29/2018	1010A

Date/Time 10/30/18 11:45 AM
 Received by: KS
 Relinquished by: RJ

Date/Time 10/30/18 13:20
 Received by: RJ
 Relinquished by: KS