



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: jignesh

Date: 10/31/2018

TEMP1 IN: 103 10/31/2018 07:10 TEMP OUT: 103 10/31/2018 08:10

TEMP2 IN: 103 10/31/2018 08:35 TEMP OUT: 103 10/31/2018 09:35

TEMP3 IN: 103 10/31/2018 11:10 TEMP OUT: 104 10/31/2018 12:35

TEMP4 IN: 104 10/31/2018 13:10 TEMP OUT: 104 10/31/2018 15:40

Run Number: LB98942

BalanceID: WC SC-3

OvenID: WC OVEN-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB98942BL	LB98942BL	2.6997	2.6997	100	2.6998	2.6998	2.6998	0.0001	1
2	LB98942BS	LB98942BS	2.9412	2.9412	100	2.9958	2.9958	2.9958	0.0546	546
3	J5728-03	EG-WATER	1.4033	1.4033	300	1.4752	1.4752	1.4752	0.0719	239.7
4	J5728-03DUP	EG-WATERDUP	1.3930	1.3930	300	1.4680	1.4680	1.4680	0.0750	250
5	J5730-01	102918-TIPC-E-D-M	1.4084	1.4084	400	1.4226	1.4226	1.4226	0.0142	35.5
6	J5730-02	102918-TIPC-E-U-M	1.4013	1.4013	400	1.4134	1.4134	1.4134	0.0121	30.3
7	J5730-03	102918-SIDI-E-D-B	1.4144	1.4144	400	1.4254	1.4254	1.4254	0.0110	27.5
8	J5730-04	102918-SIDI-E-D-M	1.4064	1.4064	400	1.4135	1.4135	1.4135	0.0071	17.8
9	J5730-05	102918-SIDI-E-D-S	1.4011	1.4011	400	1.4053	1.4053	1.4053	0.0042	10.5
10	J5730-06	102918-SIDI-E-U-B	1.4017	1.4017	400	1.4117	1.4117	1.4117	0.0100	25
11	J5730-07	102918-SIDI-E-U-M	1.4003	1.4003	400	1.4067	1.4067	1.4067	0.0064	16
12	J5730-08	102918-SIDI-E-U-S	1.4217	1.4217	400	1.4264	1.4264	1.4264	0.0047	11.8
13	J5731-01	103018-TIPC-E-D-M	1.4168	1.4168	400	1.4333	1.4333	1.4333	0.0165	41.3
14	J5731-02	103018-TIPC-E-U-M	1.4008	1.4008	400	1.4146	1.4146	1.4146	0.0138	34.5
15	J5731-03	103018-SIDI-E-D-B	1.4025	1.4025	400	1.4163	1.4163	1.4163	0.0138	34.5



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: jignesh

Date: 10/31/2018

TEMP1 IN: 103 10/31/2018 07:10 TEMP OUT: 103 10/31/2018 08:10

TEMP2 IN: 103 10/31/2018 08:35 TEMP OUT: 103 10/31/2018 09:35

TEMP3 IN: 103 10/31/2018 11:10 TEMP OUT: 104 10/31/2018 12:35

TEMP4 IN: 104 10/31/2018 13:10 TEMP OUT: 104 10/31/2018 15:40

Run Number: LB98942

BalanceID: WC SC-3

OvenID: WC OVEN-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	J5731-04	103018-SIDI-E-D-M	1.4001	1.4001	400	1.4102	1.4102	1.4102	0.0101	25.3
17	J5731-05	103018-SIDI-E-D-S	1.3965	1.3965	400	1.4068	1.4068	1.4068	0.0103	25.8
18	J5731-06	103018-SIDI-E-U-B	1.3934	1.3934	400	1.4047	1.4047	1.4047	0.0113	28.3
19	J5731-07	103018-SIDI-E-U-M	1.4022	1.4022	400	1.4094	1.4094	1.4094	0.0072	18
20	J5731-08	103018-SIDI-E-U-S	1.3971	1.3971	400	1.4037	1.4037	1.4037	0.0066	16.5
21	J5731-09	103018-DBRI-E-D-B	1.4049	1.4049	400	1.4136	1.4136	1.4136	0.0087	21.8
22	J5731-10	103018-DBRI-E-D-M	1.4088	1.4088	400	1.4159	1.4159	1.4159	0.0071	17.8
23	J5731-11	103018-DBRI-E-D-S	1.4045	1.4045	400	1.4088	1.4088	1.4088	0.0043	10.8
24	J5731-12	103018-DBRI-E-U-B	1.3959	1.3959	100	1.4052	1.4052	1.4052	0.0093	93
25	J5731-13	103018-DBRI-E-U-M	1.4096	1.4096	400	1.4163	1.4163	1.4163	0.0067	16.8
26	J5731-14	103018-DBRI-E-U-S	1.3950	1.3950	400	1.4006	1.4006	1.4006	0.0056	14
27	J5732-01	STORM-DRAIN-1	1.4085	1.4085	900	1.4089	1.4089	1.4089	0.0004	0.4
28	J5732-02	STORM-DRAIN-2	1.4010	1.4010	400	1.4013	1.4013	1.4013	0.0003	0.7
29	J5732-02DUP	STORM-DRAIN-2DUP	1.3959	1.3959	400	1.3964	1.3964	1.3964	0.0005	1.3



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: jignesh

Date: 10/31/2018

TEMP1 IN: 103 10/31/2018 07:10 TEMP OUT: 103 10/31/2018 08:10

TEMP2 IN: 103 10/31/2018 08:35 TEMP OUT: 103 10/31/2018 09:35

TEMP3 IN: 103 10/31/2018 11:10 TEMP OUT: 104 10/31/2018 12:35

TEMP4 IN: 104 10/31/2018 13:10 TEMP OUT: 104 10/31/2018 15:40

Run Number: LB98942

BalanceID: WC SC-3

OvenID: WC OVEN-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

LB98942

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS J5732

WorkList ID : 118652

Date : 10/31/2018 7:51:27 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/02/2018	Water	J5728-03 -F	TSS	Cool 4 deg C	PSEG01	I11	EG-WATER	10/30/2018	SM2540 D
11/01/2018	Water	J5730-01 -C	TSS	Cool 4 deg C	HDR103	I21	102918-TIPC-E-D-M	10/29/2018	SM2540 D
11/01/2018	Water	J5730-02 -C	TSS	Cool 4 deg C	HDR103	I21	102918-TIPC-E-U-M	10/29/2018	SM2540 D
11/05/2018	Water	J5730-03 -C	TSS	Cool 4 deg C	HDR103	I21	102918-SIDI-E-D-B	10/29/2018	SM2540 D
11/05/2018	Water	J5730-04 -C	TSS	Cool 4 deg C	HDR103	I21	102918-SIDI-E-D-M	10/29/2018	SM2540 D
11/05/2018	Water	J5730-05 -C	TSS	Cool 4 deg C	HDR103	I21	102918-SIDI-E-D-S	10/29/2018	SM2540 D
11/05/2018	Water	J5730-06 -C	TSS	Cool 4 deg C	HDR103	I21	102918-SIDI-E-U-B	10/29/2018	SM2540 D
11/05/2018	Water	J5730-07 -C	TSS	Cool 4 deg C	HDR103	I21	102918-SIDI-E-U-M	10/29/2018	SM2540 D
11/05/2018	Water	J5730-08 -C	TSS	Cool 4 deg C	HDR103	I21	102918-SIDI-E-U-S	10/29/2018	SM2540 D
11/01/2018	Water	J5731-01 -C	TSS	Cool 4 deg C	HDR103	I23	103018-TIPC-E-D-M	10/30/2018	SM2540 D
11/01/2018	Water	J5731-02 -C	TSS	Cool 4 deg C	HDR103	I23	103018-TIPC-E-U-M	10/30/2018	SM2540 D
11/06/2018	Water	J5731-03 -C	TSS	Cool 4 deg C	HDR103	I23	103018-SIDI-E-D-B	10/30/2018	SM2540 D
11/06/2018	Water	J5731-04 -C	TSS	Cool 4 deg C	HDR103	I23	103018-SIDI-E-D-M	10/30/2018	SM2540 D
11/06/2018	Water	J5731-05 -C	TSS	Cool 4 deg C	HDR103	I23	103018-SIDI-E-D-S	10/30/2018	SM2540 D
11/06/2018	Water	J5731-06 -C	TSS	Cool 4 deg C	HDR103	I23	103018-SIDI-E-U-B	10/30/2018	SM2540 D
11/06/2018	Water	J5731-07 -C	TSS	Cool 4 deg C	HDR103	I23	103018-SIDI-E-U-M	10/30/2018	SM2540 D
11/06/2018	Water	J5731-08 -C	TSS	Cool 4 deg C	HDR103	I23	103018-SIDI-E-U-S	10/30/2018	SM2540 D
11/06/2018	Water	J5731-09 -C	TSS	Cool 4 deg C	HDR103	I23	103018-DBRI-E-D-B	10/30/2018	SM2540 D
11/06/2018	Water	J5731-10 -C	TSS	Cool 4 deg C	HDR103	I23	103018-DBRI-E-D-M	10/30/2018	SM2540 D
11/06/2018	Water	J5731-11 -C	TSS	Cool 4 deg C	HDR103	I23	103018-DBRI-E-D-S	10/30/2018	SM2540 D
11/06/2018	Water	J5731-12 -C	TSS	Cool 4 deg C	HDR103	I23	103018-DBRI-E-U-B	10/30/2018	SM2540 D

Date/Time 10-31-18 8:40 AM

Received by: 20

Relinquished by: DP

Date/Time 10-31-18 11:50 AM

Received by: DP

Relinquished by: 20

1898942

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS J5732

WorkList ID : 118652

Date : 10/31/2018 7:51:27 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/06/2018	Water	J5731-13 -C	TSS	Cool 4 deg C	HDR103	I23	103018-DBRIE-U-M	10/30/2018	SM2540 D
11/06/2018	Water	J5731-14 -C	TSS	Cool 4 deg C	HDR103	I23	103018-DBRIE-U-S	10/30/2018	SM2540 D
11/06/2018	Water	J5732-01 -C	TSS	Cool 4 deg C	KWAR01	I41	STORM-DRAIN-1	10/30/2018	SM2540 D
11/06/2018	Water	J5732-02 -C	TSS	Cool 4 deg C	KWAR01	I41	STORM-DRAIN-2	10/30/2018	SM2540 D

Date/Time 10-31-18 8:00 AM
 Received by: JP
 Relinquished by: JP

Date/Time 10-31-18 11:30 AM
 Received by: JP
 Relinquished by: JP