



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: sweta

Run Number: LB98944

Ambient Barometric Pressure (mmHg): 740

Barometric Scale ID: 55971

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	81.8	10/31/2018	09:45
2	J5728-04		1	9	>100.0	>212.0	>212.0	10/31/2018	10:20
3	J5728-04DUP		1	9	>100.0	>212.0	>212.0	10/31/2018	10:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB98944

WorkList Name : fp103118

WorkList ID : 118665

Date : 10/31/2018 9:11:32 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/02/2018	Water	J5728-04	C Flash Point	Cool 4 deg C	PSEG01	I11	EG-WATER	10/30/2018	1010A

Date/Time 10/31/18 9:15 AM
 Received by: KS
 Relinquished by: CP

Date/Time 10/31/18 10:55 AM
 Received by: CP
 Relinquished by: KS