


LB 98961

7470A

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.8371e-004

C= 3.9422e-002

Rho= 0.9998098

Accept=Accepted

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|
| 0.0    | 0.000  | 0.070  | 0.070  | 108   | 0.000      | 108   |       |       |       |       |
| 0.2    | 0.200  | 0.232  | 0.032  | 678   | 0.0 %      | 678   |       |       |       |       |
| 2.5    | 2.500  | 2.414  | -0.086 | 8371  | 0.0 %      | 8371  |       |       |       |       |
| 5.0    | 5.000  | 4.977  | -0.023 | 17404 | 0.0 %      | 17404 |       |       |       |       |
| 7.5    | 7.500  | 7.410  | -0.090 | 25981 | 0.0 %      | 25981 |       |       |       |       |
| 10.0   | 10.000 | 10.096 | 0.096  | 35448 | 0.0 %      | 35448 |       |       |       |       |

## LB98961 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 01 Nov 2018 09:30:36

| Sample ID       | Extended ID       | $\mu$ Abs. | Conc.  | Std Conc | Method | Units                | Date                 | Type | Type |
|-----------------|-------------------|------------|--------|----------|--------|----------------------|----------------------|------|------|
| 0.0 - 1         | 50                | 108        | -      | 0.0000   | 7470A  | PPB                  | 01 Nov 2018 09:36:18 | S    | Std  |
| 0.2 - 1         | 50.2              | 678        | -      | 0.2000   | 7470A  | PPB                  | 01 Nov 2018 09:38:26 | S    | Std  |
| 2.5 - 1         | 52.5              | 8371       | -      | 2.5000   | 7470A  | PPB                  | 01 Nov 2018 09:40:35 | S    | Std  |
| 5.0 - 1         | 55                | 17404      | -      | 5.0000   | 7470A  | PPB                  | 01 Nov 2018 09:42:43 | S    | Std  |
| 7.5 - 1         | 57.5              | 25981      | -      | 7.5000   | 7470A  | PPB                  | 01 Nov 2018 09:44:51 | S    | Std  |
| 10.0 - 1        | 58                | 35448      | -      | 10.0000  | 7470A  | PPB                  | 01 Nov 2018 09:51:48 | S    | Std  |
| ICV50 - 1       | ICV50             | 13918      | 3.9881 | -7470A   | PPB    | 01 Nov 2018 09:55:20 | U                    | SMPL |      |
| ICB50 - 1       | ICB50             | -102       | 0.0105 | -7470A   | PPB    | 01 Nov 2018 09:57:26 | U                    | SMPL |      |
| CCV44 - 1       | CCV44             | 17790      | 5.0866 | -7470A   | PPB    | 01 Nov 2018 09:59:33 | U                    | SMPL |      |
| CCB44 - 1       | CCB44             | -109       | 0.0085 | -7470A   | PPB    | 01 Nov 2018 10:01:40 | U                    | SMPL |      |
| CRA - 1         | CRA               | 648        | 0.2233 | -7470A   | PPB    | 01 Nov 2018 10:03:46 | U                    | SMPL |      |
| HighStd - 1     | HighStd           | 35062      | 9.9868 | -7470A   | PPB    | 01 Nov 2018 10:05:53 | U                    | SMPL |      |
| ChkStd - 1      | ChkStd            | 25061      | 7.1494 | -7470A   | PPB    | 01 Nov 2018 10:08:00 | U                    | SMPL |      |
| PB114363BL - 1  | PBW               | -81        | 0.0164 | -7470A   | PPB    | 01 Nov 2018 10:10:08 | U                    | SMPL |      |
| PB114363BS - 1  | LCSW              | 14400      | 4.1248 | -7470A   | PPB    | 01 Nov 2018 10:12:17 | U                    | SMPL |      |
| J5199-08 - 1    | OK-01-103018      | -84        | 0.0156 | -7470A   | PPB    | 01 Nov 2018 10:14:24 | U                    | SMPL |      |
| J5202-08 - 1    | PL-02-103018-A    | -9         | 0.0369 | -7470A   | PPB    | 01 Nov 2018 10:16:32 | U                    | SMPL |      |
| J5202-08DUP - 1 | PL-02-103018-ADUP | 6          | 0.0411 | -7470A   | PPB    | 01 Nov 2018 10:18:40 | U                    | SMPL |      |
| J5202-08MS - 1  | PL-02-103018-AMS  | 13813      | 3.9583 | -7470A   | PPB    | 01 Nov 2018 10:20:49 | U                    | SMPL |      |
| J5202-08MSD - 1 | PL-02-103018-AMSD | 15658      | 4.4817 | -7470A   | PPB    | 01 Nov 2018 10:22:55 | U                    | SMPL |      |
| CCV45 - 1       | CCV45             | 18364      | 5.2494 | -7470A   | PPB    | 01 Nov 2018 10:25:02 | U                    | SMPL |      |
| CCB45 - 1       | CCB45             | -59        | 0.0227 | -7470A   | PPB    | 01 Nov 2018 10:27:10 | U                    | SMPL |      |
| J5202-10 - 1    | PL-02-103018-B    | -49        | 0.0255 | -7470A   | PPB    | 01 Nov 2018 10:29:16 | U                    | SMPL |      |
| J5202-12 - 1    | PL-02-103018-C    | -4         | 0.0383 | -7470A   | PPB    | 01 Nov 2018 10:31:23 | U                    | SMPL |      |
| J5202-14 - 1    | PL-03-103018-A    | 4          | 0.0406 | -7470A   | PPB    | 01 Nov 2018 10:33:29 | U                    | SMPL |      |
| J5202-16 - 1    | PL-03-103018-B    | 14         | 0.0434 | -7470A   | PPB    | 01 Nov 2018 10:35:36 | U                    | SMPL |      |
| J5205-08 - 1    | TR-04-103018-A    | 19         | 0.0448 | -7470A   | PPB    | 01 Nov 2018 10:37:44 | U                    | SMPL |      |
| J5205-10 - 1    | TR-04-103018-B    | 43         | 0.0516 | -7470A   | PPB    | 01 Nov 2018 10:39:52 | U                    | SMPL |      |
| J5377-03 - 1    | RBR-200036        | 22         | 0.0457 | -7470A   | PPB    | 01 Nov 2018 10:42:00 | U                    | SMPL |      |
| J5728-02 - 1    | EG-DRUM           | 24         | 0.0462 | -7470A   | PPB    | 01 Nov 2018 10:44:08 | U                    | SMPL |      |
| J5737-02 - 1    | TP-A              | 26         | 0.0468 | -7470A   | PPB    | 01 Nov 2018 10:46:17 | U                    | SMPL |      |
| J5737-04 - 1    | TP-B              | 36         | 0.0496 | -7470A   | PPB    | 01 Nov 2018 10:48:24 | U                    | SMPL |      |
| CCV46 - 1       | CCV46             | 17880      | 5.1121 | -7470A   | PPB    | 01 Nov 2018 10:50:32 | U                    | SMPL |      |
| CCB46 - 1       | CCB46             | 5          | 0.0408 | -7470A   | PPB    | 01 Nov 2018 10:52:39 | U                    | SMPL |      |
| PB114367BL - 1  | PBW               | -44        | 0.0269 | -7470A   | PPB    | 01 Nov 2018 11:01:30 | U                    | SMPL |      |
| PB114367BS - 1  | LCSW              | 13212      | 3.7878 | -7470A   | PPB    | 01 Nov 2018 11:03:36 | U                    | SMPL |      |
| J5728-03 - 1    | EG-WATER          | 207        | 0.0981 | -7470A   | PPB    | 01 Nov 2018 11:05:44 | U                    | SMPL |      |
| J5728-03DUP - 1 | EG-WATERDUP       | 252        | 0.1109 | -7470A   | PPB    | 01 Nov 2018 11:07:50 | U                    | SMPL |      |
| J5728-03MS - 1  | EG-WATERMS        | 13810      | 3.9574 | -7470A   | PPB    | 01 Nov 2018 11:09:58 | U                    | SMPL |      |
| J5728-03MSD - 1 | EG-WATERMSD       | 12870      | 3.6907 | -7470A   | PPB    | 01 Nov 2018 11:12:05 | U                    | SMPL |      |
| PB114363TB - 1  | PB114327TB        | -26        | 0.0320 | -7470A   | PPB    | 01 Nov 2018 11:14:13 | U                    | SMPL |      |
| CCV47 - 1       | CCV47             | 17722      | 5.0673 | -7470A   | PPB    | 01 Nov 2018 11:16:22 | U                    | SMPL |      |
| CCB47 - 1       | CCB47             | 16         | 0.0440 | -7470A   | PPB    | 01 Nov 2018 11:21:03 | U                    | SMPL |      |
| J5202-08LX5 - 1 |                   | 64         | 0.0576 | -7470A   | PPB    | 01 Nov 2018 11:23:10 | U                    | SMPL |      |
| J5202-08A - 1   |                   | 13264      | 3.8025 | -7470A   | PPB    | 01 Nov 2018 11:25:17 | U                    | SMPL |      |
| J5728-03LX5 - 1 |                   | 39         | 0.0505 | -7470A   | PPB    | 01 Nov 2018 11:29:58 | U                    | SMPL |      |
| J5728-03A - 1   |                   | 14580      | 4.1759 | -7470A   | PPB    | 01 Nov 2018 11:32:05 | U                    | SMPL |      |
| CCV48 - 1       | CCV48             | 17657      | 5.0489 | -7470A   | PPB    | 01 Nov 2018 11:34:12 | U                    | SMPL |      |
| CCB48 - 1       | CCB48             | -53        | 0.0244 | -7470A   | PPB    | 01 Nov 2018 11:36:19 | U                    | SMPL |      |