



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: Corrosivity

Supervisor Review By : JIGNESH

Run Number: LB98962

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.2	4.01	10/31/2018	18:20
2	CAL2	1	Water			21.2	7.00	10/31/2018	18:22
3	CAL3	1	Water			21.3	10.04	10/31/2018	18:24
4	ICV	1	Water			21.4	7.01	10/31/2018	18:26
5	CCV1	1	Water			21.3	2.00	10/31/2018	18:28
6	J5729-04	1	Water			16.2	6.80	10/31/2018	18:30
7	J5729-04DUP	1	Water			16.3	6.82	10/31/2018	18:32
8	J5755-14	1	Water			21.1	7.23	10/31/2018	18:34
9	CCV2	1	Water			21.2	12.00	10/31/2018	18:36

WORKLIST(Hardcopy Internal Chain)

LB 98962

WorkList Name : CORROSIONITY103118-1

WorkList ID : 118688

Date : 10/31/2018 4:42:43 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/31/2018	Water	J5729-04	Corrosivity	Cool 4 deg C	PSEG01	I12	WS1	10/31/2018	9040C
10/31/2018	Water	J5755-14	Corrosivity	Cool 4 deg C	PSEG01	I21	TP-2A	10/31/2018	9040C

Date/Time 10/31/18 6:10
 Received by: SP
 Relinquished by: CP

Date/Time 10/31/18 7:40
 Received by: CP
 Relinquished by: SP