



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: sweta

Run Number: LB98965

Ambient Barometric Pressure (mmHg): 740

Barometric Scale ID: 55971

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	81.8	11/01/2018	09:25
2	J5729-04		1	8	>100.0	>212.0	>212.0	11/01/2018	09:40
3	J5729-04DUP		1	8	>100.0	>212.0	>212.0	11/01/2018	10:02
4	J5755-14		1	7	>100.0	>212.0	>212.0	11/01/2018	10:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 98965

WorkList Name : fp111118

WorkList ID : 118711

Date : 11/1/2018 8:52:03 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/05/2018	Water	J5729-04	C Flash Point	Cool 4 deg C	PSEG01	I12	WS1	10/31/2018	1010A
11/05/2018	Water	J5755-14	C Flash Point	Cool 4 deg C	PSEG01	I21	TP-2A	10/31/2018	1010A

Date/Time 11/1/18 8:55 AM
 Received by: KS
 Relinquished by: CP

Date/Time 11/1/18 10:40 AM
 Received by: CP
 Relinquished by: KS